

Digital Color Doppler Ultrasound System

SonoMax 1/ SonoMax 2/ SonoMax 3/ SonoMax 5/ SonoMax 6
SonoMax 7/ SonoMax 7 Super/ SonoMax 7 EXP
SonoMax 8/ SonoMax 8 Super/ SonoMax 8 EXP
SonoMax 9/ SonoMax 9 Super/ SonoMax 9 EXP
SonoMax 10/ SonoMax 11/ SonoMax 22

Instruction Manual

CHGA-SonoMax-058

Jan. 29th, 2023

V1.0

CHISON Medical Technologies Co., Ltd.

Regulatory Requirement



This product conforms to the essential requirements of the Medical Device Regulation (EU) 2017/745. Accessories without the CE mark are not guaranteed to meet the Essential Requirements of the Medical Device Regulation.

This manual is a reference for the SonoMax 1/ SonoMax 2/ SonoMax 3/ SonoMax 5/ SonoMax 6/ SonoMax 7/ SonoMax 7 Super/ SonoMax 7 EXP/ SonoMax 8/ SonoMax 8 Super/ SonoMax 8 EXP/ SonoMax 9/ SonoMax 9 Super/ SonoMax 9 EXP/ SonoMax 10/ SonoMax 11/ SonoMax 22. Please verify that you are using the latest revision of this document. If you need the latest revision, contact your distributor.

NOTE:

Important

1. No part of this manual may be reduced, modified, copied or reprinted, in whole or in part, without written permission from CHISON.
2. The contents of this manual are subject to change without prior notice and without our legal obligation.
3. Before operating the system, please read and understand this manual. After reading, keep this manual in an easily accessible place. If you have any question or doubt, please contact CHISON's authorized service engineer.
4. CHISON's Warranty only cover material and parts costs for repair, but do not cover any labor cost or onsite service cost at end user's side.
5. 5.A notice to the user and/or patient that any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

NOTE:

Important information

1. It is the customer's responsibility to maintain and manage the system after delivery.
2. The warranty does not cover the following items, even during the warranty period:
 - a) Damage or loss due to misuse or abuse with system and probes, for example, drop the probe, the liquid or the metal part fall into the system.
 - b) Damage or loss caused by Acts of God such as fires, earthquakes, floods, lightning, etc.
 - c) Damage or loss caused by failure to meet the specified conditions for this system, such as inadequate power supply, improper installation or environmental conditions.
 - d) Damage or loss caused by non-approved transportation by CHISON.
 - e) Damage or loss due to use the system outside the region where the system was originally sold.
 - f) Damage or loss involving the system purchased from a source other than CHISON or its authorized agents.
3. Do not make changes or modifications to the software or hardware of this system and probes.
4. During operate the system, if user has any doubt, difficulty or any unclear, please contact CHISON's authorized service engineer immediately. Please describe the situation clearly to solve the question in time. Before solve the question, please don't operate the system.
5. This system shall not be used by persons other than fully qualified and certified medical personnel.
6. It is prohibited to use the device for fetal sex examination, except for necessary medical needs. The device can only be sold to qualified medical institutions or doctors. The users shall fully understand and master the devices before operating. The users shall have got the qualification, and shall comply with the local laws and regulations, the local religion and customs, etc.
7. The System modified or repaired by people other than CHISON's qualified service engineers, CHISON shall not be liable for the system.

8. The purpose of this system is to provide physicians with data for clinical diagnosis. It is the physician's responsibility for diagnostic procedures. CHISON shall not be liable for the results of diagnostic procedures
9. This manual contains warnings regarding foreseeable potential dangers, but user shall always be alert to dangers other than those indicated as well. CHISON shall not be liable for damage or loss that results from negligence or from ignoring the precautions and operating instructions described in this operation manual.
10. Due to negligence not following operation manual, CHISON shall not be liable for the results.
11. Each time before and after ultrasound examination, please check the probe surface, probe cable and sheath whether they are abnormal, such as cracking, peeling and deformation. Also check whether the lens is strongly fixed. Abnormal probes may cause electric shock and injure the patient. Once any abnormal, user must stop using and contact CHISON's authorized service engineer.
12. If the probe is dropped or scratched by hard part, please stop using the probe immediately. And contact CHISON's authorized service engineer to make sure the safety and effectiveness is in good condition before use.
13. If there is any liquid or metal to enter to the system, please power off the system and stop using it immediately. Please first contact CHISON's authorized service engineer to make sure it's safe before restart using it.
14. Please don't use solvents (such as paint thinner, benzine, or alcohol) or abrasive cleansers for cleaning the system (including monitor and probes, etc.). It may corrode the system and probes.
15. While the system or probe is over life time, please refer to operation manual section 9.5.
16. Important data must be backed up on external memory media. CHISON shall not be liable for loss of data stored in the memory of this system caused by operator error or accidents.
17. Please put this operation manual with the system to ensure operator and manager can reach it at any time.



CAUTION: *It is prohibited to use the device for fetal sex examination, except for necessary medical needs. The device can only be sold to qualified medical institutions or doctors. The users shall fully understand and master the devices before operating. The users shall have got the qualification, and shall comply with the local laws and regulations, the local religion and customs, etc.*



CAUTION: *The user should read the operation manual carefully before operating the devices. Turning on the device means the users have read the operation manual and accept the listed cautions, warnings, and notes in the manuals. If the users disagree and cannot accept the cautions, the users can ask for returning the device.*

TABLE OF CONTENTS

Chapter 1 Introduction.....	10
1.1 System Overview.....	10
1.2 Contact Information.....	10
1.3 Clinical Benefits.....	11
Chapter 2 System Safety.....	12
2.1 Safety Overview	12
2.2 Electrical Safety	13
2.3 Label.....	15
2.4 Patient Environmental Devices	17
2.5 Biological Safety.....	20
2.6 Scanning Patients and Education.....	21
2.6.1 Safe Scanning Guidelines	21
2.6.2 Understanding the MI/TI Display.....	23
Chapter 3 Preparing the System for Use.....	28
3.1 Site Requirements	28
3.1.1 Operation Environmental Requirements	28
3.1.2 Transport and Storage Environmental Requirements.....	29
3.1.3 Electrical Requirements	29
3.2 System Specifications	32
3.2.1 Console Overview	32
3.2.2 Physical Specification.....	35
3.2.3 Differences between models.....	36
3.2.4 Key System Features	43
3.2.5 Accessories	43
3.2.6 I/O ports.....	44
3.3 System Positioning & Transporting.....	44
3.4 Powering the System.....	45
3.4.1 Acclimation Time	45
3.4.2 Connecting and Using the System.....	45
3.4.3 Power On	46
3.4.4 Power Off.....	47
3.4.5 Standby	47
3.5 Probes.....	48
3.6 Optional installation.....	50
3.6.1 Connect the printer.....	50
3.6.2 Set the system for Video Printer	50
3.6.3 Connect the PC printer.....	51
3.6.4 Use the Gel Warmer.....	52
3.7 User Interface Control	53

3.7.1	Control Panel and Alphanumeric Keyboard	54
3.7.2	Soft-Menu Controls	61
Chapter 4	Imaging.....	62
4.1	General Description.....	62
4.2	Beginning an Exam	62
4.2.1	Selecting a Probe and an Application	62
4.2.2	Patient Data Entry	63
4.2.3	Display Interface.....	64
4.3	Optimizing the Image	66
4.3.1	Image Parameters Display	66
4.3.2	Scanning Modes.....	66
4.3.3	B Mode	67
4.3.4	FHI.....	68
4.3.5	Dual Mode	68
4.3.6	Quad Mode.....	69
4.3.7	B/M and M Mode	69
4.3.8	CFM Mode.....	70
4.3.9	B/BC Mode	73
4.3.10	CPA Mode.....	73
4.3.11	DPD Mode.....	74
4.3.12	PW Mode.....	74
4.3.13	Triplex.....	78
4.3.14	Quadplex	78
4.3.15	Curved Panoramic.....	78
4.3.16	2D Steer	79
4.3.17	Trapezoidal Imaging.....	79
4.3.18	Free Steer M Mode	79
4.3.19	Color M Mode.....	79
4.3.20	CW Mode	80
4.3.21	TDI.....	80
4.3.22	4D software package and Virtual HD	80
4.3.23	Elastography	81
4.3.24	ECG	90
4.3.25	Biopsy and Super Needle (Linear probe only)	90
4.3.26	Stress Echo	91
4.3.27	HPRF	93
4.3.28	TSS.....	93
4.3.29	Auto TGC.....	93
4.3.30	LGC	93
4.3.31	HD 3D.....	93
4.3.32	Compound.....	94
4.3.33	SRA	94
4.3.34	AIO	94
4.3.35	Q-image.....	94

4.3.36	Q-beam	94
4.3.37	Q-flow	94
4.3.38	X-con	94
4.3.39	Zoom	95
4.3.40	Auto IMT	95
4.3.41	Sono IMT+	95
4.3.42	Free NT	95
4.3.43	Human Measurement Package.....	95
4.3.44	Human Bodymark.....	95
4.3.45	SonoContrast	95
4.3.46	SoundFlow	96
4.3.47	SonoVector	97
4.3.48	Zoom	97
4.3.49	SonoFollicle	97
4.3.50	SonoOB	97
4.3.51	SonoFusion	97
4.3.52	SonoHelp.....	98
4.3.53	SonoColor	98
4.3.54	SonoBeam.....	98
4.3.55	Virtual Apex	98
4.3.56	SonoWifi	99
4.3.57	SonoBlue Tooth	99
4.3.58	Hard Disk.....	99
4.3.59	Strain and Strain Rate	99
4.3.60	Remote Control	99
4.3.61	DICOM	99
4.3.62	Optimize Image	100
4.3.63	B Mode Image Optimization	100
4.3.64	M Mode Image Optimization	102
4.3.65	Color Flow Map (CFM) Mode Image Optimization.....	102
4.3.66	CPA Mode Imaging Optimization	103
4.3.67	DPD Mode Imaging Optimization	104
4.3.68	Pulse Wave Doppler (PWD) Imaging Optimization.....	105
4.3.69	Continuous Wave Doppler (CWD) Imaging Optimization	106
4.4	After Capturing the Image	107
4.4.1	Adding Comments	107
4.4.2	Adding Body Mark	108
4.4.3	Saving Still Images	108
4.4.4	Saving Clips	108
4.4.5	Browse Images	111
4.4.6	Fast Storage	111
4.4.7	Files Management	112
4.4.8	Network Storage Function	113
4.4.9	DICOM Function	113

Chapter 5 General Measurements	120
5.1 Key for Measurement	121
5.2 Fast measurement	121
5.2.1 Enter fast measurement	121
5.2.2 Exit fast measurement	121
5.2.3 Fast measurement in B mode.....	121
5.2.4 Fast measurement in PW mode	123
5.2.5 Fast measurement in M mode	124
5.3 Measurement and Calculation	126
5.3.1 Measurement in B mode.....	126
5.3.2 Measurement in M mode	150
5.3.3 Measurement in PW mode	156
5.4 Edit measurement results	165
5.5 Report.....	165
Chapter 6 Preset	169
6.1 Recall Preset.....	169
6.2 Save User Defined Preset.....	169
6.3 Manage Preset.....	169
Chapter 7 System Setting	171
7.1 General.....	171
7.2 Measurement	176
7.3 Comment.....	178
7.4 BodyMark	181
7.5 Report.....	181
7.6 Admin	184
7.7 Remote Service	186
7.8 Scan Code	186
7.9 Network	187
7.10 System Settings	191
Chapter 8 Probes	200
8.1 General Description	200
8.2 Care and Maintenance	200
8.2.1 Inspecting Probes	201
8.2.2 Preparation at the Point of Use.....	201
8.3 Cleaning and Disinfecting	201
8.3.1 Surface probe as Non-critical	203
8.3.2 TV, TR probe as Semi-Critical	203
8.3.3 TEE probe as Semi-Critical	204
8.4 Using protective sheaths	206
Chapter 9 System Maintenance and Troubleshooting.....	207
9.1 Back up information.....	207

9.2	System Care and Maintenance	207
9.3	Safety Check	209
9.4	Troubleshooting.....	209
9.5	Service Responsibility.....	211
APPENDIX A	The Information of EU Representative	213
APPENDIX B	System one-key-recovery Function.....	214
APPENDIX C	Maximun Acoustic Output Report	216
APPENDIX D	Display Accuracy and Acoustic Measurement Uncertainties	402
APPENDIX E	Transducer Maximum Surface Temperature	403
APPENDIX F	Guidance and Manufacture's Declaration	404
APPENDIX G	Measurement Results Summary.....	409
APPENDIX H	Pulsed Wave Doppler Velocity Measurement Results Summary	410

Chapter 1 Introduction

This manual contains necessary information for safe system operation.

Read and understand all instructions in this manual before operating the system. Always keep this manual with the equipment, and periodically review the procedures for operation and safety precautions.

1.1 System Overview

Intended use

The SonoMax Series Digital Color Doppler Ultrasound System is an integrated preprogrammed color ultrasound imaging system, capable of producing high detail resolution intended for clinical diagnostic imaging applications, which is intended for use in multiple areas within a professional healthcare facility on one patient at a time.

Indications

The SonoMax Series Digital Color Doppler Ultrasound System is intended for diagnostic ultrasound imaging in B(2D/3D/4D), B/M, M, B+CFM, B+CPA(PD), B+DPD, B+PW, B+CW, B+CFM+D(PW)/CW, B+CPA(PD)+D(PW)/CW, TDI and Fusion Harmonic Imaging modes. The device is a general-purpose ultrasonic imaging instrument intended for use by a qualified clinician for evaluation of Fetal, Abdominal, Pediatric, Small Organ (breast, thyroid, testes), Neonatal Cephalic, Adult Cephalic, Cardiac (adult, pediatric), Musculo-skeletal (Conventional, Superficial), Peripheral Vascular, Trans-esophageal, Trans-rectal, Trans-vaginal, OB/GYN and Urology, which is intended to be used in a hospital or medical clinic.

Contraindication

The system is NOT intended for Ophthalmic use or any use that causes the acoustic beam to pass through the eye.

Patient target group

Adults and Children

Intended users

Qualified sonographers

1.2 Contact Information

For additional information or assistance, please contact your local distributor or the appropriate support resource shown below:

CHISON website	www.chison.com
Service Support	CHISON Medical Technologies Co., Ltd. Tel: 0086-400-8878-020; 0086-0510-85311707 Fax: 0086-0510-85310726

Placing an Order	E-mail: service@chison.com.cn
	CHISON Medical Technologies Co., Ltd.
	Tel: 0086-0510-8531-0593/0937
	Fax: 0086-0510-85310726
Manufacturer	Email: export@chison.com.cn
	CHISON Medical Technologies Co., Ltd.
	No.3 Changjiang South Road, Xinwu District, Wuxi, 214028 Jiangsu, P.R. China
US Agent	Mr. Marco Mu, 2219 Rimland Drive, Suite 301, Bellingham, Barkley Villiage Bellingham, Washington, 98226, UNITED STATES
	Phone: (702)209-5185, Fax: 360-9253199, Email: us.agent@mid-link.net
	MID-LINK INTERNATIONAL CO., LTD

1.3 Clinical Benefits

The expected clinical benefits of the SonoMax series diagnostic ultrasound systems are related to the device's intended purpose to provide diagnostic ultrasound imaging and fluid flow analysis of the human body. These clinical benefits can be broadly classified as providing real-time noninvasive or minimally invasive visualization of the internal organs and anatomy to assist in providing a medical evaluation and diagnosis to direct patients' medical care. Because the SonoMax series diagnostic ultrasound systems provide images of human anatomy without the use of ionizing radiation, the systems can provide information about a patient's health status, without the risks of some other medical imaging modalities

Chapter 2 System Safety

2.1 Safety Overview

This section discusses the measures to ensure the safety of both the operator and patient. To ensure the safety of both operator and patient, please read the relevant details in this chapter carefully before operating this system. Disregarding the warnings or violation of relevant rules may result in personal injury for operator or patient, or even loss of life.

Users should observe the following precautions:

- This system complies with Type BF general equipment, and the IEC standard. Please follow Chapter 2 “System Safety” in the operation manual to use this system properly.
- Please do not modify this system in any way. If modifications are necessary, please contact the manufacturer first to get more information and permission.
- This system has been fully adjusted at the factory. Do not adjust any fixed adjustable parts.
- In the event of a malfunction, turn off the system immediately and inform the manufacturer or its designated agents.
- The power cord of the system should be connected to a grounded power socket. Do not remove the ground cable for any reason.
- Only connect this system, either electronically or mechanically, with devices that comply with the IEC/EN60601-1 standard. Recheck the leakage current and other safety performance indices of the entire system to avoid potential system damage caused by leakage from a current superposition.
- The system does not incorporate any specialized protective measures in the event it is configured with high-frequency operation devices. The operator should use caution in these types of applications.
- The system should be installed only by personnel authorized by the manufacturer. Do not attempt to install the system by yourself.
- Only a CHISON's authorized service engineer can perform maintenance.
- Only a qualified operator, or someone under qualified supervision, can use the system.
- Do not use this system in the presence of flammable substances, otherwise an explosion may occur.
- Do not continuously scan the same part of a patient or expose the patient to prolonged scanning. Otherwise, it may harm the patient.
- When using the system for ultrasound testing, only use qualified ultrasound gel that complies with system standards.
- Do not unplug probe when the system is in active operation. Always go to transducer Selection screen when need to remove the probe.
- To prevent from arm or neck injury, the operator should not stay at the same position for too long during patient scanning without taking break.

- Remove the transesophageal probe from the patient prior to application of a defibrillator.
- Do not put liquid on top of the main unit.



WARNING: Use of this equipment adjacent to or stacked with other equipment should be avoided because it could result in improper operation. If such use is necessary, this equipment and the other equipment should be observed to verify that they are operating normally.



NOTE: To dispose of this product properly, please contact the local CHISON's Authorized Service Representative.

2.2 Electrical Safety

Type of protection against electric shock

Class I Equipment

CLASS I EQUIPMENT in which protection against electric shock does not rely on basic insulation only, but which includes an additional safety precaution in that accessible conductive parts are connected to the protective earthing conductor in the electrical installation in such a way that accessible parts cannot become live in the event of a failure of the basic insulation.

Degree of protection against electric shock

- Type BF Applied part(for Probes and ECG marked with BF symbol)

TYPE BF APPLIED PART providing a specified degree of protection against electric shock, with particular regard to allowable LEAKAGE CURRENT

Level of protection against harmful ingress of water

- Parts of **probe** likely to come into contact with **operator** or **patient** meet the requirements of **drip-proof equipment (IPX1)**
-
- Parts of **probe** intended to be immersed in **normal use** meet the requirements of **watertight equipment (IPX7)**
- The IP Classification of System is Ordinary Equipment (IPX0)
The equipment is not suitable for use in the presence of a flammable anesthetic mixed with air (with oxygen or with oxide)

Mode of operation

- Continuous Operation

For maximum safety, always follow these guidelines:

- Proper grounding of the system is critical to avoid electric shock. For protection, ground the chassis with a three-wire cable, and plug the system into three-hole outlet.
- Do not remove or circumvent the grounding wire.
- Do not remove the protective covers on the system. These covers protect users against hazardous voltages. Cabinet panels must remain in place while the system is in use. A qualified electronic technician must make all internal replacements.
- Do not operate this system in the presence of flammable gases or anesthetics.
- All peripheral devices (unless certified as medical grade) that are connected to the system must be powered through the electrical outlet with an optional isolation transformer.
- Suggest power off the system in 30 minutes if the system continuously works in 8 hours.

Notice upon Installation of Product

Separation distance and effect from fixed radio communications equipment: field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast, and TV broadcast transmitter cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the ultrasound system is used exceeds the applicable RF compliance level as stated in the immunity declaration, the ultrasound system should be observed to verify normal operation. If abnormal operation is observed, additional measures may be necessary, such as re-orienting or relocating the ultrasound system or using an RF shielded examination room may be necessary.

- Use either power supply cords provided by or designated by CHISON. Products equipped with a power source plug should be plugged into the fixed power socket which has the protective grounding conductor. Never use any adaptor or converter to connect with a power source plug (e.g. three-prong-to-two-prong converter).
- Locate the equipment as far away as possible from other electronic equipment.
- Be sure to only use the cables provided by or designated by CHISON. Connect these cables following the installation procedures (e.g. wire power cord separately from signal cables).
- Lay out the main equipment and other peripherals following the installation procedures described in this manual.

Notice against User Modification

The user should never modify this product.

User modifications may cause degradation in Electrical Safety. Modification of the product includes changes in:

- Cables (length, material, wiring, etc.)
- System configuration/components

- User modifications may cause degradation in EMC performance. Modification of the product includes changes in:
 - Cables (length, material, wiring, etc.)
 - System installation/layout
 - System configuration/components
 - Securing system parts (cover open/close, cover screwing)

2.3 Label

Label of Main Unit




Fig.2-1 Rear panel label

Label of Probes




Fig.2-2 Probe label

2.3.1 Symbols on label

 Caution, consult accompanying documents. This symbol advises the reader to consult the accompanying documents for important safety related information such as warnings and precautions that cannot be presented on the device itself.	 This mark indicates that this product contains a limited amount of hazardous substances in the Chinese Standard GB/T 26572-2011 "Limited Requirements for Restricted Substances in Electrical and Electronic Products". The numbers in the logo are the environmental protection use period of the product, indicating that under the
---	--

	normal use conditions, the harmful substances will not leak or be abrupt. The use of the product will not cause serious pollution to the environment or cause personal or property serious damage, the term unit is year.
 Do not use the following devices near this equipment: cellular phone, radio receiver, and mobile radio transmitter, radio controlled toy, etc. Use of these devices near this equipment could cause this equipment to perform outside the published specifications. Keep these devices power off when near this equipment.	 WASTE OF ELECTRICAL AND ELECTRONIC EQUIPMENT (WEEE): This symbol is used for Environment Protection, it indicates that the waste of electrical and electronic equipment must not be disposed as unsorted waste and must be collected separately. Please contact your local Authority or distributor of the manufacturer for information concerning the decommissioning of your equipment.
 This symbol is followed by the serial number of the device.	 The CE mark of Conformity indicates this equipment conforms with the Medical Device Regulation 2017/745 [MDR]
 This symbol is accompanied by the name and the address of the importing entity	 MANUFACTURER: This symbol is accompanied by the name and the address of the manufacturer.
 This symbol indicates that in the united states of America, Federal law restricts the device to sale by or on the order of a licensed practitioner or therapist.	 YYYY-MM This symbol indicates the country of manufacture of products is china, and this symbol is followed by the manufacturing date of the device in the form YYYY-MM.
 Refer to instruction manual.	 This symbol indicates the item is a medical device.
 This symbol indicates the item is a medical device	 Insulated patient application part (Type BF)
  (01) 0 6945121 40733 5 (11) 000000 (21) 123456789 This symbol indicates the UDI of the device, (01) is followed by the UDI-DI code of the device, (11) is followed by the manufacturing date of the device, (21) is followed by the serial number of the device.	 AUTHORIZED REPRESENTATIVE IN THE EUROPEAN COMMUNITY: This symbol is accompanied by the name and the address of the authorized representative in the European Community.

2.3.2 Other Device Labels

The following table describes the purpose of safety labels and other important information provided on the equipment.

Table 2-1: Symbol Icons

Icon	Meaning
IPX7	Protection against the effects of immersion (probes)

IPX0	No protection against ingress of water (system)
	In the process of push down the monitor, it is forbidden to place items to this region, otherwise there is oppression may flip monitor.
	Forbidden to push the monitor to move the machine.
	Forbidden to place heavy objects on the monitor after push down the monitor.
	Forbidden to press the monitor after push down the monitor.
	Forbidden to put the hand at the bottom of the monitor when turning the monitor.
	This part it is forbidden to impose extra upward or downward force (especially when transporting need to up and down movement machine).
	Be careful, arrow area may Clamp hand!
	No heavy load To prohibit the placing of heavy objects on a surface.

2.4 Patient Environmental Devices

Front side (refer to Fig. 3-1 in Chapter 3):

- 1 ECG port
- 4/5 Probe ports (Number of ports varies with system types)
- 1 Pencil Probe port

Left side (refer to Fig. 3-1 in Chapter 3):

- 1 DVD RW
- 2 USB ports on touchscreen

Right side (refer to Fig. 3-1 in Chapter 3):

- 2 USB ports on touchscreen

Rear panel (refer to Fig.3-1 in Chapter3):

- 2 USB ports
- 1 LAN port
- 1 S-VIDEO port
- 1 VIDEO OUTport
- L/R AUDIO port
- 1 VGA port
- 1 REMOTE port
- 1 HDMI port

Acceptable Devices

The Patient Environmental devices shown above are specified to be suitable for use within the PATIENT ENVIRONMENT.



CAUTION:

DO NOT connect any probes or accessories without approval by CHISON within the PATIENT ENVIRONMENT.

DO NOT touch patient and devices without IEC/EN 60601-1 approval to avoid the leakage current risk within the PATIENT ENVIRONMENT.

Unapproved Devices



CAUTION:

DO NOT use unapproved devices.

If devices are connected without the approval of CHISON, the warranty will be INVALID.

The system can't be used with HF surgical equipment; otherwise the burns to patient may occur.

Any device connected to this system must conform to one or more of the requirements listed below:

-IEC standard or equivalent standards appropriate to devices.

-The devices shall be connected to PROTECTIVE EARTH (GROUND).



CAUTION: *Unsafe operation or malfunction may occur. Use only the accessories, options and supplies approved or recommended in these instructions for use.*

Peripheral used in the patient environment

The system has been verified for overall safety, compatibility and compliance with the following on-

board image recording devices:

B/W video printer: Mitsubishi P93W, Sony UP-711MD, Sony UP-X898MD

Color video printer: Mitsubishi CP31W

The system may also be used safely while connected to devices other than those recommended above if the devices and their specifications, installation, and interconnection with the system conform to the requirements of IEC/EN 60601-1.

The connection of equipment or transmission networks other than as specified in the user instructions can result in an electric shock hazard or equipment malfunction. Substitute or alternate equipment and connections require verification of compatibility and conformity to IEC/EN 60601-1 by the installer. Equipment modifications, possible resulting malfunctions and electromagnetic interference are the responsibilities of the owner.

General precautions for installing an alternate off-board, remote device or a network would include:

- The added device(s) must have appropriate safety standard conformance and CE Marking.
- There must be adequate mechanical mounting of the device and stability of the combination.
- Risk and leakage current of the combination must comply with IEC/EN 60601-1.
- Electromagnetic emissions and immunity of the combination must conform to IEC/EN 60601-1-2.

Peripheral used in the non-patient environment

The system has been verified for compatibility, and compliance for connection to a local area network (LAN) via a wire LAN. The provided LAN components are IEC/EN 60950 compliant.

General precautions for installing an alternate off-board, remote device or a network would include:

- The added device(s) must have appropriate safety standard conformance and CE Marking.

The added device(s) must be used for their intended purpose having a compatible interface.



CAUTION: Make sure using *ONLY* the dedicated USB disk or removable media to save or back up data. Before connecting to the ultrasound system, make sure using the latest antivirus software on the USB disk or removable media to clean any virus. It is user's responsibility to ensure the USB disk or removable media is virus-free. Improper use of USB disk or removable media may cause the virus infections of system and eventually malfunction may occur. Such malfunction may impact the stability, effectiveness and safety of the system and probes, and users should immediately stop using the system and probes until CHISON authorized engineer has checked the system and confirm the effectiveness and safety of the system and probes.



CAUTION: Use only secure Local Area Network connection. Don't connect the ultrasound system to Internet. Make sure your hospital's firewall software is configured correctly, thus blocking incoming connection requests from Internet. Improper use of network connection may cause the virus infections of system and eventually malfunction may occur.

2.5 Biological Safety

This product, as with all diagnostic ultrasound equipment, should be used only for valid reasons and should be used both for the shortest period of time and at the lowest power settings necessary (ALARA - As Low As Reasonably Achievable) to produce diagnostically acceptable images. The AIUM offers the following guidelines:

Clinical Safety Quoted from AIUM

Approved March 26, 1997

Diagnostic ultrasound has been in use since the late 1950s. Given its known benefits and recognized efficacy for medical diagnosis, including use during human pregnancy, the American Institute of Ultrasound in Medicine herein addresses the clinical safety of such use:

There are no confirmed biological effects on patients or instrument operators caused by exposures from present diagnostic ultrasound instruments. Although the possibility exists that such biological effects may be identified in the future, current data indicate that the benefits to patients of the prudent use of diagnostic ultrasound outweigh the risks, if any that may be present.

Heating: Elevating tissue temperature during obstetrical examinations creates medical concerns. At the embryo development stage, the rise in temperature and the length of time exposed to heat combine to determine potential detrimental effects. Exercise caution particularly during Doppler/Color exams. The Thermal Index (TI) provides a statistical estimate of the potential temperature elevation (in centigrade) of tissue temperature. Three forms of TI are available: Soft Tissue Thermal Index (TIS), Bone Thermal Index (TIB) and Cranial Bone Thermal Index (TIC).

Soft Tissue Thermal Index (TIS). Used when imaging soft tissue only, it provides an estimate of potential temperature increase in soft tissue.

Bone Thermal Index (TIB). Used when bone is near the focus of the image as in the third trimester OB examination, it provides an estimate of potential temperature increase in the bone or adjacent soft tissue.

Cranial Bone Thermal Index (TIC). Used when bone is near the skin surface as in transcranial examination, it provides an estimate of potential temperature increase in the bone or adjacent soft tissue.

Cavitations: Cavitations may occur when sound passes through an area that contains a cavity, such as a gas bubble or air pocket (in the lung or intestine, for example). During the process of cavitations, the sound wave may cause the bubble to contract or resonate. This oscillation may cause the bubbles to explode and damage the tissue. The Mechanical Index (MI) has been created to help users accurately evaluate the likelihood of cavitations and the related adverse effects.

MI recognizes the importance of non-thermal processes, cavitations in particular, and the Index is an attempt to indicate the probability that they might occur within the tissue.

2.6 Scanning Patients and Education

The Track-3 or IEC60601-2-37 output display standard allows users to share the responsibility for the safe use of this ultrasound system. Follow these usage guidelines for safe operation:

- In order to maintain proper cleanliness of the probes, always clean them between patients.
- Always use a disinfected sheath on all EV/ER probes during every exam.
- Continuously move the probe, rather than staying in a single spot, to avoid elevated temperatures in one part of the patient's body.
- Move probe away from the patient when not actively scanning.
- Understand the meaning of the TI, TIS, TIB, TIC and MI output display, as well as the relationship between these parameters and the thermal/cavitation bioeffect to the tissue.
- Expose the patient to only the very lowest practical transmit power levels for the shortest possible time to achieve a satisfactory diagnosis (ALARA - As Low As Reasonably Achievable).

2.6.1 Safe Scanning Guidelines

- Ultrasound should only be used for medical diagnosis and only by trained medical personnel.
- Diagnostic ultrasound procedures should be done only by personnel fully trained in the use of the equipment, in the interpretation of the results and images, and in the safe use of ultrasound (including education as to potential hazards to the patient and the operator).
- Operators should understand the likely influence of the machine controls, the operating mode (e.g. B-mode, color Doppler imaging or spectral Doppler) and probe frequency on thermal and cavitations hazards.
- Select a low setting for each new patient. Output should only be increased during the examination if penetration is still required to achieve a satisfactory result, and after the Gain control has been adjusted to its maximum value.
- Maintain the shortest examination time necessary to produce a useful diagnostic result.
- Do not hold the probe in a fixed position for any longer than is necessary. It should be removed from the patient whenever there is no need for real-time imaging or spectral Doppler acquisition. The frozen frame and Cine loop capabilities allow images to be reviewed and discussed without exposing the patient to continuous scanning.
- Do not use endo-cavitary probes if there is noticeable self heating of the probe when operating in the air. Although applicable to any probe, take particular care during trans-vaginal exams during the first eight weeks of gestation.

- Take particular care to reduce output and minimize exposure time of an embryo or fetus when the temperature of the mother is already elevated.
- Take particular care to reduce the risk of thermal hazard during diagnostic ultrasound when exposing: an embryo less than eight weeks after gestation; or the head, brain or spine of any fetus or neonate.
- Operators should continually monitor the on-screen thermal index (TI) and mechanical index (MI) values and use control settings that keep these settings as low as possible while still achieving diagnostically useful results. In obstetric examinations, TIS (soft tissue thermal index) should be monitored during scans carried out in the first eight weeks after gestation, and TIB (bone thermal index) thereafter. In applications where the probe is very close to bone (e.g. trans-cranial applications), TIC (cranial bone thermal index) should be monitored.

MI > 0.3

There is a possibility of minor damage to neonatal lung or intestine. If such exposure is necessary, reduce the exposure time as much as possible.

MI > 0.7

There is a risk of cavitations if an ultrasound contrast agent containing gas micro-spheres is being used. There is a theoretical risk of cavitations without the presence of ultrasound contrast agents. The risk increases with MI values above this threshold.

TI > 0.7

The overall exposure time of an embryo or fetus should be restricted in accordance with Table 2-2 below as a reference:

TI	Maximum exposure time (minutes)
0.7	60
1.0	30
1.5	15
2.0	4
2.5	1

Table 2-2 Maximum recommended exposure times for an embryo or fetus

- Non-diagnostic use of ultrasound equipment is not generally recommended. Examples of non-diagnostic uses of ultrasound equipment include repeated scans for operator training, equipment demonstration using normal subjects, and the production of souvenir pictures or videos of a fetus. For equipment of which the safety indices are displayed over their full range of values, the TI should always be less than 0.5 and the MI should always be less than 0.3. Avoid frequent repeated exposure of any subject. Scans in the first trimester of pregnancy should not be carried out for the sole purpose of producing souvenir videos or photographs, nor

should their production involve increasing the exposure levels or extending the scan times beyond those needed for clinical purposes.

- Diagnostic ultrasound has the potential for both false positive and false negative results. Misdiagnosis is far more dangerous than any effect that might result from the ultrasound exposure. Therefore, diagnostic ultrasound system should be performed only by those with sufficient training and education.

2.6.2 Understanding the MI/TI Display

Track-3 follows the Output Display Standard for systems that include fetal Doppler applications. The acoustic output will not be evaluated on an application-specific basis, but the global maximum de-rated I_{spta} must be $\leq 720 \text{ mW/cm}^2$ and either the global maximum MI must be ≤ 1.9 or the global maximum de-rated I_{sppa} must be $\leq 190 \text{ W/cm}^2$. An exception is for ophthalmic use, in which case the $TI = \max(TIS_{as}, TIC)$ is not to exceed 1.0; $I_{spta.3} \leq 50 \text{ mW/cm}^2$, and $MI \leq 0.23$. Track-3 gives the user the freedom to increase the output acoustic power for a specific exam, and still limit output acoustic power within the global maximum de-rated $I_{spta} \leq 720 \text{ mW/cm}^2$ under an Output Display Standard.

For any diagnostic ultrasonic systems, Track-3 provides an Output Indices Display Standard. The diagnostic ultrasound systems and its operation manual contain the information regarding an ALARA (As Low As Reasonably Achievable) education program for the clinical end-user and the acoustic output indices, MI and TI. The MI describes the likelihood of cavitations, and the TI offers the predicted maximum temperature rise in tissue as a result of the diagnostic examination. In general, a temperature increase of 2.5°C must be present consistently at one spot for 2 hours to cause fetal abnormalities. Avoiding a local temperature rise above 1°C should ensure that no thermally induced biologic effect occurs. When referring to the TI for potential thermal effect, a TI equal to 1 does not mean the temperature will rise 1 degree C. It only means an increased potential for thermal effects can be expected as the TI increases. A high index does not mean that bioeffects are occurring, but only that the potential exists and there is no consideration in the TI for the scan duration, so minimizing the overall scan time will reduce the potential for effects. These operator control and display features shift the safety responsibility from the manufacturer to the user. So it is very important to have the Ultrasound systems display the acoustic output indices correctly and the education of the user to interpret the value appropriately.

RF: (De-rating factor)

In Situ intensity and pressure cannot currently be measured. Therefore, the acoustic power measurement is normally done in the water tank, and when soft tissue replaces water along the ultrasound path, a decrease in intensity is expected. The fractional reduction in intensity caused by attenuation is denoted by the de-rating factor (RF),

$$R_F = 10^{(-0.1 a f z)}$$

Where α is the attenuation coefficient in $\text{dB cm}^{-1} \text{ MHz}^{-1}$, f is the transducer center frequency, and z is the distance along the beam axis between the source and the point of interest.

De-rating factor R_F for the various distances and frequencies with attenuation coefficient $0.3\text{dB cm}^{-1} \text{ MHz}^{-1}$ in homogeneous soft tissue is listed in the following table. An example is if the user uses 7.5MHz frequency, the power will be attenuated by .0750 at 5cm, or $0.3 \times 7.5 \times 5 = -11.25\text{dB}$. The De-rated Intensity is also referred to as ' I_s ' at the end (e.g. $I_{spta.3}$).

Distance (cm)	Frequency (MHz)			
	1	3	5	7.5
1	0.9332	0.8128	0.7080	0.5957
2	0.8710	0.6607	0.5012	0.3548
3	0.8128	0.5370	0.3548	0.2113
4	0.7586	0.4365	0.2512	0.1259
5	0.7080	0.3548	0.1778	0.0750
6	0.6607	0.2884	0.1259	0.0447
7	0.6166	0.2344	0.0891	0.0266
8	0.5754	0.1903	0.0631	0.0158

$I' = I \cdot R_F$ Where I' is the intensity in soft tissue, I is the time-averaged intensity measured in water.

Tissue Model:

Tissue temperature elevation depends on power, tissue type, beam width, and scanning mode. Six models are developed to mimic possible clinical situations.

	Thermal Models	Composition	Mode	Specification	Application
1	TIS	Soft tissue	Unscanned	Large aperture ($>1\text{cm}^2$)	Liver PW
2	TIS	Soft tissue	Unscanned	Small aperture ($<1\text{cm}^2$)	Pencil Probe
3	TIS	Soft tissue	Scanned	Evaluated at surface	Breast color
4	TIB	Soft tissue and bone	Scanned	Soft tissue at surface	Muscle color
5	TIB	Soft tissue and bone	Unscanned	Bone at focus	Fetus head PW
6	TIC	Soft tissue and bone	Unscanned/scanned	Bone at surface	Transcranial

Soft tissue:

Describes low fat content tissue that does not contain calcifications or large gas-filled spaces.

Scanned: (auto-scan)

Refers to the steering of successive burst through the field of view, e.g. B and color mode.

Unscanned:

Emission of ultrasonic pulses occurs along a single line of sight and is unchanged until the transducer is moved to a new position. For instance, the PW, and M mode.

TI:

TI is defined as the ratio of the In Situ acoustic power (W_3) to the acoustic power required to raise tissue temperature by 1°C (W_{deg}), $TI = W_3/W_{\text{deg}}$.

Three TIs corresponding to soft tissue (TIS) for abdominal; bone (TIB) for fetal and neonatal cephalic; and cranial bone (TIC) for pediatric and adult cephalic, have been developed for applications in different exams.

An estimate of the acoustic power in milli-watts necessary to produce a 1°C temperature elevation in soft tissue is:

$W_{\text{deg}} = 210/f_c$, for model 1 to 4, where f_c is the center frequency in MHz.

$W_{\text{deg}} = 40 K D$ for model 5 and 6, where K (beam shape factor) is 1.0, D is the aperture diameter in cm at the depth of interest.

MI:

Cavitation is more likely to occur at high pressures and low frequencies in pulse ultrasound wave in the tissue, which contains the bubble or air pocket (for instance, the lung, intestine, or scan with gas contrast agents). The threshold under optimum conditions of pulsed ultrasound is predicted by the relation of the peak pressure to the square root of the frequency.

$$MI = Pr' / \sqrt{f_c}$$

Pr' is the de-rated (0.3) peak rare-fractional pressure in Mpa at the point where PII is the maximum, and f_c is the center frequency in MHz. PII is the Pulse Intensity Integral that the total energy per unit area carried by the wave during the time duration of the pulse. The peak rare- fractional pressure is measured in hydrophone maximum negative voltage normalized by the hydrophone calibration parameter.

Display Guideline:

For different operation modes, different indices must be displayed. However, only one index needs to be shown at a time. Display is not required if maximum MI is less than 1.0 for any setting of the operating mode, or if maximum TI is less than 1.0 for any setting of the operating mode. For TI, if the TIS and TIB are both greater than 1.0, the scanners need not be capable of displaying both indices simultaneously. If the index falls below 0.4, no display is needed. The display increments are no greater than 0.2 for index value less than one and no greater than 1.0 for index values greater than one (e.g. 0.4, 0.6, 0.8, 1, 2, 3).

Display and Report in Different Mode

Located on the upper middle section of the system display monitor, the acoustic output display provides the operator with real-time indication of acoustic levels being generated by the system.

For B-Scan Mode

Only display and report MI, and start from 0.4 if maximum MI > 1.0, display in increments of 0.2.

For Color Mode

Only display and report TIS or TIB and start from 0.4 if maximum TI > 1.0, display in increments of 0.2 for values of indices of 2.0 or less, and 0.5 for values of indices greater than 2.0.

For Doppler Mode

Only display and report TIS or TIB and start from 0.4 if maximum TI > 1.0, display in increments of 0.2 for values of indices of 2.0 or less, and 0.5 for values of indices greater than 2.0.

Below is a simple guideline for the user when TI exceeds one limit exposure time to $4^{(6-TI)}$ minutes based on the 'National Council on Radiation Protection. Exposure Criteria for Medical Diagnostic Ultrasound: I. Criteria Based on Thermal Mechanisms. Report No.113 1992'.

Operator Control Features:

The user should be aware that certain operator controls may affect the acoustic output. It is recommended to use the default (or lowest) output power setting and compensate using Gain control to acquire an image. Other than the output power setting in the soft-menu, which has the most direct impact on the power; the PRF, image sector size, frame rate, depth, and focal position also slightly affect the output power. The default setting is normally around 70% of the allowable power depending on the exam application mode.

Controls Affecting Acoustic Output

The potential for producing mechanical bioeffects (MI) or thermal bioeffects (TI) can be influenced by certain controls.

Direct: The Acoustic Output control has the most significant effect on Acoustic Output.

Indirect: Indirect effects may occur when adjusting controls. Controls that can influence MI and TI are detailed under the bioeffect portion of each control in the Optimizing the Image chapter.

Always observe the Acoustic Output display for possible effects.

Best practices while scanning

HINTS: Raise the Acoustic Output only after attempting image optimization with controls that have no effect on Acoustic Output, such as Gain and TGC.



WARNING: *Be sure to have read and understood control explanations for each mode used before attempting to adjust the Acoustic Output control or any control that can affect Acoustic Output. Use the minimum necessary acoustic output to get the best diagnostic image or measurement during an examination. Begin the exam with the probe that provides an optimum focal depth and penetration.*

Acoustic Output Default Levels

In order to assure that an exam does not start at a high output level, the system initiates scanning at a reduced default output level. This reduced level is preset programmable and depends upon the exam icon and probe selected. It takes effect when the system is powered on or New Patient is selected. To modify acoustic output, adjust the Power Output level on the Soft Menu.

Cable Information

Except for cables sold as spare parts for internal components, the use of accessories and cables outside the regulations may result in increased emission or a decreased immunity of the powered stapler for single use. This device contains the following cables:

Cable or transducer	Manufacturer	Model and specification
Power cable	KING-CORD	2m long, 60227IEC 53 (RVV) 3×0.75mm ² 300/500V unshielded wire
Cable of probe	CHISON	2m, shielded wire
Foot Switch	ZHEJIANG KACON ELECEIC CO.	2m, unshielded wire
ECG three-lead wire	Shenzhen Med-link Electronics Tech Co.	3.6m, unshielded wire

WIFI

Model: XPC240200B

RF Protocols: 5GHz Wi-Fi 20M 802.11n

Antenna Type and Antenna Gain: On-Board antenna 4dBi

Modulation: OFDM with BPSK, QPSK, 16-QAM, 64-QAM

Effective Transmitted Power (ERP): 13dBm

Basic performance

- ▮ Display images
- ▮ Doppler audio and spectrum display
- ▮ The measurement function is normal
- ▮ The ultrasonic output is normal
- ▮ The ECG function is triggered normally
- ▮ The system interface functions normally

Chapter 3 Preparing the System for Use

3.1 Site Requirements

3.1.1 Operation Environmental Requirements

The following environmental conditions are within system tolerances for operation:

Temperature:	10° C ~ 40° C
Relative Humidity:	30%~75%, non-condensing
Atmosphere Pressure:	700hPa ~ 1060hPa

Strong radiation sources or powerful electromagnetic waves (e.g. electro-magnetic waves from radio broadcasting) may result in image ghosting or noise. The system should be isolated from such radiation sources or electromagnetic waves.

To prevent damage to the system, do not use in the following locations:

- Exposed to direct sunlight
- Subject to sudden changes in temperature
- Dusty
- Subject to vibration
- Near heat generators
- High humidity



NOTE:

- ***This equipment generates, uses and can radiate radio frequency energy. The equipment may cause radio frequency interference to other medical and non-medical devices and radio communications. To provide reasonable protection against such interference, this product complies with emissions limits for a Group 1, Class A Medical Devices Directive as stated in IEC/EN 60601-1-2. However, there is no guarantee that interference will not occur in a particular installation.***
- ***If this equipment is found to cause interference (which may be determined by turning the equipment on and off), the user (or qualified service personnel) should attempt to correct the problem by one or more of the following measure(s):***
 - ***reorient or relocate the affected device(s).***
 - ***increase the separation between the equipment and the affected device.***
 - ***power the equipment from a source different from that of the affected device.***
 - ***consult the point of purchase or service representative for further suggestions.***

3.1.2 Transport and Storage Environmental Requirements

The following environmental transport and storage conditions are within system tolerances:

Temperature: -5° C ~ 40° C

Relative Humidity: ≤80% non-condensing

Atmosphere Pressure: 700hPa ~ 1060hPa

3.1.3 Electrical Requirements

Power Requirements

AC 100-240V, 50-60Hz

Fuse Requirements

Fuse specification is 250V, 6.3 A (time-lag), the model is 50CT-063H



CAUTION: *Please use the fuse provided by manufactory if the fuse need replace. User can't buy and exchange the fuse by themselves.*

Power Consumption: 600 VA

Voltage Fluctuation



WARNING: *Maintain a fluctuation range of less than ±10% of voltage labeling on rear panel of the system, otherwise the system may be damaged.*

Grounding

Before connecting the power cord, connect the attached ground protection cable from Equipotentiality terminal on system rear panel to a specialized grounding device.



NOTE:

- ***Please follow the outlined power requirements. Only use power cords that meet the system guidelines—failure to follow these procedures may result in system damage.***
- ***Line power may vary in different geographic locations. Refer to the detailed ratings on the rear panel of the system for detailed information.***

Built in battery specifications

Battery model	BT-3000	
Capacity	6450mAh	
Rated voltage	14.4V	
Standard charge voltage	16.8V	
Discharge closing voltage	11V	
Discharge time	Less than 20 minutes	
Standard charge current	1800mA	
Maximum continuous discharge current	9000mA	
Battery structure	4S3P	
Cycle life	300 times	
Charging time	About 150 minutes	
Operating temperature	Charge	0℃～55℃
	Discharge	-20℃～65℃
	Storage	-20℃～60℃ for less than 1 month; -20℃～30℃ for less than 6 months
Battery status indicator	1%-100%	Power balance display
	Charge tips	Charge indicator

Adapter specifications

Adapter model	MKP150-19A
Input	100-240V~,50-60Hz
Output	19V  11.84A



NOTE:

- *To avoid the battery bursting, igniting, or fumes from the battery to cause the equipment damage.*
- *Do observe the following precautions:*
 - 1) *Do not immerse the battery in water or allow it to get wet.*
 - 2) *Do not put the battery into a microwave oven or pressurized container.*
 - 3) *If the battery leaks or emits an odor, remove it from all possible flammable sources.*
 - 4) *If the battery emits an odor or heat, is deformed or discolored, or in a way appears abnormal during use, recharging or storage, immediately remove it and stop using it.*
 - 5) *The battery shall be storied within -10℃～45℃ range environmental condition .If stored for a long time (exceed three months), the cell should be stored in dry and*

cooling place. The cell's storage voltage should be 14.0~14.8V and the cell is to be stored in a condition as Temperature: $23\pm 5^{\circ}\text{C}$, Humidity: $65\pm 20\%\text{RH}$.

6) Upon receipt of the SonoMax series and before first time usage, it is highly recommended that the customer performs one full discharge/charge cycle. If the battery has not been used for over 2 months, the customer is recommended to perform one full discharge/charge cycle. One Full Discharge/Charge Cycle Process: 1. Full discharge of battery to let the SonoMax series automatically shut down. 2. Charge the SonoMax series to 100% FCC (full current capacity). When storing packs for more than 3 months, charge the pack at least once during the 3 month timeframe to prevent leakage and deterioration in performance.

7) Do not dismantle the battery. If need to change the battery, please contact CHISON's authorized service engineer.

8) To avoid the battery bursting, igniting, or fumes from the battery to cause the equipment damage.



CAUTION:

1. Do not immerse the battery in water or allow it to get wet.
2. Do not use or store the battery near sources of heat such as a fire or heater.
3. Do not use any chargers other than those recommended.
4. Do not put the battery into a fire or apply direct heat to it.
5. Do not short-circuit the battery by connecting wires or other metal objects to the positive (+) and negative (-) terminals.
6. Do not pierce the battery casing with a nail or other sharp object, break it open with a hammer, or step on it.
7. Do not strike, throw or subject the battery to severe physical shock.
8. Do not attempt to disassemble or modify the battery in any way.
9. Do not place the battery in a microwave oven or pressurized container.
10. Do not use the battery if it gives off an odor, generates heat, becomes discolored or deformed, or appears abnormal in any way. If the battery is in use or being recharged, remove it from the device or charger immediately and discontinue use.
11. Do not use or store the battery where is exposed to extremely hot, such as under window of a car in direct sunlight in a hot day. Otherwise, the battery may be overheated. This can also reduce battery performance and/or shorten service life.
12. If the battery leaks and electrolyte gets in your eyes, do not rub them. Instead, rinse them with clean running water and immediately seek medical attention. If left as is, electrolyte can cause eye injury.

3.2 System Specifications

3.2.1 Console Overview



Fig. 3-1 a: Console Overview

 NOTE:

The appearance may vary slightly with types. Please refer to the actual machine you have purchased.

The following pictures show the system in different views.




Fig. 3-1 b: Front Side View





Fig. 3-1 c: Right Side View






Fig. 3-1 d: Back Side View

3.2.2 Physical Specification

Dimensions of main unit (approx.): 685mm (Length) *604mm (Width) *1728.5mm (Height)

Net weight of main unit (approx): 58kg (no probe included)

3.2.3 Differences between models

Model	SonoMax 1	SonoMax 2	SonoMax 3	SonoMax 5	SonoMax 6	SonoMax 7	SonoMax 7 Super	SonoMax 7 EXP	SonoMax 8	SonoMax 8 Super	SonoMax 8 EXP	SonoMax 9	SonoMax 9 Super	SonoMax 9 EXP	SonoMax 10	SonoMax 11	SonoMax 22
5th Probe connector	option	option	option	option	option	option	option	option	option	option	option	√	√	√	√	√	√
LED	√	√	√	√	√	√	√	√	√	√	√	√	√	√	√	√	√
LED Size	21.5	21.5	21.5	21.5	21.5	23.8	23.8	23.8	23.8	23.8	23.8	23.8	23.8	23.8	23.8	23.8	23.8
Touch Screen	√	√	√	√	√	√	√	√	√	√	√	√	√	√	√	√	√
CD/DVD	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Gel Warmer	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Integrated Battery	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Footswitch	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
B mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
FHI	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
B/M mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
M mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
Dual mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
Quad mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
CFM mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
CPA mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
DPD mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
PW mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
CW mode	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
B/BC mode	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
2D Steer	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Triplex	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Quadplex	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Free Steer M	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option

Model	SonoMax 1	SonoMax 2	SonoMax 3	SonoMax 5	SonoMax 6	SonoMax 7	SonoMax 7 Super	SonoMax 7 EXP	SonoMax 8	SonoMax 8 Super	SonoMax 8 EXP	SonoMax 9	SonoMax 9 Super	SonoMax 9 EXP	SonoMax 10	SonoMax 11	SonoMax 22
mode																	
Free Curved M	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
HPRF	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
TDI	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Color M mode	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
TSS	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Curved Panoramic	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoContrast	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SoundFlow	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoVector	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Auto TGC	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
LGC	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
HD 3D	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Stress Echo	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Strain and Strain Rate	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Zoom	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
DICOM	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Trapezoidal Imaging	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Compound	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SRA	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
ECG	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Hard disk	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
Human Bodymark	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Auto IMT	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option

Model	SonoMax 1	SonoMax 2	SonoMax 3	SonoMax 5	SonoMax 6	SonoMax 7	SonoMax 7 Super	SonoMax 7 EXP	SonoMax 8	SonoMax 8 Super	SonoMax 8 EXP	SonoMax 9	SonoMax 9 Super	SonoMax 9 EXP	SonoMax 10	SonoMax 11	SonoMax 22
Sono IMT+	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Free NT	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Biopsy	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Super Needle	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Human measurement package	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
4D software package	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Virtual HD	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoFollicle	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoOB	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoFusion	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoHelp	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
X-con	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Q-image	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Q-flow	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Q-beam	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoColor	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoNeedle	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoPW	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Extended Imaging	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Intelligent doppler	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Volume Flow	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Elastography	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoCNS	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoBreast	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option

Model	SonoMax 1	SonoMax 2	SonoMax 3	SonoMax 5	SonoMax 6	SonoMax 7	SonoMax 7 Super	SonoMax 7 EXP	SonoMax 8	SonoMax 8 Super	SonoMax 8 EXP	SonoMax 9	SonoMax 9 Super	SonoMax 9 EXP	SonoMax 10	SonoMax 11	SonoMax 22
SonoThyroid	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoBladder	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Lung Assistant	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Auto B-Line	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoPleural	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoWork	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Auto EF	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Photo Tool	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoCoach	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Echo Compare	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Screen Saver	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
Report	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
Measurement	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
Comment	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
Bodymark	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
Hot Key	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard	standard
HL7	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoFIS	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoVocal	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Remote Control	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Image Sharing	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Printer Manager	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Remote Update	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Barcode	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option

Model	SonoMax 1	SonoMax 2	SonoMax 3	SonoMax 5	SonoMax 6	SonoMax 7	SonoMax 7 Super	SonoMax 7 EXP	SonoMax 8	SonoMax 8 Super	SonoMax 8 EXP	SonoMax 9	SonoMax 9 Super	SonoMax 9 EXP	SonoMax 10	SonoMax 11	SonoMax 22
Reader																	
Electronic Signature	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SWE	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Raw Data	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
NanoFlow	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Microflow MVI	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
ATI	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
PWV	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Auto Face	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoBeam	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Virtual Apex	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
AIO	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoWifi	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
SonoBlue tooth	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
Standby	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L4-10	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L4-10R	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L6-18	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L3-8	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L4-15B	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L3-10	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L5-14	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L4-15	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L5-15P	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L11-20	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option
L6-15i	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option	option

[illegible]

Components of L5-10W and S1-5W

1. Transducer part: Allowable maximum temperatures for skin contact with ME EQUIPMENT APPLIED PARTS.
2. Cover part: Allowable maximum temperatures for ACCESSIBLE PARTS that are likely to be touched by operators.



3.2.4 Key System Features

- Display B, B/M, M, Dual, Quad, CFM, CPA, DPD, PW, B/BC, 2D Steer, Triplex, Quadplex, CW, Free Steer M, HPRF, TDI, Color M, TSS, Curved Panoramic, SonoContrast, SoundFlow, SonoVector, HD 3D, Stress Echo, Trapezoidal Imaging, ECG, Biopsy, Super Needle, FHI, Elastography, 4D, VirtualHD, SonoOB, SonoFusion, SonoColor, SonoBeam, Virtual Apex
- X-con, Q-image, Q-flow, Q-beam, AIO, Zoom, Compound, SRA
- LGC, Auto TGC
- Human measurement package
- Remote Control
- Touch Screen, CD/DVD, Gel Warmer, Integrated Battery, Footswitch, Hard Disk
- SonoWifi, SonoBlue Tooth, Standby, Screen Saver
- DICOM
- Auto IMT, Sono IMT, Free NT, SonoFollicle, Strain and Strain Rate
- SonoHelp

3.2.5 Accessories

Transducers

No.	Probe Type	Probe Name
1	Linear probe	L4-10
2	Linear probe	L4-10R
3	Linear probe	L6-18
4	Linear probe	L3-8
5	Linear probe	L4-15B
6	Linear probe	L3-10
7	Linear probe	L5-14
8	Linear probe	L4-15
9	Linear probe	L5-15P
10	Linear probe	L11-20
11	Linear probe	L6-15i
12	Convex probe	C1-5
13	Convex probe	C1-6
14	Convex probe	C1-7
15	Convex probe	C1-7P
16	Micro convex probe	C2-6
17	Micro convex probe	C4-11
18	Micro convex probe	E4-10
19	Micro convex probe	E3-14
20	Micro convex probe	E4-13
21	Micro convex probe	BL3-12
22	Volume probe	VE4-10
23	Volume probe	V2-6
24	Pencil probe	CW2
25	Tee probe	T4-6

No.	Probe Type	Probe Name
26	Tee probe	MT4-6
27	Phased array probe	S1-5P
28	Phased array probe	S1-5
29	Phased array probe	S2-8
30	Phased array probe	S4-12
31	Linear probe	L5-10W
32	Phased array probe	S1-5W
33	Linear probe	ML4-6
34	Convex probe	MC1-5
35	Phased array probe	MS1-5
36	Phased array probe	LA1-15

3.2.6 I/O ports

- VGA output for external monitor
- S-VIDEO, VIDEO output for B&W video printer or Color video printer
- Remote port for video printer
- LAN port output for PC printer, DICOM and image review station
- 6 USB 3.0 ports for flash drive
- 2 AUDIO ports for audio R/L transport
- HDMI port for video transport

3.3 System Positioning & Transporting

Moving the System

When moving or transporting the system, take the precautions described below to ensure maximum safety for personnel, the system and other equipments.

Before Moving the System

- Completely switch off the system. See Section 3.4.4 “Power Off” for more information.
- Unplug the power cord (if the system is plugged into wall outlet).
- Disconnect all cables from off-board peripheral devices (external printer, etc.) from the console.



NOTE: To prevent damage to the power cord, DO NOT pull excessively on the cord or sharply bend the cord while wrapping it.

- Disconnect all probes from main unit. See Section 3.5 “Probes” for more information.
- Store all probes in their original cases or wrap them in soft cloth or foam to prevent damage.
- Replace gel and other essential accessories in the appropriate storage case.
- Ensure that no loose items are left on the main unit.

When Moving the System

Move the system by pushing or pulling the handle. Use extra care when crossing door or elevator thresholds.



NOTE: Always use the handle to move the system. In order to avoid possible injury or equipment damage :

- Walk slowly and carefully when moving the system.
- Do not let the system strike walls or doorframe.

Transporting the System

Use extra care when transporting the system in a vehicle. After preparing the system as described above, take the following additional precautions:

- Before transporting, place the system in its original storage case.
- Ensure that the system is firmly secured while inside the vehicle.
- Load the unit aboard the vehicle carefully and over its center of gravity.
- Keep the storage case still and upright. Secure that the system firmly with straps or as directed within the vehicle to prevent movement during transport.
- Any movement, coupled with the weight of the system, could cause it to break loose. Drive carefully to prevent damage from vibration.
- Avoid unpaved roads, excessive speeds, and erratic stops or starts.

3.4 Powering the System

3.4.1 Acclimation Time

After being transported, the unit requires one hour for each 2.5 °C increment if its temperature is below 10 °C or above 40 °C.




NOTE: Please keep at least 20 to 30 cm spare space away from the back of the system to ensure well ventilation. Otherwise, with the increasing of the temperature inside the unit, malfunction may occur.

3.4.2 Connecting and Using the System



NOTE: Before using the ultrasound system with a mains power supply connection, it is necessary to connect one end of the grounding wire from the accessories to the grounding rod of the ultrasound system, and the other end should be connected to a well-grounded metal object in contact with the earth. The diagram below illustrates the equipotential connection:



To connect the system to the electrical supply:

- Check the power voltage input labeling at rear panel of the system.
- Ensure that the wall outlet is the appropriate type and well grounded.
- Ensure that the system powers off.
- Unwrap the power cord, and allow sufficient slack in the cable so that the plug will not be pulled out of the wall outlet if the system is moved slightly.
- Attach the power plug to the system and secure it in place by using the retaining clamp.
- Push the power plug securely into the wall outlet.

 NOTE:

Only use the power cord provided by Manufacturer.

Use caution to ensure that the power cord does not disconnect during system use.

If the system is accidentally unplugged, data may be lost.

 WARNING:

To avoid risk of fire, the system power must be supplied from a separate, properly rated outlet.

Under no circumstances should the AC power plug be altered, changed, or adapted to a configuration rated less than specified. Never use an extension cord or adapter plug.

To help assure grounding reliability, connect to a “hospital grade” or “hospital only” grounded power outlet.

3.4.3 Power On

 **NOTE:** *Turn on the green power switch (main power circuit breaker switch, see Fig. 3-1 d in Section 3.2.1 Console Overview) at the back of the system, and then press the Power button on the top left of control panel to turn on the system.*

Power up Sequence:

The system is initialized and start-up status is reflected on the monitor:

- control panel flashing and then getting dark
- system checking BIOS data

- booting the operation system
- loading software
- entering examination status

HINTS

The power up procedure takes about approx. 100 seconds. If a problem occurs, take a picture and record the error information for service reference.



- ***After power off the system, please wait for more than 3 minutes to power on again.***
- ***When the system is powered on, for safety, please avoid moving the system.***

3.4.4 Power Off

To power off the system:

- Press the Power button on the top left of control panel.
- When the screen shows “Turn Off”, “Standby”, and “Cancel”, press “Turn off” to shutdown the system.



If the system is down or has not fully shut down, press and hold the Power button located on the top left of control panel for more than 4 seconds and release it, this will force the system to shut down completely.

- ***Disconnect the probes. Clean or disinfect all probes as necessary. Store them in their original cases to avoid any damage.***
- ***To ensure the system is disconnected from the power source, disconnect power plug from the wall outlet.***

3.4.5 Standby

To enter standby:

- Press the power button on the top left of control panel.
- Select “Standby” to enter into the standby status.

To exit standby: press the power button.



NOTE:

- ***Power off the system if you will not use the system for a long period of time (including storage/transportation condition), and you should not allow the system in standby status, otherwise the batteries will be out of power and permanently damaged.***
- ***If you will not use the system for a long period of time, DO NOT leave the system in the standby status, you should shut down the system, disconnect mains power, and turn off powers of all connected peripherals.***

3.5 Probes



NOTE: Only use the probes approved by Manufacturer.

Selecting probes

- Choose the probe according to the different examination.
- Begin the scanning session by choosing the correct application and preset for the examination.

Connecting the Probe

When you connect the probes, please ensure that the probe ports are not active. Place the system in “Transducer Selection” interface by pressing PROBE-key to deactivate the probe ports.

To connect a probe:

- Place the probe’s carrying case on a stable surface and open the case.
- Carefully remove the probe and unwrap the probe cord.
- DO NOT allow head of the probe hang freely. Impact to head of the probe could result in irreparable damage.



NOTE: Inspect the probe before and after each use for damage or degradation to the housing, strain relief, lens, seal and connector. DO NOT use a probe that appears damaged until its functional and safe performance is verified. A thorough inspection should be performed during the cleaning process.

- Align the connector with the probe port and carefully push into place with the cable facing the back of the system.

- Turn the probe connector locking lever to “lock” status.
- Carefully position the probe cord so it is free to move and is not resting on the floor.
- When the probe is connected, the system will be automatically recognized.

 CAUTION:

- ***Fault conditions can result in electric shock hazard. DO NOT touch the surface of probe connector that is exposed when the probe is removed. DO NOT touch the patient when connecting or disconnecting a probe.***
- ***Take precautions with probe cables. DO NOT bend the cable acutely.***




Fig.3-2 a Probe connector “Unlock” status




Fig.3-2 b Probe connector “Lock” status

Deactivating the Probe

When deactivating the probe, the probe is automatically placed in a standby mode.

To deactivate a probe:

- Ensure the system is in “Transducer Selection” interface. If necessary, press the PROBE-key to return.
- Gently wipe off the excess gel from the probe surface.
- Carefully slide the probe toward the probe holder, and place the probe gently in the probe holder.

Disconnecting the Probe

Probes can be disconnected when the system is "Transducer Selection" interface.

To disconnect a probe:

- Turn the connector locking lever to an "Unlock" position.
- Pull the probe and connector straight out of the probe port.
- Carefully slide the probe and connector away from the probe port.
- Ensure that the head of the probe is clean before placing the probe in its storage box.

Transporting the Probe

When transporting a probe a long distance, store it in its original carrying case.

Storing the Probe

It is recommended that all probes should be stored in the original carrying case.

- Place the probe connector into the carrying case.
- Carefully wind the cable into the carrying case.
- Carefully place the probe head into the carrying case. DO NOT use excessive force or impact on the probe head.

3.6 Optional installation

3.6.1 Connect the printer

- 1) It needs three cables: Remote cable, Video signal cable, Power cord.
- 2) Connect the remote cable to remote port on the back panel of ultrasound system.
- 3) Connect the video signal cable to the Video Out port on the back panel of the ultrasound system.
- 4) Connect the Power cord to the Power output of auxiliary mains on the back panel of the ultrasound system.



NOTE: *If you don't connect remote cable, you still can do the printing by pressing the key on printer.*

3.6.2 Set the system for Video Printer



CAUTION: *Please confirm the video printer is turned on and connected well with the main unit, then you can do below setting.*

- 1) Press the  key to enter "System Settings" interface, select "General", click "Keyboard".
See picture in Fig.3-3.




Fig.3-3

- 2) Choose “Video Print” under Print Key or Touch Screen Print Button Option, and set the “Video Print & Output Option”.

“Only Picture” means only print the ultrasound image.

“Picture and Info” means print the ultrasound image with patient information.

“Full Screen” means print the full screen image.

- 3) Press the print key on keyboard or use touch screen for printing.



NOTE:


- **You need to restart the system after connect the cables between Video printer and the System.**
- **You can't print the system information.**

3.6.3 Connect the PC printer

1. Place the printer smoothly.
2. Connect the printer to the system.
3. Set the print manager. Please see more information in 7.10.
4. Choose “PC Print & Image Save Selection” in system setting, chooses “Picture and Info”, or “Only Picture”.
5. Choose “PC Print” under Print Key.
6. Press the print key on keyboard for printing.

3.6.4 Use the Gel Warmer

Insert the interface of the gel warmer into the circular concave hole in the lower left corner at the back of the keyboard. It can be fixed through the fixing screw hole and the auxiliary fixing slot.



1. Button Description

Button Name	Operation	Function Description
Temperature setting button	Press (less than 1 second)	Setting temperature: It can be set to 37°C and 40°C. (Save the settings automatically and call the settings automatically every time you start up)
	Long Press (more than 1 second)	Power On/Off

2. Indicator Light Description

Indicator Light Status	Description
Indicator light turns green for 37℃	Temperature reaches to 37℃
Indicator light turns green for 40℃	Temperature reaches to 40℃
Two green indicators blink slowly at the same time (Blinking at 2Hz)	Overtemperature alarm (over 42.5℃) 【Indicators turn normal after malfunction is resolved】
Two green indicators blink quickly at the same time (Blinking at 0.5Hz)	Temperature sensor malfunction 【Indicators turn normal after malfunction is resolved】
	Voltage input is too low (Lower than 7V) 【Indicators turn normal after malfunction is resolved】
	Voltage input is too high (Higher than 24V) 【Indicators turn normal after malfunction is resolved】
Off	Power off

3. Overtemperature Protection

In the case of single fault, the warmer has overtemperature protection function, which can automatically cut off the heating power when the temperature of the aluminum drum rises to $50\pm 10^{\circ}\text{C}$. The reset temperature of the overtemperature protection device is $\geq 33^{\circ}\text{C}$.

3.7 User Interface Control

- B gain, Color gain and Doppler gain
- TGC
- Brightness
- Acoustic power
- Gamma
- Smooth
- Edge enhance
- Persistence
- Depth control
- Focal position/number
- Dynamic range selection
- Audio volume control
- Q-image
- Space compound imaging
- Freeze/Unfreeze
- Image storage
- Scanning width
- Zoom


- Dual display: Dual B or color
- Quad display
- L/R inversion
- U/D inversion
- Biopsy guide
- PRF
- Wall filter
- Blood Effecton
- Steering
- Color ROI panning
- Doppler sample volume adjustment
- Doppler angle correction
- Baseline movement
- Time base scrolling speed
- Annotation
- Patient data entry
- Measurement and calculation package
- File management and image archiving
- Clip image saving
- DICOM setting
- User defined preset


3.7.1 Control Panel and Alphanumeric Keyboard





Fig. 3-4: Overview of Control Panel


The main function of each key is introduced as below.

 POWER	✧ Press system power key momentary on the left top of control panel to turn on the system. Press this key and choose “turn off” to turn off the system. Press this key longer than 4 seconds to force the system to shut down in case the system is down. Press this key and choose “standby” to enter stand by status.
 SK1-SK6	✧ Press or rotate the keys to change the parameters of the corresponding menu at the bottom of the touch screen.
 TGC Sliders	✧ Manipulate the TGC (Time Gain Compensation) with 10 pairs of sliders.
 STILL SAVE	✧ Store still images.
 CINE SAVE	✧ Store selected clips.
 FREEZE	✧ Freeze/Un-Freeze the ultrasound image and enter/quit the Cine mode automatically.
	✧ Push it left for 3D function; push it right for 4D function. ✧ Angle: Rotate it for angle correction. In the Spectral Doppler mode, the default angle correction feature remains active. In the real time or cine modes, rotate this knob to adjust the Doppler Angle Correction by lining up the cursor with the vessel wall for an accurate reading. The Doppler Angle Correction setting can be adjusted 1 degree at a time. Rotate it to adjust the probe direction on the body mark status.
 B	✧ Press the B-knob to turn on B mode imaging. The system will stay in B mode if the current state is B, or return to B-mode if the current state is not B (e.g. M, color, Duplex Doppler). ✧ Rotate this knob to change the overall B gain throughout the image.

 CPA	Activate/turn off CPA Mode (also named as Color Power Angio mode). ✧ Press the CPA-key to turn on the CPA mode if the system is in B mode; ✧ Press the CPA-key for second time to turn off CPA and return to the previous mode (either B-mode or duplex Doppler).
 DEPTH	✧ Change the image depth of view.
 FOCUS	✧ Change the focus position.
 C	✧ Press the C-knob to turn on the Color Flow Map (CFM) mode if the system's current state is B; ✧ Press this knob can turn on Color if the system's current state is duplex Doppler; Press the C-knob for second time to turn off color and return to the previous mode (either B-mode or Duplex Doppler). ✧ Rotate C-knob to change the overall Color gain for CFM (PD) mode. ✧ In 4D mode, rotate C-knob to rotate image by X-axis.
 PW	✧ Press the PW-knob to turn on the duplex Doppler duplex mode if the current mode is B; ✧ Rotate the knob to change the overall Doppler gain for PW mode, when activate the spectral Doppler mode; ✧ In 4D mode, rotate PW-knob to rotate image by Y-axis.
	✧ Fingerprint recognition
 ZOOM	✧ In B and CFM mode, rotate it to zoom the image. Press twice to enter multiple enlargement.
 VOL.	✧ Rotate it to adjust the volume of the Doppler spectrum.
 CW	✧ If the probe supports CW mode, e.g. a phased array probe, press CW-key starts CW mode. The CW control operates in the same manner as the PW.

 M	✧ Press the M-knob to enter B/M mode if the current mode is B; ✧ Press the M-knob for second time to enter M-mode without B- mode ✧ Press the M-knob for third time to go back to B-mode ✧ Rotate the knob to change the overall M gain throughout the image. ✧ In 4D mode, rotate M-knob to rotate image by Z-axis.
 AIO	✧ AIO means Automatic Imaging Optimization. During image scanning, press this key will optimize the image for a better quality in resolution automatically.
 PRINT1	✧ Print the images when the printer is working.
 PRINT2	✧ Print the images when the printer is working.
 UPDATE	✧ Press the UPDATE key after the sample volume gate is defined to activate the Spectral Doppler mode. ✧ In Measurement mode, it can be used to switch between start point and end point (distance), long-axis and short-axis (ellipse), and return back to last position in trace measurement before the measurement is finished.
 CALC	✧ Use this key to activate calculation packages under different applications. This feature supports the optional OB, GYN, Vascular, Urology, Cardiac, Small parts, Pediatrics, Carotid Abdomen, TCD and General calculation packages. Refer to Measurement & Calculation section for details.
 MEASURE	✧ Enter into fast measurement.
 CURSOR	✧ Press to display cursor.

 P1	✧ Function config key. Set the function on the system settings interface.
 P2	✧ Function config key. Set the function on the system settings interface.
 P3	✧ Function config key. Set the function on the system settings interface.
 P4	✧ Function config key. Set the function on the system settings interface.
 Clear	✧ Delete all comments, arrows, bodymarks and measure tracks on the screen.
 Comment	✧ Enter to the Comment function.
 Bodymark	✧ Enter to the Bodymark function.
 Single	✧ Enter to the Single display.
 Dual	✧ Enter to the Dual display.
 Arrow	✧ Enter to the ARROW mark function.
 Elastic	✧ Enter to Elastography mode.
 SonoContrast	✧ Enter to SonoContrast mode.

 SonoOB	✧ Activate SonoOB function.
 Remote Service	✧ Activate Remote Service function.
 Enter	✧ Confirm the command entry; ✧ Confirm EXAM mode and menu setting; ✧ Confirm caliper and measurement setting; ✧ Toggle Trackball function between Re-sizing and Re-positioning for the CROI, and Doppler Sample Volume Gate.
	✧ It works as UPDATE button.
 Trackball	✧ Position calipers in measurement; ✧ Position 'mouse' cursor for exam mode selection; ✧ Position the M-mode, PW cursor; ✧ Select entry in soft-menu; Select EXAM mode; ✧ Position and re-size the Color Region of Interest (CROI); ✧ Position and re-size the Doppler Sample Volume Gate; ✧ Control digital cine review frames.
 Alphanumeric Keyboard	The Aplhanumeric Keyboard hides below the control panel. ✧ Input comment by alphanumeric keyboard. ✧  : Enter to the system recovery mode. ✧  : Turn down the volume. ✧  : Turn up the volume. ✧  : Delete all comments, arrows, bodymarks and measure tracks on the screen. ✧  : (Function key) user can set the function on the system setting.


	✧ P5 : (Function key) user can set the function on the system setting.
--	---

3.7.2 Soft-Menu Controls

The Soft-Menu is activated depending on the current active mode. The Soft-Menu will provide a second level control to set the parameters in the system. The default setting depends on different applications.

Soft-Menu provides the user with an easy and flexible approach to access additional system controls. The system will display the appropriate menus for the selected mode and functions.

There are two soft-Menu controls, user can press or rotate the SK1-SK6 keys on the control panel or touch the parameters on the touch screen to set the value.



Press or rotate the SK1-SK6 to adjust the value of the corresponding parameters

Fig. 3-5: Overview of touch screen

Chapter 4 Imaging

4.1 General Description

- How to begin an exam
- How to select a probe and an application
- How to optimize the image
- The operations after getting the image: adding annotation and body mark, storing and recalling the image

4.2 Beginning an Exam

Begin an exam by entering new patient information. You should enter as much information as possible, such as patient ID, patient name.

The patient name and ID number are retained with each patient's image and transferred with each image during archiving or hard copy printing.



CAUTION: To avoid patient identification errors, always verify the identification with the patient. Make sure the correct patient identification appears on all screens and hard copy prints.

4.2.1 Selecting a Probe and an Application

The system has five/six built-in probe connectors, so it can connect five/six probes at the same time.

Press **[Probe]** button on the touch screen to back to "Transducer Selection" interface.

1. Select the proper probe.
2. Select the proper Application.
3. Click the preset to enter into the default preset.




Fig.4-1

4.2.2 Patient Data Entry

Press the **[Patient]** button on the touch screen to display the patient interface.

Fig.4-2


Method of entering the patient information:

- 2) Move the cursor to input area and input the patient information through the keyboard. Move trackball to exchange between each option. E.g., patient name, birthday (birthday result will be automatic calculated after entering the age), age (age result will be automatic calculated after

entering the birthday), sex, weight, height.etc.

- 3) Click “OK” key to save the data after entering the appropriate information, then the system will go back to B mode.

4.2.3 Display Interface



- Image parameter area: Display the parameters under current mode. The different modes have different parameters.
- Thumbnail area of saved images: This area displays the thumbnail images of current patient. Selecting thumbnail images can recall this image immediately.
- Image preview area: This area displays the image when the image is selected in the thumbnail area.
- Operation note area: The first line shows the status of the current system. The second line indicates the current function of the trackball. In the picture Fig.4-3, the first line display “freeze menu”. It means the system is in frozen status. The second line display “recall image”. It means moving the trackball can recall the images in current status.
- System information area: Display the current system information. Moving the mouse to different icons will display detailed information accordingly.



Display current network situation. It will display the IP address after connecting to the

network.



Display the status of current removable disks. If exist removable disks, click this icon can quickly enter into the storage manager interface and do the operation for disks.




Fig.4-4



Display space size of all the drivers.



Printer management, refer to section 7.10.



Task manager: display the state of current task.




Fig.4-5



Show the present state of charge and connection of adapter.



Click this icon to pop up remote control login box.



Click this icon to show the remote service interface for customer service.

4.3 Optimizing the Image

4.3.1 Image Parameters Display

B	Meaning	CFM/ PD	Meaning	PW	Meaning	M	Meaning
Freq.	Frequency	Freq.	Frequency	Freq.	Frequency	SR	Sweep Speed
FPS	Frame rate	PRF	PRF	PRF	PRF	D/G	Dynamic range G:Gain
D/P	D:Dynamic range P: Persistence	WF	Wall Filter	WF	Wall Filter		
GN	B Gain	GN	Color Gain	GN	Doppler Gain		
AP	Acoustic power	C/P	C: Color map /direction power P:Persistence	DA	Doppler angle		
D	Display depth			SV	Sample volume gate		
X-con	X-contrast						

4.3.2 Scanning Modes

The system can support the following modes:

- B mode
- B/M mode
- M mode
- Dual mode
- Quad mode
- CFM mode
- CPA mode
- DPD mode
- PW mode
- B/BC mode
- 2D Steer
- Triplex
- Quadplex
- CW mode
- Free Steer M mode
- TDI
- Color M Mode
- TSS
- Curved Panoramic
- SonoContrast
- SoundFlow
- SonoVector
- HD 3D
- Stress Echo
- Strain and Strain Rate
- Trapezoidal Imaging
- ECG
- Biopsy
- Super Needle
- 4D
- SonoOB
- SonoFusion
- FHI
- SonoColor
- Virtual HD
- LGC
- Virtual Apex
- HPRF
- Auto TGC
- Elastography

4.3.3 B Mode

Intended Use:

B-mode is intended to provide two-dimensional images and measurement capabilities concerning the anatomical structure of soft tissue. Press B-knob to enter into B mode. Rotate B-knob to adjust B gain.




Fig.4-6 B Mode

Extended Imaging

In B mode, click Extended Imaging in the B menu to extend the display area for image.

SonoWork

SonoWork is to apply the automatic workflow template protocol to each inspection part in ultrasonic system to carry out automatic scanning. It is designed to reduce the button operation of the user's, standardize the measurement process and improve the measurement efficiency. Click [SonoWork] in the B menu, the system will automatically pop up the item to be measured and guide the user to complete the whole measurement.

SonoCoach

SonoCoach is used to help operators to do examination of different body parts.

When operators change different presets, display some guide image of scanning parts and methods on the screen. Press [Help] button in the keyboard to open SonoCoach.

Echo Compare

In the System Setting, users can set the shortcut key in the Button Configuration interface of General

setting for Echo Compare function. Press the corresponding shortcut key to activate the function. The current scanning image is displayed in the right and the left side is blank.

When scanning, double click a certain saved image or movie in the thumbnail area, which is to make comparison with the current image or movie, the recalled image or movie is displayed in the left side. This function is also available in the Easyview and Archive module.

SonoVocal

SonoVocal is option. This function supports voice annotation and voice command. Voice annotation and voice command should be enabled before using. And a microphone should be connected to the main unit.

In the System Setting, click [General], then click [SonoVocal] to make settings for voice annotation and voice command.

Users can tick [Voice Annotation] and set the maximum time for voice annotation.

Tick [Voice Command] and a list of command in factory default is displayed.

Photo Tool

After the device is successfully connected to the main unit, choose [Photo Tool] in the device to enter the photo tool mode. Meanwhile, enable the photo tool permission for the device on the main unit. Then users can upload image in the device to the main unit. The scanning interface displays the uploaded image.

4.3.4 FHI

Fusion Harmonic Imaging technique can improve the signal-to-noise ratio and resolution of the image, display the inspected organ more clearly, enhance tissue contrast and reduce artifacts. Based on low frequency fundamental wave emission, the sound wave penetration is improved.

In B mode real-time status, click [FHI] in B menu on the touch screen to turn it on/off.

4.3.5 Dual Mode

In B mode, press  key, system will display the current image at the left side of the screen,

Press  key for a second time, system will freeze the image displayed at the left side and

activate the image displayed at the right side at the same time. Press  key continuously to achieve exchange of the freeze/real status between left side image and right side image. Press



key to go back to B mode.

! NOTE:

Only one image can be activated at a time.

Dual display is also available for the Color Mode and PD Mode and DPD Mode.




Fig.4-7 B/B Mode

4.3.6 Quad Mode

In B Mode, press  key on the touch screen, the image which is activated will be displayed at the upper left side of the screen, press  key continuously will freeze and activate the upper right image, lower left image, and lower right image in order. Press  key to go back to B mode.

! NOTE:

Only one image can be activated at a time.

Quad display is also available for the Color Mode.

4.3.7 B/M and M Mode

Intended Use:

M mode is used to determine patterns of motion for objects within the ultrasound beam. The most common use is for viewing motion patterns of the heart.

This system provides B mode image and M mode image at the same time. In B mode, press M knob to enter into prepare mode. Press M knob a second time to enter into B/M mode. Press M knob once more to enter M mode. M mode is fit for heart scanning and measurement.




Fig.4-8 B/M Mode

M-mode Exam Procedure:

- Get a good B-mode image. Survey the anatomy and place the area of interest near the center of the B-mode image.
- Press the M-knob, move the trackball to position the M cursor over the area that you want to display in M-mode.
- Adjust the Scanning Speed, TGC, Gain etc, as needed.
- Press the FREEZE-key to stop the M scanning.
- Record the image to hard disk or to the printer (hard copy device).
- Press FREEZE-key to continue scanning.
- Press M-knob to enter into M-mode.
- Press M-knob again to exit M-mode.

M-mode Scanning Hints:

Speed: control speed of M-mode update.

Color map: choose the color map type for M mode.

Dynamic range: adjust the dynamic range of M mode.

Layout: adjust the B/M mode layout.

4.3.8 CFM Mode

Intended Use:

CFM is a Doppler mode intended to add color-coded qualitative information concerning the relative velocity and direction of fluid motion within the B-mode image.

CFM is useful to see flow in a broad area. It allows visualization of flow in the CROI, whereas

Doppler mode provides spectral information in a smaller area. CFM is also used a stepping stone to Doppler mode. You can use CFM to locate flow and vessels prior to activating Doppler.

In CFM mode, move the trackball to change the position of sampling box. [STEER] menu is used to adjust the angle of color sampling box (if current probe is linear probe). Press ENTER key to fix the position of color sampling box and move the trackball to adjust the size of color sampling box. Press ENTER key again and move trackball to change the color sampling position again.

Press C knob to enter into CFM mode; after C knob light is on, rotate the knob for adjusting the gain of CFM.

CFM-mode Exam Procedure:

- Follow the same procedure as described under B-mode to locate the anatomical area of interest.
- After optimizing the B-mode image, add Color Flow.
- Move the color region of interest CROI as close to the center of the image as possible.
- Optimize the color flow parameters so that a high frame rate can be achieved and appropriate flow velocity can be visualized.
- Press FREEZE-key to hold the image in cine memory.
- Record color flow image as necessary.




Fig.4-9 CFM Mode

CFM Scanning Hints:

PRF: increase/decrease the PRF on the color bar. Imaging of higher velocity flow requires increased velocity scale values to avoid aliasing.

Wall Filter: affect low flow sensitivity versus motion artifact.

Color Map: allow you to select a specific color map. It shows the direction of the flow and highlights the higher velocity flows.

Twinkle Flow uses dynamic information of blood flow, such as velocity, signal strength and variance defined as different colors and elevations, to present blood flow in 3D effect, so as to improve spatial resolution, dynamic information display and continuity of subtle blood flow display.

This function is available under convex probe and supports all the blood flow modes: C, CPA and DPD.

Click [NanoFlow] to enter this mode. This mode mimics the movement of red blood cells.



Click [MVI] to enter microflow mode.




Fig. 4-11 Microflow MVI Mode

4.3.9 B/BC Mode

In active CFM mode, press the [B/BC] key on the touch screen to display a real B mode image at the right side of the screen and active color mode image at the left side of the screen.




Fig.4-10 B/BC Mode

4.3.10 CPA Mode

CPA Mode is a color flow mapping technique used to map the strength of the Doppler signal coming from the flow rather than the frequency shift of the signal. Using this technique, the ultrasound system plots color flow based on the number of reflectors that are moving, regardless of their velocity. CPA does not map velocity, therefore it is not subject to aliasing.

Press CPA -key to enter into the CPA mode and then CPA light is on.




Fig.4-13 CPA Mode

4.3.11 DPD Mode

In CPA mode, press DPD key on the touch screen to enter into DPD mode.

If you need go back to PD mode from DPD mode, you can press CPA key.

4.3.12 PW Mode

Intended Use:

Doppler is intended to provide measurement data concerning the velocity of moving tissues and fluids. PW Doppler lets you examine blood flow data selectively from a small region called the Sample Volume.

The X axis represents time while the Y axis represents velocity in either a forward or reverse direction.

PW Doppler is typically used for displaying the speed, direction, and spectral content of blood flow at selected anatomical sites.

PW Doppler can be combined with B-mode for quick selection of the anatomical site for PW Doppler examination. The site where PW Doppler data is derived appears graphically on the B-mode image (Sample Volume Gate). The Sample Volume Gate can be moved anywhere within B-mode image.




Fig.4-11 PW Mode

PW mode Exam Procedure:

- Get a good B-mode image. Press C-knob to help locate the vessel you wish to examine.
- Press PW-knob to display the sample volume cursor and gate.
- Position the sample volume cursor by moving the Trackball left and right. Position or re-size the sample volume gate by moving the Trackball up and down, then press ENTER-key.
- Press UPDATE-key to display PW Doppler spectrum and the system will run in combined B+Doppler mode. The Doppler signal can be heard through the speakers.
- Optimize the PW Doppler spectrum, as necessary.
- Ensure that the sample line is parallel to the blood flow.
- Press FREEZE-key to hold the trace in cine memory and stop imaging.
- Perform measurements and calculations, as necessary.
- Record results with your recording devices.
- Press FREEZE-key to resume imaging.
- Repeat the above procedure until all relevant flow sites have been examined.
- Replace the probe in its respective holder.

When pressing PW-knob for the first time, the Doppler spectrum is not activated. The Doppler Sample Volume appears in the default position, and the B mode image or 2D (either B or Color) mode are active. Moving the Trackball will change the Sample Volume position. Press the ENTER-key to toggle the Trackball function between Sample Volume Gate position and size. Press the UPDATE-key after the Sample Volume Gate is defined to activate the Spectral Doppler mode. Press the UPDATE-key for second time to toggle back to 2D (B or Color) update and deactivate the Spectral Doppler.

Doppler mode Scanning Hints:

The best Doppler data will be got when the scanning direction is parallel to the direction of the blood flow; when the scanning direction is perpendicular to the anatomic target, you can get the best B mode image, so you should keep the balance as you don't usually get both an ideal B-

mode image and ideal Doppler data simultaneously.

PRF: adjust the velocity scale to accommodate faster/slower blood flow velocity. Velocity scale determines pulse repetition frequency.

Wall Filter: remove the noise caused by vessel or heart wall motion at the expense of low flow sensitivity.

Baseline: adjust the baseline to accommodate faster or slower blood flows to eliminate aliasing.

Angle Correct: optimize the accuracy of the flow velocity. It estimates the flow velocity in a direction at an angle to the Doppler vector by computing the angle between the Doppler vector and the flow to be measured. This is special useful in vascular applications where you need to measure velocity.

Doppler Gain: allow you to control the background information of spectral.

Sweep Speed: control speed of spectral update.

Doppler Sample Volume Gate Position and Size (Trackball and Enter)

Move the sample volume on the B-mode's Doppler cursor. The gate is positioned over a specific position within the vessel.

- To move Doppler cursor position, turn the trackball left or right until positioned over the vessel.
- To move sample volume gate position, move the trackball up or down until positioned inside the vessel.
- To size sample volume gate, press ENTER-key to toggle trackball function from sample volume gate positioning to sizing, then move the trackball to change sample volume gate size.

SonoPW

In D mode, click [SonoPW] on the touch screen.

Under SonoPW, adjust the [Sample Number] to adjust the number of sampling lines.

It can collect the information of hemodynamics or motor coordination of multiple blood streams or tissues in the same cardiac cycle in real time.




Fig. 4-15 SonoPW Mode

Intelligent Doppler

Press the PW-knob twice to activate PW mode. Click [Intelligent Doppler] on the touch screen to open this function.

PWV


Start PWV

Click [PWV] in the right column of the touch screen in B mode to enter PWV tab page.


- Supported presets: Carotid artery, upper limb, lower limb
- Click [Start] to activate PWV automatic measurement.

Automatic Measurement


- A sampling gate is placed at the left and right ends of the blood vessel respectively, and the distance d between the center points of the two sampling gates is automatically obtained.



- The time difference t for identifying the starting point of two gate cardiac cycles on PW spectrum



- AutoCalc PWV= d/t



4.3.13 Triplex

- While in B mode, press C-knob and PW-knob to access the B+Doppler+Color mode, press Triplex key on the touch screen to enter the Triplex active mode.
- While in B mode, press CPA-key and PW-knob to access the B+Doppler+CPA mode, press Triplex key on the touch screen to enter the Triplex active mode.

4.3.14 Quadplex

- While in B mode, press C-knob and PW-knob to access the B+Doppler+Color mode, press Quadplex key on the touch screen to enter the Quadplex active mode.
- While in B mode, press CPA-key and PW-knob access the B+Doppler+CPA mode, press Quadplex key on the touch screen on to enter the Quadplex active mode.

4.3.15 Curved Panoramic

In B real time mode, turn on the Curved Panoramic on the touch screen, then move the probe

linearly to get a series of B mode images, and the system piecing together multiple B images into a single, extended B image. Curved Panoramic is available for linear probes.




Fig.4-12 Curved Panoramic Mode

4.3.16 2D Steer

2D Steer is available for linear probes. It can steer the beam to obtain the left or right image and enlarge the area without to rotate the probes.

In B mode, press [2D Steer] on the touch screen to adjust.

4.3.17 Trapezoidal Imaging

Trapezoidal image is available for linear probes. In B mode, press Trapezoidal Mode on the touch screen to enter to Trapezoidal Mode.

4.3.18 Free Steer M Mode

Free steer M mode is only available for phased array probe. This mode can give you the ability to manipulate the cursor at different angle and position. The M-mode display changes as per the M cursor position.

User can activate the Steer M mode using touch screen. Turn the item Free M Mode on, and turn Angle to rotate free steering M line. The system provides maximum 3 free steering M line and you can select either of them with ENTER-key.

Trackball: used to move the free steer M line.

Angle: used to rotate the free steer M line.

4.3.19 Color M Mode

Color M mode is used for fetal cardiac applications. Color flow overlays color on the M-mode image

using velocity and variance color maps. The color flow wedge overlays the B-mode image and M-mode timeline. The color flow maps available in M-mode are the same as in CFM mode.

Color M-mode is a Doppler mode intended to add color-coded qualitative information concerning the relative velocity and direction of fluid motion within the M-mode image.

If the system is in color mode and the probe supports Color M mode (e.g. phased array probe), press M-knob to enter Color M mode.

4.3.20 CW Mode

Continuous Wave Doppler allows examination of blood flow data all along the Doppler cursor rather than from any specific depth. Gather samples along the entire Doppler beam for rapid scanning of the heart. Range gated CW allows information to be gathered at higher velocities.

It works with a phased array or pediatric probe.

If the velocity of the blood flow is even too high for the HPRF mode to detect, you have to try CW mode. Press CW-key to enter CW prepare mode. Press CW-key again to activate the CW mode.




Fig.4-13 CW Mode

4.3.21 TDI

TDI mode is tissue Doppler mode, which is intended to provide information of low-velocity tissue motion, specifically for cardiac movement. Only phased array probe is available for TDI function.

Press TDI-key on the touch screen to enter into TDI mode.

TDI Parameter:

These parameters can only be changed when TDI mode is enabled.

Color Invert, Color Map, Persistence, Frequency, Baseline, PRF, Wall Filter.

4.3.22 4D software package and Virtual HD

Push the 4D key right to enter into 4D scanning. The software is used with volume probe to get the images, and then build the 3D image and display in the screen in real time.

In 4D mode, choose Virtual HD to enter Virtual HD mode. With a moveable virtual light source and the advanced skin rendering techniques, Virtual HD improves the 4D image to allow doctors to envision the fetus more vividly.

4D Parameters:

These parameters apply to 4D mode.

Layout, render mode, direction, quality, threshold, smooth.

Auto Face

After clicking the 4D button, click the Auto Face button on the touch screen to enter the function. The system will automatically obtain fetal facial features.

4.3.23 Elastography

Elastography shows the biological tissue elasticity properties with the ultrasound imaging system. It records the ultrasound signals before and after tissue distortion caused by applying external or internal forces. Based on the recorded signals, it analyses, estimates and displays the strain of biological tissue.

In B mode, press elastography key on the touch screen to enter into elastography mode.

The Elastography screen:




Fig.4-14 Elastography Mode

1--Elastography imaging 2--Elastography ROI 3--Color bar

Change the size and the position of the Elastography box by using the trackball.

Change between size and position by using the ENTER key.

SWE

Probes supporting shear wave elastic imaging: linear array probe and convex array probe


Enter/Exit SWE

Enter SWE

- 1) Press the [ELASTO] button under B mode on the touch screen to enter the elasto tab page.



- 2) Click [SWE] to enter the shear wave elastic imaging mode.



- 3) Tab page switches to SWE and SWE mode menu item displays.



Exit SWE

Press [B] button or press SK6 to exit the shear wave elastic imaging mode and return to B mode.

SWE Mode Menu Item

- 1)  Single
 Dual

In single mode, a shear wave image is displayed.

In dual mode, the shear wave image is displayed on the left and the 2D image is displayed on the right, which is the default mode.

By clicking this two menu items, user can switch between single and dual display mode.

2) Single Trigger

Click [Single Trigger] to trigger single frame shear wave elastic imaging, and the image will be frozen automatically.


3) Repeating

Click [Repeating] to start continuous shear wave elastic imaging and obtain shear wave elastic image in real time.

4) Confidence Map

Click [Confidence Map] under dual mode, the layout of the two images is as follows:

The left image shows the shear wave image, and the right one shows the confidence image.



Confidence map value: 0%~100% (Images with value over 50% are reliable.)

The shear-wave image whose confidence value is less than 50% is displayed transparently.

5) Map Type

Two map types are available: kPa, m/s

When kPa is selected, the shear wave map corresponds to the elastic map in kPa, and the unit of measurement result is kPa.

When m/s is selected, the shear wave map corresponds to the velocity map in m/s, and the unit of measurement result is m/s.

6) Propagation


In dual mode, click [Propagation], only single trigger function is available. The [Repeating] button is gray and not available.

Click [Single Trigger], the two images show in the following layout.


The image on the left shows the shear wave image (still image), and the image on the right shows the image of the propagation process of the shear wave.

After the propagation process is completed, the image is automatically frozen. Before freezing, the prompt at the bottom of the image shows: please wait a moment.... And no other operation is allowed.

After frozen, the prompt disappears. And move the trackball to view the image of each frame of the propagation image.



In this mode, the parameter bar is as follows:



Transparency (SWE) : Adjust the transparency of the shear wave image area accordingly. 5% adjustable each time. Rotate SK3 to adjust.

Transparency (Propagation) : Adjust the transparency of the propagation image area accordingly. 5% adjustable each time. Rotate SK5 to adjust.

7) Transparency (SWE)

Refer to the transparency (SWE) in propagation.

8) Transparency (Confidence)

Transparency of confidence image area

Adjust the transparency of the confidence map accordingly. 5% adjustable each time. Rotate SK5 to adjust.

9) Interval

Set the interval time for repeating function. Rotate SK4 to adjust.

Adjust ROI

When entering the shear wave elastic imaging, the ROI box will appear in the B image area. The size and shape of the ROI box are related to the probe.


In dual mode, the ROI box synchronous displays in both two images.

Move the trackball to change the position of the ROI box. Press [Enter] to confirm.

To adjust the size of the ROI box, press [Enter] button to activate the ROI box and then move the trackball upwards, downwards, left or right to resize of the ROI box. Press [Enter] button to confirm the adjustment.

Display Area

After the ROI box is set ready, click [Single Trigger] to obtain the dual images. Shear wave imaging needs time to refresh, and the 2D image and shear wave elastic image refresh synchronously.




Save


Press save image or save cine button to store the shear wave elastic image.
The saved image will be displayed in the upper right corner of the screen.

Measure

According to the selected preset, press [Calc] or [Freeze] button to enter the corresponding measurement package (general or small parts).



Measurement method for SWE



1. SWE (Circle)




The measurement method is ellipse and the ROI is circle, which is the default measurement method.

Operation procedure:

- 1) The first symbol "+" appears, and move the trackball to the starting point. Press [Enter] to confirm the starting point.
- 2) The second symbol "+" appears, and move the trackball to fix the circle.
- 3) Press [Enter] to confirm the end point, the result value of shear wave elasticity will be displayed, and the next measurement item will be automatically activated. 10 measurement items are available. If it is the last measurement item, the measurement ends.



2. SWE (Trace)



The measurement method is trace, and the lesion area is manually traced for measurement.

Operation procedure:

- 1) The first symbol "+" appears, and move the trackball to the starting point. Press [Enter] to confirm the starting point.
- 2) The second symbol "+" appears, and move the trackball to trace the edge of the lesion to be measured and press [Enter] button to set the end point.
- 3) The end point and the starting point of the trace will be automatically connected with a straight line. The result value of shear wave elasticity will be displayed, and the next measurement item will be automatically activated. 10 measurement items are available. If it is the last measurement item,

the measurement ends.



NOTE:

- 1) If the circled/traced area exceeds the elastic image area, only the value of the area with elastic value is calculated.
- 2) If no elastic image is in the circled/traced region, no elastic value shows.

Report

After finishing the exam, press the Report-key to pop up report interface, edit and print the report. The measurement result value of SWE will be added into the general or small parts report page.

Report Interface:

ATI

In the real-time single B mode of liver application under convex probe, press [ATI] in the right column of touch screen to enter the ATI tab page. Press [SK6] to exit ATI mode.

4.3.24 ECG

The ECG module is a device that provides the 3 lead ECG signal acquisition for cardiac application. It is not intent for the ECG diagnostic purpose as in the 12-lead module. In the cardiac application, the ECG trace is displayed on the bottom of the screen. For echo-stress, the R-wave triggering is used to gate or synchronize the image acquisition. The ECG has 3 leads: LL (left leg, RED), LA (left arm, BLACK), RA (right arm, WHITE). LA is for reference, which usually provides a bias voltage from the ECG module, and the LL, LA are the two signals from the body and going to the differential input of the ECG isolation amplifier.

The ECG control is in the soft-menu available for the cardiac probe, it allows the user to set up the following control:

ECG ON/OFF: turn on/off the ECG trace.

ECG GAIN: increase or decrease the ECG gain.

ECG POS: set the ECG trace position.

ECG Velocity: set the ECG Velocity.

Dynamic Range: adjust dynamic range.

4.3.25 Biopsy and Super Needle (Linear probe only)

1. How to enter into Biopsy

Press the Biopsy key on the touch screen in B mode to display biopsy line and verify the biopsy line before biopsy. Press the Biopsy again to turn off the biopsy.

2. How to adjust the biopsy

Press ENTER-key to edit the biopsy line and move the trackball to change the biopsy line position. Press ENTER-key again to fix the biopsy line.

3. Super needle

Super needle is used for enhance the needle image in the B mode image. After turn on the super needle, super needle and needle angle function will be active and user can adjust the needle angle to optimize the image for needle only (The angle is 5 degree per step.).



CAUTION:

Please confirm the biopsy line before using biopsy function. The biopsy tool provided by CHISON must be used for this diagnostic ultrasound instrument. Otherwise, the accuracy of biopsy cannot be guaranteed.

4. SonoNeedle

SonoNeedle enhances the puncture visibility. The system will identify and mark the position and movement track of the needle, which is convenient for users to observe.

This function is only for linear probe.

4.3.26 Stress Echo

It works with a phased array probe.

Press Protocol item on the touch screen to turn stress echo function on.




Fig.4-19

When the Protocol turns on, select the Default 4x4 template to start the study. User can also open the Timer1 for the study if needed. Save the cines for the study. After finish the study, enter into the Utility menu to select the Analyze to enter the Protocol Analyze interface. User can edit the Protocol Score for the study.




Fig.4-20

Press OK to exit the Protocol Analyze.

User also can edit the Template, define the Number levels and Number projection views.




Fig.4-21

4.3.27 HPRF

High Pulse Repetition Frequency (HPRF) is a special operating mode of PW Doppler. In HPRF mode, multiple energy pulses are used. This allows higher velocities to be detected without causing aliasing artifacts. HPRF mode is used when detected velocities exceed the processing capabilities of the currently selected PW Doppler scale or when the selected anatomical site is too deep for the selected PW Doppler scale.

HPRF is invoked when you are operating in PW Doppler mode and conditions activate HPRF when the velocity scale factor or sample volume gate depth exceeds certain limits. When HPRF is active, multiple sample volume gates appear along the Doppler cursor. Doppler information can be received from any of the multiple sample volume gates. The Doppler signals from all the gates are added together and displayed in one spectrum.

It works with a phased array or pediatric probe.

4.3.28 TSS

TSS is used to calculate the wave velocity synthesis of sound speed in different tissues, so as to obtain the best resolution in various modes. Press TSS button in the B menu to open this function. Press TSS button to make the selection of TSS. The TSS has 4 selections: General, Muscle, Flow and Fat.

4.3.29 Auto TGC

Time Gain Compensate (TGC) is also called as Depth Gain Compensation. In the process of ultrasonic propagation, the energy will decay with distance/time. Time-gain compensation compensates for this energy attenuation mainly in the direction of acoustic propagation (longitudinal).

4.3.30 LGC

In addition to the attenuation of ultrasonic energy along with depth, the energy of ultrasonic in the transverse direction will also be significantly reduced due to refraction, scattering, absorption and other reasons. In order to compensate for the attenuation of ultrasonic energy in a Lateral area, Lateral gain compensation (LGC) function is needed. By using this feature, we can compare and visualize structures and lesions in areas that are obscured by lateral sound, cartilage, intermuscular sound shadows, and so on.

4.3.31 HD 3D

HD 3D is a kind of three-dimensional ultrasono-graphy, which can generate 3D images and then enter into the 3D reconstruction state. Compared with traditional two-dimensional images, three-

dimensional ultrasound imaging has the advantages of intuitive image display and accurate location of diseased organs.

4.3.32 Compound

Compound (spatial and frequency composite) combines electronic acoustic beam control and traditional linear array technology to obtain real-time images from different perspectives. Composite images improve image quality by reducing blobs, cluttered echoes and other acoustic artifacts. They also better define structure boundaries and improve contrast resolution.

4.3.33 SRA

SRA uses noise reduction algorithm to remove fine speckle noise, smooth tissue image, reduce noise, enhance signal, and improve signal-to-noise ratio.

4.3.34 AIO

Automatic image optimization is to optimize the parameters of the image according to the tissue characteristics of in current scanning.

4.3.35 Q-image

Q-image can improve the image quality and make the image more delicate.

4.3.36 Q-beam

This function is a kind of image optimization technology based on the principle of traditional double beamformer. The multi-beamformer can improve the processing speed, thus improve the frame rate of image display and reduce noise, meet the clinical diagnosis of moving organs, improve line density, avoid or reduce interpolation processing, and restore the true information.

4.3.37 Q-flow

Q-flow is a kind of image optimization technology, which is based on advanced filtering technology and two-dimensional blood flow algorithm, can optimize the sensitivity of image color and facilitate the observation of peripheral circulation and micro-blood flow.

4.3.38 X-con

This function can optimize the contrast of ultrasound images of different tissues and improve the image quality by improving the signal-to-noise ratio.

4.3.39 Zoom

The system supports multilevel magnification. In the real-time scanning state, rotate the Zoom key to zoom the image. Press the Zoom key twice for local magnification. It supports multilevel magnification. Rotate Zoom key to adjust the magnification. In the zoom state, move the trackball to move the enlarged image so that it can be viewed locally. Press the Zoom key, then press the UPDATE key, select the size of ROI by sliding the trackball, press the UPDATE key again to fix the size of ROI. Select the location of ROI by sliding the trackball, and finally press ENTER or Zoom key again to achieve local amplification. To exit Zoom, press the Zoom key again or press FREEZE key.

4.3.40 Auto IMT

Auto IMT is a measurement item in B vascular measurement, which is to measure the thickness of intima. It is a semi-automated function.

In vascular preset, get a good B image for intima, and then click this measurement.

Set start position and end position, be sure to contain the intima in the box.

The system will calculate the thickness of intima, the result of IMT will be displayed on the screen.

4.3.41 Sono IMT+

Sono IMT+ is a function. In vascular preset, get a good B image for intima. Click [Sono IMT+] on the touchscreen, the system will automatically mark the sampling box, in which the intima is included.

4.3.42 Free NT

Free NT is a measurement item in B OB measurement, which is to measure the thickness of nuchal translucency.

4.3.43 Human Measurement Package

In B, M and PW mode, press 【CAL】 key to enter into corresponding human measurement package.

Please refer to Chapter 5.3.

4.3.44 Human Bodymark

The system configures with different human bodymarks according to the presets. Please refer to Chapter 4.4.2 for how to add a bodymark.

4.3.45 SonoContrast

SonoContrast utilize the characters that the contrast agent has a strong reflection on ultrasound when it enters the blood circulation of the human body to observe blood flow in various parts of the body through contrast agents with 2D images. SonoContrast possess better resolution and sensitivity than flow mode.

Press SonoContrast button to open the function.



SonoContrast can only be used for Liver, Cardiac, and Vesicoureteral reflux (VUR). Not for other parts of the body.

When entering contrast imaging mode, the screen displays the contrast image. Both the contrast image (marked with “Con”) and tissue image (marked with “C”) are displayed.

Function	Description
Center Line	Turn on/off the center line of the image.
Flash	Destroy the micro-bubbles by the previous contrast imaging, or observe the reinfusion effect in a continuous agent injecting process.
Timer1/Timer2	The two timers are used to record the total time of contrast imaging and the single time of one contrast exam.
Visual	Display ● Tissue ● Contrast ● Mix Mix combines the contrast image with tissue image

Perform image scanning, freeze the image and select a range of images for analysis, or select a desired cine loop from the stored images.

Click “TIC Analysis” in freeze menu to activate the time-intensity analysis function.

Mark out the part of interest (ROI). If necessary, perform curve fitting on the time-intensity curve. Analyze the parameters of the curve.

Click “Exit” in the menu or press “Freeze” to exit TIC Analysis.

4.3.46 SoundFlow

SoundFlow mode increases the sensitivity of detecting blood flow.

In B mode, press SoundFlow button on the touch screen to enter SoundFlow mode.

Press SoundFlow button on the touch screen again to exit SoundFlow Mode.

4.3.47 SonoVector

SonoVector is used to get motion information for each point in the blood flow and mark it with a straight line with an arrow.

Press SonoVector button under C mode to open the function. Use straight line with an arrow to display the motion of each point in the blood flow. Arrow display the direction and length display the velocity.

Press SonoVector button again to close SonoVector.

4.3.48 Zoom

Zoom is used to change the magnification of the global image, and the target area can be moved to the field of vision by moving the trackball.

Press Zoom knob to enter zoom mode. Rotate Zoom knob to change the magnification. Display thumbnails when images cannot be fully displayed in the image area. Move the trackball, and the area marker box will move with it. The image area will change synchronously.

Press Zoom knob to exit zoom mode. Enter PW mode, M mode will exit zoom mode automatically.

4.3.49 SonoFollicle

SonoFollicle can automatically identify the follicle, trace the follicle, and calculate the area and length. Press Calc on the control panel, and press Change on the control panel or Select GYN on the touch Screen to open the GYN measurement package. Select Fo Auto, select Follicle 1(/n), then click the follicle to be tested. The follicle is automatically traced and the area and length are calculated.

4.3.50 SonoOB

SonoOB can automatically identify and calculate the corresponding feature organization on the standard section, it includes Auto BPD, Auto OFD, Auto HC, Auto AC, Auto FL.

Press Calc on the control panel, and press Change on the control panel or select OB on the touch Screen to open the OB measurement package.

Select Auto BPD to automatically identify and calculate the BPD on the standard section.

Select Auto OFD to automatically identify and calculate the OFD on the standard section.

Select Auto HC to automatically identify and calculate the HC on the standard section.

Select Auto AC to automatically identify and calculate the AC on the standard section.

Select Auto FL to automatically identify and calculate the FL on the standard section.

4.3.51 SonoFusion

SonoFusion is used for spatial orientation with a sensor equipped on the ultrasound probe, further to

make real-time ultrasound image match with CT/MR image. With the advantages on the real-time display of the ultrasound image and premium image quality of CT/MR, it can improve the diagnostic efficiency greatly by moving the probe to change the CT/MR data in real-time.

Only the linear and convex probes support the SonoFusion in B/Color/CPA/DPD mode.

Press SonoFusion on the touch screen to open the function.

4.3.52 SonoHelp

When in SonoCoach mode, click [User Manual] on the touchscreen to activate SonoHelp function.

4.3.53 SonoColor

SonoColor can measure the value of blood flow velocity in C mode with color.

Open the measurement package in C mode and click SonoColor on the touch screen to measure the value of blood flow velocity.

4.3.54 SonoBeam


SonoBeam uses the Quad beam former to improve the processing speed, to improve the frame rate.

Press SonoBeam on the touch screen to open the function in Color mode.

4.3.55 Virtual Apex

The Virtual Apex technology is applied to the phased-array probe, which generates a virtual point and is topped up. The generated ultrasonic image covers the whole probe surface and expands the imaging range of the probe.

Press Virtual Apex on the touch screen to open the function.



4.3.56 SonoWifi

SonoWifi is used to transfer data through Wifi.

Double click SonoWifi button in the system icon area to open it. The data can be transfer using Wifi connection.

4.3.57 SonoBlue Tooth

SonoBlue tooth is used to transfer data through Blue tooth.

Double click SonoBlue tooth button in the system icon area to open it. The data can be transfer using Blue tooth connection.

4.3.58 Hard Disk

The system configures hard disk to save data and ultrasound software.

4.3.59 Strain and Strain Rate

Strain is an analysis of myocardial movement in the left ventricle. The myocardium is marked with spots and strain is calculated according to the spots' displacement. For a movie file, strain can be calculated for each frame of image. Taking time as the horizontal axis and strain as the vertical axis, the curve of strain changing with time is drawn, which is the strain-time curve. Strain rate is the ratio of strain to time.

In phased-array probe, press [FREEZE] to enter freeze menu. Choose [Strain] to activate this function. Choose area of interest and then click [Play], the system will automatically calculate the strain and strain rate. In the strain tab page, click [FREEZE] to exit.

4.3.60 Remote Control

This function realizes the remote control of the ultrasound main unit by tablet or phone.


After the device is successfully connected to the main unit, Choose [Remote Control] in the device to enter the remote control mode. Meanwhile, enable the remote control permission for the device on the main unit. Then users can operate on the device to control the main unit.

In the System Setting, click [Network], then click [Remote Control] to make settings.

Tick to open this function. The device name, SN and QR code are displayed.

In the device list, the bound device is displayed.

4.3.61 DICOM



After the DICOM is configured after optimization, the DICOM-related configuration will be automatically re-verified after the network cable is unplugged.

4.3.62 Optimize Image

Methods for optimizing:

1. Use soft menu

Press corresponding mode on the touch screen to display and exchange the soft menu.

2. Use shortcut menu

SK1-SK6 shortcut key is correspondence to NO.1—NO.6 menu, respectively. Rotate shortcut key will choose the menu function corresponded.

3. Use other keys

4.3.63 B Mode Image Optimization

Soft menu adjustment:

Menu	Function
Quad mode	Activate Quad B mode.
SRA	Turn on the SRA function.
Q-Image	Ajust the parameter of the Q-Image.
Compound	Turn on the compound function. SRA and Compound can't work at the same time.
Focus Num	Adjust focus number.
Full Screen	Turn on the full screen function.
Scan width	Adjust scan width of B image; frame rate will be fast if scan width is small.

Image Rotate	Change the direction of B image. The angle will be increased at 90° interval in clockwise direction.
X-contrast	Adjust X-contrast function.
A. power	Adjust acoustic power.
Persistence	Adjust the persistence function.
Smooth	Adjust the image smooth.
Edge enhance	Adjust edge enhance.
Density	Choose from low, mid and high.
2D Map	Choose the format of 2D map.
Gamma	Adjust gamma value.
Noise Reject	Adjust noise reject.
TSS	Activate TSS function.
4D	Activate 4D mode.

Shortcut adjustment:

Shortcut (Action)	Menu	Menu function
SK1 (rotate)	Frequency	Adjust frequency of the probe.
SK2 (rotate)	Dynamic	Adjust the image dynamic range, increase or decrease, System dynamic range and contrast resolution.
SK3 (rotate)	Depth	Adjust the image depth.
SK4 (rotate)	Zoom	Adjust the size of the image.
SK5 (rotate)	Focus Position	Adjust the position of the Focus.
SK6 (rotate)	Chroma	Adjust the value of chroma

Other adjustments:

1. Depth (DEPTH key)

DEPTH toggle key is used for the adjustment of image depth. Depth will increase if adjust upward; Depth will decrease if adjust downward. For the best resolution of focusing and edge enhancement, it needs to adjust TGC after the depth has been adjusted.

2. Gain (B knob)

Adjust gain of B mode could enhance or reduce quantities of echo information in images. Adjust the main gain can adjust sensitivity of overall images (brightness). Rotate the knob of B mode could adjust B gain after the B light is on. The gain will be increased if rotate clockwise; the gain will be decreased if rotate contra. Adjust range is from 0 to 255. TGC will influence on each other between gain adjustment and TGC adjustment.

3. TGC

TGC with 10 adjustable slides: Slide the set of slider bars can change the depth gain of 2D images.

TGC returns signal amplifier to correct for the attenuation caused when adding depth. TGC

balance the image and make the density of echo uniformly distributed in images. Similarly the enlargement of every sliding channel area is also different.

The range of TGC will redistribute according to the new depth while depth is changing.

Move the slide bar Left/Right can reduce/increase TGC.

4. Image magnification (ZOOM knob)

Press ZOOM knob, press UPDATE-key to choose the size of ROI box, and then press UPDATE-key again to choose the position. Press ENTER-key can realize multiple enlargements. Press ZOOM knob to exit operation.

4.3.64 M Mode Image Optimization

Soft menu adjustment:

Menu	Function
Layout	Choose from LR and UD
Display Format	Choose from 1:2,1:1,2:1
Color Map	Adjust color map of M image
Chroma	Adjust chroma of M image
2D Map	Adjust 2D map of M image

Shortcut menu adjustment:

Shortcut (action)	Menu	Menu Function
SK3 (rotate)	Dynamic	Adjust the image dynamic range.
SK4 (rotate)	Angle	Adjust the angle.
SK5 (rotate)	Speed	Adjust scanning speed.

4.3.65 Color Flow Map (CFM) Mode Image Optimization

Soft menu adjustment:

Menu	Function
B/BC	Turn on B/BC Mode
Color Invert	Realize invert of the color for blood flow
Q-beam	Turn on Q-beam function
Q-flow	Turn on Q-flow function
Color Map	Adjust the color map
Persistence	Adjust the value of persistence
Color Mode	Choose between Vel and Variance
Wall Filter	Adjust the wall filter
Quad mode	Activate Quad C mode.
Density	Change the density

Auto Color	Blood flow velocity class: low, normal, high
------------	--

Shortcut menu adjustment:

Shortcut (action)	Menu	Menu Function
SK1 (rotate)	Frequency	Adjust the frequency of probe.
SK2 (rotate)	Baseline	Change the baseline of color-map.
SK3 (rotate)	PRF	Change the value of PRF.
SK4 (rotate)	Steer	Adjust angle of sampling box of blood flow under the linear probe.
SK5 (rotate)	Blood effect	Adjust the angle.
SK6 (rotate)	Wall Thre.	Adjust the wall threshold.

Other adjustments:

1. CFM gain control.

Rotate C- knob after C lights is on. Adjust the color gain and adjust range is from 0 to 255.

2. Sampling box.

Moves sampling box through trackball to the area which you are interested and press the ENTER-key to confirm the position of sampling box then adjust the size of box through trackball and press ENTER-key to confirm.

Rotate SK4 to change the steer angle of sampling box under the linear probe.



NOTE: This function can only support linear probe.

4.3.66 CPA Mode Imaging Optimization

Soft menu adjustment:

Menu	Function
B/BC	Turn on B/BC Mode.
Dual mode	Turn on Dual Mode.
Q-beam	Turn on Q-beam function.
Q-flow	Turn on Q-flow function.
DPD	Exchange between CPA Mode and DPD Mode.
Color Map	Adjust the color map.
Persistence	Adjust the value of persistence.
Wall Filter	Adjust the wall filter.
Density	Change the density.
Wall Thre.	Adjust wall threshold.
Auto Color	Blood flow velocity class: low, normal, high

Shortcut menu adjustment:

Shortcut (action)	Menu	Menu Function
SK1 (rotate)	Frequency	Adjust the probe launch frequency.
SK2 (rotate)	Dynamic	Adjust the image dynamic range, increase or decrease, System dynamic range and contrast resolution.
SK3 (rotate)	PRF	Adjust PRF value.
SK4 (rotate)	Steer	Adjust angle of sampling box of blood flow under the linear probe.
SK5 (rotate)	SV	Adjust the size of the sample volume gate.

Other adjustments:

1. CPA gain control

Rotate C knob after CPA light is on, then adjust the gain and the range is from 0 to 255.

2. Sampling box.

Moves sampling box through trackball to the area which you are interested and press the ENTER-key to confirm the position of sampling box then adjust the size of box through trackball and press ENTER-key to confirm.

Rotate SK4 to change the steer angel of sampling box under the linear probe.



NOTE: This function can only support linear probe.

4.3.67 DPD Mode Imaging Optimization

Soft menu adjustment:

Menu	Function
B/BC	Turn on B/BC Mode.
Single mode	Back to single mode when in dual mode.
Dual mode	Turn on Dual Mode.
Q-Beam	Turn on Q-beam function.
Q-Flow	Turn on Q-flow function.
CPA	Exchange between DPD Mode and CPA Mode.
Color Invert	Realize invert of the color for blood flow.
Color Map	Adjust the color map.
Persistence	Adjust the value of persistence.
Wall Filter	Adjust the wall filter.
Density	Change the density.
Wall Thre.	Adjust wall threshold.
Auto Color	Blood flow velocity class: low, normal, high

Shortcut menu adjustment:

Shortcut (action)	Menu	Menu Function
SK1 (rotate)	Frequency	Adjust the probe launch frequency.
SK2 (rotate)	Dynamic	Adjust the image dynamic range, increase or decrease, System dynamic range and contrast resolution.
SK3 (rotate)	PRF	Adjust PRF value.
SK4 (rotate)	Steer	Adjust angle of sampling box of blood flow under the linear probe.
SK5 (rotate)	SV	Adjust the size of the sample volume gate.

4.3.68 Pulse Wave Doppler (PWD) Imaging Optimization

Soft menu adjustment:

Menu	Function
Triplex	Open the Triplex Ffunction
Quadplex	Open the Quadplex function
Auto Cal	Turn on Auto Cal function
Invert	Reverse the spectrum according to the baseline.
Display Format	1:1,1:2,2:1 can be chosen
QuickAngle	Turn on the QuickAngle function
Color Map	Change the color map
2D Map	Adjust the color of spectrum
Wall Filter	Change Wall Filter
Auto Cal Parameter	Set the auto Calc parameters
Dynamic Range	Adjust the dynamic range of the spectrum
Spectrum Enhance	Adjust the brightness of the spectrum
DTrace Smooth	Adjust the smooth of the trace of the spectrum
Threshold	Adjust the threshold of the spectrum
Trace Area	Set the auto cal range
Layout	Choose from LR and UD
Auto Color	Blood flow velocity class: low, normal, high

Shortcut menu adjustment:

Shortcut (action)	Menu	Menu Function
SK1 (rotate)	Freq.	Adjust the probe launch frequency
SK2 (rotate)	Baseline	Adjust the baseline position
SK3 (rotate)	PRF	Adjust PRF value

SK4 (rotate)	Speed	Adjust scanning speed
SK5 (rotate)	SV	Adjust the size of the sample volume gate.
SK6 (rotate)	Audio	Adjust the audio of Doppler spectrum

Other adjustments:

1. Sampling line

In PW mode, move the trackball left and right to adjust the position of sampling line.

2. Sampling gate (Sampling volume)

Move the sample volume on the B-mode's Doppler area. The gate is positioned over a specific position within the vessel.

To move Doppler line position, move the Trackball left or right until positioned over the vessel.

To move sample volume gate position, move the Trackball up or down until positioned inside the vessel.

To size sample volume gate, press Enter-key to toggle Trackball function from sample volume gate positioning to sizing, then move the Trackball to change sample volume gate size.

3. Gain

Rotate PW knob and adjust gain of PWD mode when PW light is on. Gain will be increased if rotate clockwise; Gain will be decreased if rotate anti-clockwise. Adjustable range is from 0 to 255.

4.3.69 Continuous Wave Doppler (CWD) Imaging Optimization

Soft menu adjustment:

Menu	Function
Invert	Reverse the spectrum according to the baseline.
2D Map	Adjust the color of spectrum.
Spectrum Enhance	Adjust the brightness of the spectrum.
Dynamic	Adjust the dynamic range of the CW spectrum.
Audio	Adjust from 1% to 100%.
Wall Filter	Adjust wall filter.
Color Map	Change the color map.
QuickAngle	Turn on the QuickAngle function.
Stress Echo	It works with a phased array probe. Turn on Stress Echo function.
Display Format	1:1,1:2,2:1 can be chosen
Auto Color	Blood flow velocity class: low, normal, high

Shortcut menu adjustment:

Shortcut (action)	Menu	Menu Function
SK2 (rotate)	Baseline	Adjust the baseline position

SK3 (rotate)	PRF	Adjust PRF value
SK4 (rotate)	Speed	Adjust scanning speed
SK5 (rotate)	Angle	Adjust the Doppler angle

Other adjustments:

1. Sampling line

In PW mode, move the trackball left and right to adjust the position of sampling line.

2. Gain

Rotate PW knob and adjust gain of CWD mode when CW light is on. Gain will be increased if rotate clockwise; Gain will be decreased if rotate anti-clockwise. Adjustable range is from 0 to 255.

4.4 After Capturing the Image

4.4.1 Adding Comments

Comments means input the words or symbols on an ultrasound image to bring attention, notate or communicate information observed during the examination. You can add comments to: zoomed image, cine review image, real-time image, frozen image. You can input characters as comments; add custom comment bar; insert the pre-defined comments from the comment library.

Input characters

1. Press Comment key will display “|” on the screen, and then system will enter into the comment status.
2. Move the cursor to the position where need to comments.
3. Input characters at cursor position by keyboard then press ENTER-key to confirm.
4. Press Comment key to exit.

Add comment form comment library

1. Press Comment key will display “|” on the screen, and then system will enter into the comment status.
2. The comments from the comment library will display on the touch screen.
3. Press the comment that wants to be added.
4. Press Comment key to exit the comment status.

Adjust font size of comments

1. Rotate SK3 in comment status.
2. Move the cursor to the comment note. Press ENTER-key and adjust the font size of comment note. Press ENTER-key again to confirm revision.

Adjust position of comments

1. In comments status, moves the cursor to the comment which need to be changed and then press ENTER-key to select it.
2. Moves the comment to the position wanted to be and press ENTER-key to confirm.

4.4.2 Adding Body Mark

Operation:

1. Press **[Bodymark]** key to enter into the body mark status. The body marks will display on the touch screen.
2. Press the bodymark that want to be added, move trackball after adding the bodymark image and then adjust the position of probe. Rotate SK5 knob to adjust the probe direction. Press ENTER-key to confirm after finishing the adjustment.
3. Press UPDATE-key and move the Trackball to change the position of the body mark.
4. Press ENTER-key to confirm when adjustment is finished.
5. If you want to exit the body mark function, press **[Bodymark]** key

Delete comments, body marks and measure tracks

Press **[Clear]** key to delete all comments, arrows, body marks and measure tracks.

4.4.3 Saving Still Images

Save single image: Press **[STILL SAVE]** key to save the single image in the real time and frozen statue. The saved thumbnail images will be displayed under the image area for users to playback or post-process quickly.

Re-call single image: Press cursor key in re-call status, then the mouse will appear on the screen. Move the mouse to thumbnail images then click.

4.4.4 Saving Clips

Save forward cine files: Press **[CINE SAVE]** key in the real time to save forward cine files.

Save backward cine files: Press **[CINE SAVE]** key in frozen statue to save backward cine files.

The saved cine images will be displayed on the right of the image area for users to playback or post-process quickly.

Playback cine files: Press cursor key in recall status, then the mouse will appear on the screen. Move the mouse to the cine thumbnail image then click.




Fig.4-19




Fig.4-15

Play/Pause: Press this key to Play or Pause cine playback.

Play Speed: Rotate the SK1 knob to adjust the playback speed of cine.

Start Pos: Rotate the SK2 knob to set the frame of start position.

End Pos: Rotate the SK3 knob to set the frame of final position.

Reset Range: Press the SK4 to reset the range of cine playback, the default range of playback is 1-maximum frame of current cine.

Manual playback cine: Press cuosor key and back to recall status then manual playback through trackball.



NOTE: When you save the still images or cine images, if you do not establish a new patient, the system will establish a new ID according to the current date of the system

automatically and save the data and operation under the folder of this ID.

Quick transmission of image:

Image transmission: Press CURSOR key to activate the cursor and then choose the image which

you want to transfer or delete. Press  icon and then three icons will display in the thumbnail area.




Fig.4-16



Enter into the export interface.

Image Sharing

The system supports image sharing function.

Raw Data

This function allows user to make measurement and zoom after freezing the image.



Enter into the EasyView interface.



Delete the file.

4.4.5 Browse Images

Press **[EasyView]** key on the touchscreen to enter into the EasyView interface.



Fig.4-22

Info: press this key can check the information of current patient.

Report: press this key to enter into the report interface.

Close: click this key and choose to close information of patient.

Close all: close all current patients.

Continue/Edit Exam: continue to check the current patient and back to B mode if press this key. If current exam is 48 hours later than last exam, the system will not allow you to continue the exam. The function will be changed to "Edit".

New Study: create a new patient, then the system will automatically go to this new page.

Archive: Pop up Files Management interface. Multiple patients can be compared and browsed.

Cancel: Exit.

4.4.6 Fast Storage

Setting: Enter into System Settings. Press General Menu and choose General Sub Menu to do fast storage setting.



Fig.4-23

After confirming time, back to examination.

For example, choose 5 seconds in the fast storage setting. The system begins to save cine in the 5 seconds after pressing **[CINE SAVE]** key.

If you choose Manual in the fast storage setting, you should press **[CINE SAVE]** key for two times. First time is to start saving and the second time is to finish.

If you choose Export to removable media, it will save the images to the removable media.

4.4.7 Files Management

In this interface you can manage patient's files. Press **[EasyView]** key on the touch screen to enter into the EasyView interface. Then click "Archive" to enter into file management interface.




Fig.4-17

Search: Can search the patient files after enter a search term. Search will be more accurate when input more conditions.

Information: View the selected patient information.

Backup: Click this item will store the database of this patient to the portable storage medium

Recover: Click this item can recover the images or video from USB to the system

Send to: Select the content and click, can send files to other medium.

Delete: delete patient's files, images etc.

Select all: Select all patients.

Patients view: Under this view, one patient with multiple inspection data will be listed in column and also you can check each exam file in details.

Study view: This view can list exam types one by one. Different exam type of one patient will not be listed in one column.

New study: Create a new patient and recall into the patient information interface.

Continue/Edit Exam: Continue to check the current patient and back to B mode if press this key. If current exam is 24 hours later than last exam, the system will not allow you to continue the exam. The function will be changed to "Edit".

Review: Select patient and click it to enter into images browse.

Cancel: Click "Cancel" to exit.

4.4.8 Network Storage Function

Before you use this function, please refer to Section 7.7 to set network storage parameters.

Click **[EasyView]** button on the touch screen to enter into the Easy Review interface, select the



picture you want to send, click  to open the export interface, select the network drive and click "OK" to send.

After transmit into the target computer, open the shared folder, you can see the pictures you sent.

4.4.9 DICOM Function

The system support DICOM storage, DICOM print, MPPS, WORKLIST, DICOM SR storage.

1. Connect the machine into LAN before setting the DICOM.
2. Enter into DICOM Setting interface.
3. Press Utility key to enter into system setting interface. Press DICOM page to enter into DICOM interface.



NOTE: If the DICOM function is not open, the system setting interface will not display DICOM page. Please make sure the DICOM function is open before you use DICOM function.




Fig.4-18

Add: Press this icon to add a DICOM service link.

Modify: Select the DICOM link which need to be modified. Press this icon to modify the parameters inside.

Delete: Select the DICOM link which needs to be deleted. Press this icon to delete the DICOM link.

Verify: Select the DICOM link which need to be verified. Press this icon to verify the current DICOM link status.

Default: If there are several service links of same type, you can select one of them as the default one.

DICOM storage

- DICOM storage parameters setting.

The screenshot shows a software window titled "Add DICOM Setting". It contains several input fields and checkboxes. On the left, under the "Service" header, there are fields for "Service" (set to "Storage"), "AE Title", "Port" (set to "104"), and "Time Out" (set to "60000"). On the right, there are fields for "CharSet", "HOST IP", "Alias", and "Mode" (set to "Manual"). Below these fields are two checkboxes: "DICOM SR" and "De-Identification" (which has a sub-label "De-Identification" next to it). At the bottom right of the window are three buttons: "Clear", "OK", and "Cancel".

Fig.4-19

Service: Click the drop list to select storage function.

Charset: Click the drop list to select between ISO_IR 100, ISO 2022 IR 6 and ISO 2022 IR 100.

AE title: Name the server at will.

Host IP: Set the IP of the service host.

Port: Make the port of the work station the same as the port of the server.

Alias: The name of current link.

Time out: time delay.

Mode: batch, follow, manual selectable

Batch: all the saved images in current exam will be sent to the server in one time.

Follow: When one image is saved, the system will send this image to the server.

Manual: The user has to select the image by manual.

After all the settings are completed, please click OK to save current setting. Click clear to clear all the setting. Clicking cancel will not save the current setting.

- DICOM storage

Batch mode: In batch mode, when the patient finishes his exam, press End key to end the current exam. All the saved images will be saved as DCM format and sent to the server automatically.

Follow: In follow mode, during the exam, press **[STILL SAVE]** key to save the image. The current image will be saved as DCM format and sent to the server.

Manual: In manual mode, the user has to select the image by manual and do DICOM transfer.

Press **[EasyView]** key on the touch screen to enter into easyview interface. Select the images

need to be sent, press  key to send.

DICOM print

- DICOM print parameters setting.



Fig.4-20

Service: Click the drop list to select print function.

AE title: Name the server at will.

Host: Set the IP of the server.

Port: Make the port of the work station the same as the port of the server.

Alias: The name of current link.

Time out: time delay.

Mode: batch, follow, manual selectable.

Batch: all the saved images in current exam will be sent to the server and printed automatically in one time.

Follow: When one image is saved, the system will send this image to the server and print it.

Manual: The user has to select the image by manual and print it.

After all the settings are completed, please click OK to save current setting. Click clear to clear all the setting. Clicking cancel will not save the current setting.

- DICOM print

Batch mode: In batch mode, when the patient finishes his exam, press End key to end the current exam. All the saved images will be saved as DCM format and sent to the server automatically to be printed.

Follow: In follow mode, during the exam, press save key to save the image. The current image will be saved as DCM format and sent to the server to be printed.

Manual: In manual mode, the user has to select the image by manual and do DICOM transfer.

Press browse key to enter into review interface. Select the images need to be sent, click icon to print.



DICOM Worklist

- DICOM Worklist parameters setting.

Fig.4-28

Service: Click the drop list to select worklist function.

Charset: Click the drop list to select between ISO_IR 100, ISO 2022 IR 6 and ISO 2022 IR 100.

AE title: Name the server at will.

Host: Set the IP of the server.

Port: Make the port of the work station the same as the port of the server.

Alias: The name of current link.

Time out: time delay.

For “Update Method” parameter group, the worklist update can only started by user request for “Only on user request”, or user can set N(1~60) minutes to update worklist automatically for “On Startup and Every X minutes”.

For “Scheduled Station AE Title” parameter group, the schedule station AE title, which is one of a matching key, will be set to blank for “Any”, “LocalAE” for “This System” or user customized string for “Another”. And only the matched items will be queried for each worklist update.

For “Start Date” parameter group, it is the matching key for “Scheduled Procedure Step Start Date”, and there are five “Start Date” types to select.

- DICOM Worklist

CHISON Patient Information

Basic Information

ID: 2022-07-23-163813

First Name:

Middle Name:

Last Name:

Birthday:

Age: 0 Y 0 M 0 D

Gender:

Detail Information

Study Type: General

Height: cm Weight: kg BSA: m²

Study Description: Accession #:

Primary Indications: CPT4 Code:

Secondary Indications: CPT4 Description:

Diag. Physician: Ref. Physician: Sonographer:

Comment:

Search In


Fig.4-29

Press **[Patient]** button on the touch screen, and press “WorkList”.

Press “Search” to start a worklist update by the matching key specified in Worklist Configuration, all the matched items will be queried in worklist table. User also can add additional for matching keys, which are “PatientID”, “PatientName”, “Accession#” and “ProcedureId”, for a more precise query.

Select one matched item and press “Apply” for the current exam application. At this moment, a MPPS service will be started which will be introduced in MPPS section.

DICOM MPPS



The image shows a software dialog box titled "Add DICOM Setting". It contains several input fields for configuring DICOM settings. On the left, under a "Service" header, there is a dropdown menu for "Service" (currently showing "MPPS"), a text field for "AE Title", a text field for "Port" (containing "108"), and a text field for "Time Out" (containing "60000"). On the right, there is a text field for "HOST IP" and a text field for "Alias". At the bottom right of the dialog are three buttons: "Clear", "OK", and "Cancel".

Fig.4-30

Service: Click the drop list to select MPPS function.

AE title: Name the server at will.

Host: Set the IP of the server.

Port: Make the port of the work station the same as the port of the server.

Alias: The name of current link.

Time out: time delay.

Applying the selected worklist item will result in automated creation of an MPPS instance.

Chapter 5 General Measurements

Introduction

Measurements and calculations derived from ultrasound images are intended to supplement other clinical procedures available to the physician. The accuracy of measurements is not only determined by system accuracy, but also by the use of proper medical protocols by the operator. When appropriate, make sure to note any protocol associated with a particular measurement or calculation.

CAUTION:

Please select the most appropriate ultrasound images, measurement tools and measurement methods for measurements according to your diagnosis needs. The final measurement results must be determined and verified by a physician. Measurement accuracies are affected by many non-technical factors, for example operator's experience, patient's status. Please do not only use the ultrasound measurement results as the sole basis for diagnosis, please always use other clinical information to do integrated diagnostics.

Overview

This section provides information about taking measurements and describing calculations available in each mode. It includes the following topics:

- List of generic measurements.
- Mode Measurements: Step-by-step instructions for taking specific measurements, organized by mode.
- Measurement results display and erasing.

5.1 Key for Measurement

◆ 【Trackball】

Trackball is used to move the cursor, main functions are as follows:

Before starting a measurement, use the trackball to choose the menu options;

After starting a measurement, move the trackball to move the cursor, during the measurement, the cursor should not be moved out image area;

During the Ellipse method measurement, use trackball to change the length of short axis.

◆ 【ENTER】

During the measurement, the functions of 【ENTER】 key are as follows:

When cursor is on the menu, press the key to choose the options and start the measurement.

During the measurement, press the key to anchor the start point and end point.

◆ 【CLEAR】

Main functions are as follows:

In frozen status, press the CLEAR-key, delete all the measurement results, comments and traces.

◆ 【UPDATE】

During the measurement, 【UPDATE】 key is used to switch the start point and end point, long axis and short axis when the measurement is not finished.

During the distance measurement, press the 【ENTER】 to fix the start point, when the end point is not fixed, press the 【UPDATE】 key to switch the start point and end point.

During the Ellipse measurement, when fix the long axis, but the short axis is not fixed, press the 【UPDATE】 key to switch the long axis and short one.

5.2 Fast measurement

The system use 【MEASURE】 button to make fast measurement. Users can choose the measurement items on the touch screen.

5.2.1 Enter fast measurement

Press 【MEASURE】 key to enter fast measurement.

5.2.2 Exit fast measurement

During measurement, press the 【MEASURE】 key for a second time or press SK6 to exit.

5.2.3 Fast measurement in B mode

Distance

1. In B mode fast measurement status, press the **[Dis]** on the touch screen to bring up the first white plus sign (“+”) cursor.

2. Use the Trackball and the **【ENTER】** key to anchor the start point of the desired distance to be measured.
3. Move the Trackball and a second white plus sign cursor will automatically appear. Move the second cursor to the specified position. With the movement of cursor the system will update the measurement distance in real time in Result Window on the right side of the screen.
4. Press **【UPDATE】** key can change the activated point, and fit the other point of the line.
5. Press the **【ENTER】** key to fix the second cursor, the first measurement results appear on the screen.
6. When two points have been defined, a white plus sign cursor of next distance pair will pop out. Repeat the same steps to create other distance pairs.

Trace

1. In B mode fast measurement status, press the **【Trace】** on the touch screen to bring up the trace start cursor plus sign (“+”).
2. Use the trackball to move the cursor, press **【ENTER】** key to set the start point.
3. Use the Trackball to trace the perimeter of the object to be measured, and press the **【ENTER】** key to set the end point. The end point and the starting point of the trace will be automatically connected with a straight line. The area and the Circumference of the shape will be calculated. With the movement of the cursor, the result will appear on the right side of the screen, and will change along with the moving of the trace.
4. To get the multiple data, repeat the same steps to create other distance pairs.

Area Auto

1. In B mode fast measurement status, press the **【Area Auto】** on the touch screen to bring up the trace start cursor plus sign (“+”).
2. Move the Trackball and press **【ENTER】** key to set start point.
3. The second point marker is displayed by a line connecting the first and second points. Use the Trackball to lengthen the line in order to change the diameter and press the **【ENTER】** key to fix the length.
4. Use the Trackball to change the length of another axis. When fix the long axis, but the short axis is not fixed, press **【UPDATE】** key to switch the long axis and short one. Press the **【ENTER】** key to fix the length. The total area and circumference will be displayed in the result window.
5. To get the multiple data, repeat the same steps to create other distance pairs.

Volume - 1Dis1Ellip

1. In B mode fast measurement status, press the **【1Dis1Ellip】** in the dropdown list of **【Volume】** on

- the touch screen to bring up the trace start cursor plus sign (“+”).
2. To measure distance, refer to the steps from 2~5 in the Distance measurement above; to measure ellipse, refer to the steps from 2~ 4 in the Area Auto measurement above.
 3. After finishing the measurement of 1 line and 1 ellipse, the result will be displayed in the result area.

Volume - 3Dis

1. In B mode fast measurement status, press the **[3Dis]** in the dropdown list of **[Volume]** on the touch screen to bring up the trace start cursor plus sign (“+”).
2. To measure distance, refer to the steps from 2~ 5 in the Distance measurement above.
3. After finishing the measurement of 3 lines, the result will be displayed in the result area.

Angle

1. In B mode fast measurement status, press the **[Angle]** on the touch screen to bring up the first white plus sign (“+”) cursor.
2. Line1 measurement starts. Refer to the steps from 2~ 5 in the Distance measurement above.
3. Line2 measurement starts. Refer to the steps from 2~ 5 in the Distance measurement above.
4. After the measurement, the result of the angle will be displayed on the screen.

5.2.4 Fast measurement in PW mode

Peak

1. In PW mode fast measurement status, press the **[Peak]** on the touch screen to bring up the sample marker.
2. Move the marker to the measurement start point with the Trackball, press **【ENTER】** key, velocity and pressure of the current point will appear on the screen automatically.
3. Go on to measure Vd, after getting the result, the system will calculate S/D, RI, Tiime automatically.

Auto

1. In PW mode fast measurement status, press the **[Auto]** on the touch screen, to enter automatic trace measurement. The system will finish the trace of spectrum automatically. The icon “+” will appear on the screen.
2. Move the trackball to choose a start point of one cycle, press the **【ENTER】** key to confirm.
3. The second icon “+” will appear on the screen automatically, move the trackball to the end point of current cycle, and press the **【ENTER】** key to set.
4. The measurement result and the other calculated parameters will appear on the screen

automatically.

5. If the spectrum automatic envelop is not accurate, press SK3 and reselect cycle to correct.

Manual

1. In PW mode fast measurement status, press the the **[Manual]** on the touch screen to enter manual trace measurement.
2. Move the trackball to choose a start point of one cycle, press the **【ENTER】** key to confirm.
3. Move the trackball to trace the spectrum, press the **【ENTER】** key to finish the trace.
4. The measurement results and the other calculated results will appear on the screen automatically.

Acceleration

Measure the slope of the flow rate between two points on the Doppler image.

1. In PW mode fast measurement status, press the **[Acceleration]** on the touch screen to bring up the sample marker.
2. Move the trackball to the starting point of Doppler spectrum measurement and press **【ENTER】** key to confirm.
3. A second icon “+” displays. Move the trackball to the ending point of the measurement and press **【ENTER】** key to confirm.
4. The Vel, time and acceleration values are automatically displayed on the screen.

Volume Flow

1. In PW mode fast measurement status, press the **[Volume Flow]** on the touch screen, an icon “+” displays in B image area.
2. Move the trackball and press **【ENTER】** key to set the start.
3. A second icon “+” appears and move the trackball and press **【ENTER】** key to set the end.
4. A dotted vertical line with an icon “+” shows in PW image area.
5. Move the trackball to move the line and press **【ENTER】** key to fix it.
6. Another dotted vertical line with an icon “+” will appear.
7. Move the trackball and press the **【ENTER】** key to fix the line.
8. The measurement result will appear on the screen automatically.

5.2.5 Fast measurement in M mode

MDistance

1. In M mode fast measurement status, press the **[MDistance]** on the touch screen, an icon “+” on a dotted line will appear on the screen.
2. Move the trackball left or right to adjust the line position. Move the trackball up or down to adjust the icon position. Press **【ENTER】** key to fix the line and icon.
3. One dotted line and the second icon “+” will appear on the screen.
4. Move the icon by moving the trackball to the end point in the dotted line vertically, press the **【ENTER】** key to confirm.
5. The measurement result will appear on the screen automatically.

MTime

1. In M mode fast measurement status, press the **[MTime]** on the touch screen, an icon “+” on a dotted line will appear on the screen.
2. Move the trackball to move the line, press **【ENTER】** key to fix it.
3. One dotted line and a second icon on a line will appear.
4. Move the trackball and press the **【ENTER】** key to fix the line.
5. The measurement result will appear on the screen automatically.

MSpeed

1. In M Cine mode, press the **[MSpeed]** on the touch screen, an icon “+” on a dotted line will appear on the screen.
6. Move the trackball left or right to adjust the line position. Move the trackball up or down to adjust the icon position. Press **【ENTER】** key to fix the line and icon.
2. One dotted line and the second icon “+” will appear on the screen.
7. Move the trackball left or right to adjust the line position. Move the trackball up or down to adjust the icon position. Press **【ENTER】** key to fix the line and icon.
3. The measurement result will appear on the screen automatically.

HR

Heart rate is used to calculate the number of heart beats per minute from cardiac image.

1. In M mode fast measurement status, press the **[HR]** on the touch screen, it will display a dotted vertical line with “+” icon in the M image area.
2. The remaining measurement steps are the same as those from 2~4 in MTime.
3. After the above measurement, the calculated heart rate result is displayed in the measurement result area.

5.3 Measurement and Calculation

There are corresponding measurement menu in different modes. Press **【CAL】** key to recall the measurement menu.

Press  icon to switch between different measurement packages of different modes. Press **【CAL】** key again to close the measurement menu.



Click on the touch screen to select the measurement item on the measurement menu and move the trackball to do the measurement in the image.

After finishing the measurement, press **[Report]** key on the touch screen to generate the report template to view the measurement results or print the report.

5.3.1 Measurement in B mode

Press **【CAL】** Key to enter measurement package menu. Press  icon to select measurement package.

5.3.1.1 General Measurement in B mode

1. Distance

- 1) Move the trackball to move the cursor to the “distance” item. Press ENTER-key to select it and a mark “+” will appear on the screen
- 2) Move the trackball to anchor the first mark and press ENTER-key to set it.
- 3) The second mark will appear on the screen. Move the trackball to place the mark to the aimed place. The value of the measurement will be displayed on the right side of the screen and will be changed along with the trace of the measurement.
- 4) Press ENTER-key to fix the second mark and the measurement results will be fixed and displayed on the screen.
- 5) To get the multiple data, repeat the same steps to create other distance pairs.
- 6) Press **【CAL】** key to exit.

2. Area(Ellipse)

- 1) Rotate the trackball to move the cursor to the “Area (ellipse)” item. Press ENTER-key to select it. At this time, a mark “+” will appear on the screen.
- 2) Rotate the trackball to anchor the first mark and press ENTER-key to fix it.
- 3) The second mark will appear on the screen. Rotate the trackball to change the length of the long

axis of the ellipse. Press ENTER-key to confirm the length.

- 4) Rotate the trackball to change the length of another axis of the ellipse and press ENTER-key to fix it. The value of the area and perimeter will be displayed on the right side of the screen.
- 5) To get the multiple data, repeat the same steps to create other distance pairs.
- 6) Press 【CAL】 key to exit.

3. Area(Trace)

- 1) Rotate the trackball to move the cursor to the “Area (trace)” item. Press ENTER-key to select it. At this time, a mark “+” will appear on the screen.
- 2) Move the trackball to move the mark to the start point of the measurement. Press ENTER-key to confirm.
- 3) Rotate the trackball to trace the edge of the aimed object and press ENTER-key to fix the end point. The second mark will appear on the screen. Rotate the trackball to place the mark to the aimed place. The value of the measurement will be displayed on the right side of the screen and will be changed along with the moving of the measurement.
- 4) To get the multiple data, repeat the same steps to create other distance pairs.
- 5) Press 【CAL】 key to exit.

4. Volume(1distance)

Calculate the volume of the object by one line.

$$\text{Volume} = \text{distance} \times \text{distance} \times \text{distance} \times 3.1415926/6$$

5. Volume (1ellipse)

Through 1 ellipse measurement, calculate the volume of the object.

$$\text{Volume} = \text{axis A} \times \text{Axis B} \times \text{axis B} \times 3.1415926/6$$

6. Volume(2distance)

Through 2 distance measurement, calculate the volume object.

$$\text{Volume} = \text{distance1} \times \text{distance2} \times \text{distance2} \times 3.1415926/6 \quad (\text{distance1} > \text{distance2})$$

$$\text{Volume} = \text{distance2} \times \text{distance1} \times \text{distance1} \times 3.1415926/6 \quad (\text{distance2} > \text{distance1})$$

7. Volume(3 distance)

Through 3 distance measurement, calculate the volume of object

$$\text{Volume} = \text{distance1} \times \text{distance2} \times \text{distance3} \times 3.1415926/6$$

8. Volume(1distance1ellipse)

Through 1distance measurement and 1 ellipse measurement, calculate the volume of object.

$$\text{Volume} = \text{distance} \times \text{axis A} \times \text{axis B} \times 3.1415926/6$$

9. Volume(2 ellipse)

Through 2 ellipse measurement, calculate the volume of the object.

$$\text{Volume} = d1 \times d2 \times d4 \times 3.1415926 / 6$$

d1 and d2 are axis of ellipse1; d3 and d4 are axis of ellipse 2: $|d2 - d3| \leq |d1 - d4|$, $d2 > d3$

10. Ratio

Through 2 distance measurement, calculate the ratio of the 2 distance

$$\text{Ratio} = \text{distance1} / \text{distance2}$$

11. Angle

- 1) Rotate the trackball to move the cursor to the "Angle" item. Press ENTER-key to select it and d1 line will appear on the screen.
- 2) Rotate the trackball to move d1 and rotate the Angle button to adjust the direction of the line, then press ENTER-key to fix it.
- 3) Fix other 2 lines of d2 and d3 as step2).
- 4) After finishing the measurement, the result of the angles named α and β will be displayed on the left side of the screen. α is the angle between line d1 and line d2. β is the angle between line D1 and D3.
- 5) Press 【CAL】 key to exit.

12. Strain Ratio

Calculate the ratio between Strain A and Strain B.

13. B-Histogram

Histogram is used to calculate the gray distribution of the ultrasound echo signals within a specified area. Elliptic histogram measurement is to measure the desired area by an enclosed ellipse, square or tracing line. The result is shown in the form of histogram.

The horizontal axis represents the gray scale of the image ranging from 0 to 250.

The vertical axis represents the distribution ratio of each gray scale. The value shown on the top of vertical axis represents the percentage of the maximally distributed gray in the whole gray distribution.

Histogram can only be measured on frozen images.

B-Histogram (Squar)

1. Click this measurement item to activate the measurement.
2. Press [Enter] to confirm the one vertex of the rectangle.
3. Move the trackball to move the cursor to fix the vertex of the diagonal of the rectangle. And then press [Enter] again to confirm the measurement area.
4. The diagram for histogram of Square displays.

B-Histogram (Circle)

Measure the histogram by ellipse: The method is the same as ellipse measurement.

B-Histogram (Trace)

14. Measure the histogram by trace: The method is the same as trace measurement.

B-Histogram (Spline)


Measure the histogram by spline.

1. Click this measurement item to activate the measurement.
2. Move the cursor to the target area.
3. Press [Enter] to confirm the first reference point of the curved line.
4. Move along the target area to set the second reference point.
5. Move the cursor again to show the closed spline curve defined by the first and second reference points and the moving cursor points.
6. Go on to move along the target area to set the reference points to make the shape of the closed area approaches to the target area as much as possible. A maximum of 12 reference points can be set.
7. Press [Enter] to finish the measurement and the diagram for histogram of Spline displays.

Lung Assistant

SonoPleura

Select lung preset and scan Pleura under B mode. Click [SonoPleura] on the touch screen to activate SonoPleura.




Pneumonia Measurement - Pleura Line

Select lung preset and scan Pleura under B mode, and freeze the image. Select Pleura Line under Lung Menu to show pleura line automatically.



Pneumonia Measurement – Auto B-Line

Click [Auto B-Line] under Lung menu to show B-Line automatically.



B-Lines Edit: Activate this button to edit B-Lines.

B-Lines Delete: Delete B-Lines.

B-Lines: Select the B-line which is activated.

B-Lines Add: Add new B-Lines by manual.

Pneumonia Report

Press report button to get into report page, and press Comparison Data, get into comparison page.
Press OK to Show comparison sheet in report.

5.3.1.2 Cardiology Measurement in B mode

1. Auto EF

EF (Ejection Fraction): It is the amount of blood pumped out from the heart. Generally speaking, EF is the percentage of blood pumped out from the left ventricle when the heart is systolic.

Auto EF is that the system identifies end-systolic and end-diastolic sections of the heart and calculates EF value by Simpson. It works in cardiac preset under phased array probe. Turn on ECG to obtain images of four-chamber, two-chamber, or long-axis sections of the cardiac, and freeze the images. Or recall the cardiac movie file with ECG. In the cardiac measurement package, Auto EF is available. During the measurement, user can adjust the parameters and then confirm the measurement value.

2. Free Curved M

Free curved M is used to evaluate the myocardium under the heart. It works with a phased probe.

3. Teichholz

Calculation formula for volume

$$EDV = \frac{7 \times LVIDd^3}{2.4 + LVIDd}$$

$$ESV = \frac{7 \times LVIDs^3}{2.4 + LVIDs}$$

<<Items to be measured>>

Meas. item	Description	Meas. Tool
LVIDd [cm]	Left ventricular internal diameter at end diastole	Line
LVIDs [cm]	Left ventricular internal diameter at end systole	Line

<< Items to be calculated >>

Calc. item	Description	Calc. formula
EDV [ml]	End-diastolic left ventricular Volume	Please refer to <<Calculation formula for volume>>
ESV [ml]	End-systolic left ventricular Volume	Please refer to <<Calculation formula for volume>>
SV [ml]	Stroke Volume	SV=EDV – ESV
CO [l/min]	Cardiac Output	CO=SV*HR/1000
EF [%]	Ejection Fraction	EF=SV/EDV
FS [%]	Fractional Shortening	FS=(LVIDd-LVIDs)/LVIDd
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

4. Simpson SP

This method calculates the left ventricular volumes by using the 2D mode long-axis image. When the left ventricle is traced and its long axis is specified on a 2D mode cross-sectional image, this method allows the system to automatically divide the long axis into 20 segments to calculate the volume of the left ventricle as the 2D mode cross-sectional image is rotated.

Calculation formula for volume

$$EDV = \pi \times \frac{LVLd}{20} \times \sum_{i=1}^{20} r_i^2$$

$$ESV = \pi \times \frac{LVLs}{20} \times \sum_{i=1}^{20} r_i^2$$

r_i : Radius of the i -th circle

LVLd: Left ventricular long-axis diameter at end diastole

LVLs: Left ventricular long-axis diameter at end systole

<<Items to be measured>>

Meas. item	Description	Meas. Tool
EDV(A4C) [ml]	End-diastolic volume (A4C)	Measurement (trace)
ESV(A4C) [ml]	End-systolic volume (A4C)	Measurement (trace)

<< Items to be calculated >>

Calc. item	Description	Calc. formula
EDV [ml]	End-diastolic left ventricular Volume	Please refer to <<Calculation formula for volume>>
ESV [ml]	End-systolic left ventricular Volume	Please refer to <<Calculation formula for volume>>
SV [ml]	Stroke Volume	SV=EDV – ESV
CO [l/min]	Cardiac Output	CO=SV*HR/1000
EF [%]	Ejection Fraction	EF=SV/EDV
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

5. Simpson Biplane

This method allows the volume of the left ventricle to be calculated by repeating “Simpson SP method” on two 2D mode cross-sectional images (two-chamber and four-chamber cross-sectional images).

Calculation is also possible for only two-chamber cross-sectional images or only four-chamber cross-sectional images.

Calculation formula for volume

$$EDV (A2C) = \pi \times \frac{LVLd_2}{20} \times \sum_{i=1}^{20} r_{2i}^2$$

$$ESV (A2C) = \pi \times \frac{LVLS_2}{20} \times \sum_{i=1}^{20} r_{2i}^2$$

r_{2i}: Radius of i-th ellipse for 2CH

LVLd₂: Length of the left-ventricular long axis at end diastole for 2CH.

LVLS₂: Length of the left-ventricular long axis at end systole for 2CH .

$$EDV (A4C) = \pi \times \frac{LVLd_4}{20} \times \sum_{i=1}^{20} r_{4i}^2$$

$$ESV (A4C) = \pi \times \frac{LVLS_4}{20} \times \sum_{i=1}^{20} r_{4i}^2$$

r_{4i}: Radius of i-th ellipse for 4CH

LVLd4: Length of the left-ventricular long axis at end diastole for 4CH.

LVLs4: Length of the left-ventricular long axis at end systole for 4CH .

$$EDV = \pi \times \frac{\text{MAX}\{LVLd_2, LVLd_4\}}{20} \times \sum_{i=1}^{20} (r_{2i} \times r_{4i})$$

$$ESV = \pi \times \frac{\text{MAX}\{LVLs_2, LVLs_4\}}{20} \times \sum_{i=1}^{20} (r_{2i} \times r_{4i})$$

<<Items to be measured>>

Meas. item	Description	Meas. Tool
EDV(A2C) [ml]	End-diastolic volume (A2C)	Measurement (trace)
ESV(A2C) [ml]	End-systolic volume (A2C)	Measurement (trace)
EDV(A4C) [ml]	End-diastolic volume (A4C)	Measurement (trace)
ESV(A4C) [ml]	End-systolic volume (A4C)	Measurement (trace)

<< Items to be calculated >>

Calc. item	Description	Calc. formula
EDV[ml]	End-diastolic left ventricular volume	Please refer to <<Calculation formula for volume>>
ESV[ml]	End-systolic left ventricular volume	Please refer to <<Calculation formula for volume>>
SV [ml]	Stroke Volume	SV=EDV – ESV
CO [l/min]	Cardiac Output	CO=SV*HR/1000
EF [%]	Ejection Fraction	EF=SV/EDV
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

6. Modify Simpson

This method calculates the left ventricular using the 2-D mode long-axis image, the short-axis image at the level of the mitral valve, and the short-axis image at the level of the papillary muscle.

Calculation formula for volume:

$$EDV = \frac{LVLd}{9} \times (4 \times LVAMd + 2 \times LVAPd + \sqrt{LVAMd \times LVAPd})$$

$$ESV = \frac{LVLs}{9} \times (4 \times LVAMs + 2 \times LVAPs + \sqrt{LVAMs \times LVAPs})$$

<<Items to be measured>>

Meas. item	Description	Meas. Tool
------------	-------------	------------

LVLd[cm]	Left ventricular long-axis length at end diastole	Refer to "Distance Measurement (Distance)"
LVLs[cm]	Left ventricular long-axis length at end systole	Refer to "Distance Measurement (Distance)"
LVAMd[cm ²]	Left ventricular short-axis area at the level of the mitral valve at end diastole	Refer to "Area/Circumference Measurement (Area)"
LVAMs [cm ²]	Left ventricular short-axis area at the level of the mitral valve at end systole	Refer to "Area/Circumference Measurement (Area)"
LVAPd [cm ²]	Left ventricular long-axis area at the level of the papillary muscle at end diastole	Refer to "Area/Circumference Measurement (Area)".
LVAPs[cm ²]	Left ventricular long-axis area at the level of the papillary muscle at end systole	Refer to "Area/Circumference Measurement (Area)".

<<Items to be calculated>>

Calc. item	Description	Calc. formula
EDV [ml]	End-diastolic left ventricular volume	Refer to << Calculation formula for volume>>
ESV[ml]	End-systolic left ventricular volume	Refer to << Calculation formula for volume>>
SV [ml]	Stroke volume	SV=EDV – ESV
CO [L/min]	Cardiac output	CO=SV x HR /1000
EF	Ejection fraction	EF=SV/ EDV
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

7. Cube

This method calculates the left ventricular volume by approximating the given region to a cube.

Calculation formula for volume:

$$EDV = LVIDd^3$$

$$ESV = LVIDs^3$$

<< Items to be measured>>

Meas. item	Description	Meas. Tool
IVSd[cm]	Interventricular septal thickness at end diastole	Line
LVIDd[cm]	Left ventricular short-axis diameter at end diastole	Line
LVPWd[cm]	Left ventricular posterior wall thickness at end diastole	Line
IVSs[cm]	Interventricular septal thickness at end systole	Line
LVIDs[cm]	Left ventricular short-axis diameter at end systole	Line

LVPWs[cm]	Left ventricular posterior wall thickness at end systole	Line
-----------	--	------

<<Items to be calculated>>

Calc. item	Description	Calc. formula
EDV[ml]	End-diastolic left ventricular volume	Refer to << Calculation formula for volume>>
ESV[ml]	End-systolic left ventricular volume	Refer to << Calculation formula for volume>>
SV [ml]	Stroke volume	SV=EDV – ESV
CO	Cardiac output [L/min]	CO=SV x HR /1000
EF	Ejection fraction	EF=SV/ EDV
FS	Fractional shortening	FS=(LVIDd-LVIDs)/LVIDd
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

8. Bullet Volume

This method calculates the left ventricular volume using the 2D mode long-axis image and the short-axis image at the level of the mitral valve.

Calculation formula for volume:

$$EDV = \frac{5}{6} \times LVLd \times LVAMd$$

$$ESV = \frac{5}{6} \times LVLs \times LVAMs$$

<<Items to be measured>>

Meas. item	Description	Meas. Tool
LVLd[cm]	Left ventricular long-axis length at end diastole.	Line
LVLs[cm]	Left ventricular long-axis length at end systole.	Line
LVAMd[cm]	Left ventricular short-axis area at the level of the mitral valve at end diastole.	Area
LVAMs[cm]	Left ventricular short-axis area at the level of the mitral valve at end systole.	Area

<<Items to be calculated>>

Calc. item	Description	Calc. formula
EDV[ml]	End-diastolic volume	Refer to << Calculation formula for volume>>
ESV[ml]	End-systolic volume	Refer to << Calculation formula for volume>>
SV[ml]	Stroke Volume	SV=EDV – ESV
CO[L/min]	Cardiac output	CO=SV * HR /1000
EF	Ejection Fraction	EF=SV/EDV

SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

9. Gibson

Calculation formula for volume:

$$EDV = \frac{\pi}{6} \times (0.98 \times LVIDd + 0.59) \times LVIDd^2$$

$$ESV = \frac{\pi}{6} \times (0.98 \times LVIDs + 0.59) \times LVIDs^2$$

<<Items to be measured>>

Meas. item	Description	Meas. Tool
LVLd[cm]	Left ventricular long-axis length at end diastole.	Line
LVLs[cm]	Left ventricular long-axis length at end systole.	Line
LVAmd[cm]	Left ventricular short-axis area at the level of the mitral valve at end diastole.	Area
LVAms[cm]	Left ventricular short-axis area at the level of the mitral valve at end systole.	Area

<<Items to be calculated>>

Calc. item	Description	Calc. formula
EDV[ml]	End-diastolic volume	Refer to << Calculation formula for volume>>
ESV [ml]	End-systolic volume	Refer to << Calculation formula for volume>>
SV[ml]	Stroke Volume	SV=EDV – ESV
CO[L/min]	Cardiac output	CO=SV * HR /1000
EF	Ejection Fraction	EF=SV/EDV
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

10. Mitral Valve

<<Items to be measured>>

Meas. item	Description	Meas. Tool
MV Diam[cm]	Mitral valve diameter	Line
MV Area[cm ²]	Mitral valve area	Ellipse
LAD[cm]	Left ventricular diameter	Line
AOD[cm]	Aortic dimension	Line

<<Items to be calculate>>

Calc. item	Description	Calc. formula
LAD/AOD	LAD/AOD	AOD/AOD

11. Aortic Valve

<<Items to be measured>>

Meas. item	Description	Meas. Tool
AV Diam[cm]	Aortic valve diameter	Line
AV Area[cm ²]	Aortic valve area	Ellipse

12. Pulmonary Valve

<<Items to be measured>>

Meas. item	Description	Meas. Tool
PV Diam[cm]	Pulmonary valve diameter	Line
PV Area[cm ²]	Pulmonary valve area	Ellipse

13. Tricuspid Valve

<<Items to be measured>>

Meas. item	Description	Meas. Tool
TV Diam[cm]	Tricuspid valve diameter	Line
TV Area[cm ²]	Tricuspid valve area	Ellipse

14. LVOT

<<Items to be measured>>

Meas. item	Description	Meas. Tool
LVOT Diam[cm]	Left ventricular outflow tract diameter	Line
LVOT Area[cm ²]	Left ventricular outflow tract area	Ellipse

15. RVOT

<<Items to be measured>>

Meas. item	Description	Meas. Tool
RVOT Diam[cm]	Right ventricular outflow tract diameter	Line
RVOT Area[cm ²]	Right ventricular outflow tract area	Ellipse

16. PISA

PISA MR

Calculation formulaMR Flow Rate(ml/s)=2πR Flow²* MR Als.Vel

<<Items to be measured>>

Meas. item	Description	Meas. Tool
MR Rad[cm]	Radius of mitral stenosis	Line
MR Als Vel [cm/s]	Input manually	

17. LV Mass

Cube

<<Items to be measured>>

Meas. item	Description	Meas. Tool
IVSd[cm]	Interventricular septal thickness at end diastole	Line
LVIDd[cm]	Left ventricular short-axis diameter at end diastole	Line
LVPWd[cm]	Left ventricular posterior wall thickness at end diastole	Line

<<Items to be calculated>>

Calc. item	Description	Calc. formula
LV Mass(g)	Left ventricular mass	$1.04 * [(LVPWd + IVSd + LVIDd)^3 - LVIDd^3] - 13.6$
LV Mass index	Left ventricular mass index	LV Mass/BSA

A-L (Area-Length)

<<Items to be measured>>

Meas. item	Description	Meas. Tool
LVAd sax Epi[cm ²]	Left ventricular epicardial short axis area at the level of the papillary muscle tips at end diastole	Area
LVAd sax Endo[cm ²]	Left ventricular endocardial short axis area at the level of the papillary muscle tips at end diastole	Area
LVLd apical	Left ventricular long axis length at end diastole	Line

<<Items to be calculated>>

Calc. item	Description	Calc. formula
LV Mass(g)	Left ventricular mass	$1.05 \times \frac{5}{6} (LVAd \text{ sax Epi} \times LVLd \text{ apical} + \sqrt{\frac{LVAd \text{ sax Epi}}{\pi}} - \sqrt{\frac{LVAd \text{ sax Endo}}{\pi}}) - LVAd \text{ sax Endo} \times LVLd \text{ apical}$
LV Mass index	Left ventricular mass index	LV Mass/BSA

T-E (Area-Length)

<<Items to be measured>>

Meas. item	Description	Meas. Tool
------------	-------------	------------

LVAAd sax Epi[cm ²]	Left ventricular epicardial short axis area at the level of the papillary muscle tips at end diastole	Area
LVAAd sax Endo[cm ²]	Left ventricular endocardial short axis area at the level of the papillary muscle tips at end diastole	Area
a[cm]	From the widest long axis radius to the semi-major axis tips	Line
d[cm]	From the widest long axis radius to the semi-major axis tips of mitral valve	Line

<<Items to be calculated>>

Calc. item	Description	Calc. formula
LV Mass(g)	Left ventricular mass	$10.5\pi \times \left\{ (b+t)^2 \times \left[\frac{2(a+t)}{3} + d - \frac{d^3}{3(b+t)^2} \right] - b^2 \times \left(\frac{2a}{3} + d - \frac{d^3}{3a^2} \right) \right\}$
LV Mass index	Left ventricular mass index	LV Mass/BSA

$$t(\text{cm}) = \sqrt{\frac{\text{LVAAd sax Epi}}{\pi}} - \sqrt{\frac{\text{LVAAd sax Endo}}{\pi}}$$

$$b(\text{cm}) = \sqrt{\frac{\text{LVAAd sax Endo}}{\pi}}$$

18. Qp/Qs

Qp: Pulmonary flow

Qs: Systemic flow

<<Items to be measured>>

Meas. item	Description	Meas. Tool
AV Diam	Aortic valve diameter	Area
PV Diam	Pulmonary valve diameter	Area
AV VTI	Aortic valve Velocity Time Integral	Trace- PW Mode
PV VTI	Pulmonary valve Velocity Time Integral	Trace- PW Mode

<<Items to be calculated>>

Meas. item	Description	Calc. formula
AV HR	Aortic valve Heart Rate	HR=VTI/Mean
AV SV(ml)	Aortic valve Stroke Volume	SV=0.785*AV Diam * AV Diam*abs (VTI
PV HR	Pulmonary valve Heart Rate	HR=VTI/Mean
PV SV	Pulmonary valve Stroke Volume	SV=0.785*AV Diam * AV Diam*abs (VTI
PV CO	Pulmonary valve Cardiac output	CO=SV*HR/1000
Qp/Qs	The ratio of Pulmonary flow volume	PV CO/ AV CO

	and Systemic flow volume	
Qp-Qs	The balance of Pulmonary flow volume and Systemic flow volume	PV CO - AV CO

19. RV/LV

<<Items to be measured>>

Meas. item	Description	Meas. Tool
RVIDd	Right ventricular short-axis diameter at end diastole	Line
LVIDd	Light ventricular short-axis diameter at end diastole	Line

<<Items to be calculated>>

Meas. item	Description	Calc. formula
RV/LV	The ratio of RV and LV	RVIDd/LVIDd

20. IVC

<<Item to be measured>>

Meas. item name	Description	Meas. method
IVC Ins[mm]	Inferior Vena Cava inspiration	Measurement (Distance)
IVC Exp[mm]	Inferior Vena Cava expiration	Measurement (Distance)

<<Items to be calculated>>

Meas. item	Description	Calc. formula
IVC Ins / IVC Exp	The ratio of IVC Ins and IVC Exp	IVC Ins/IVC Exp

21. RA/LA

<<Item to be measured>>

Meas. item name	Description	Meas. method
RA[cm ²]	Right Atrium	Measurement (Area)
LA[cm ²]	Left Atrium	Measurement (Area)

<<Items to be calculated>>

Meas. item	Description	Calc. formula
RA / LA	The ratio of RA and LA	RA/LA

22. AO/LA

<<Item to be measured>>

Meas. item name	Description	Meas. method
AO[mm]	Aorta Diameter	Measurement (Distance)
LA[mm]	Left Atrium Diameter	Measurement (Distance)

<<Items to be calculated>>

Meas. item	Description	Calc. formula
AO / LA	The ratio of AO and LA	AO/LA

5.3.1.3 Urology Measurement in B mode

SonoBladder:

Get a good bladder image in B mode, click [SonoBladder], the system will automatically trace the bladder and calculate the circumference and area of the bladder.

Meas. item name	mark	unit	Meas. method
Residual Vol.	Pre Length	cm	Refer to "Distance Measurement (Distance)".
	Pre Width	cm	Refer to "Distance Measurement (Distance)".
	Pre Height	cm	Refer to "Distance Measurement (Distance)".
	Pre Vol.	ml	$\text{Pre Vol.} = (\pi/6) * \text{Pre Length} * \text{Pre Width} * \text{Pre Height}$
	Post Length	cm	Refer to "Distance Measurement (Distance)".
	Post Width	cm	Refer to "Distance Measurement (Distance)".
	Post Height	cm	Refer to "Distance Measurement (Distance)".
	Post Vol.	ml	$\text{Post Vol.} = (\pi/6) * \text{Post Length} * \text{Post Width} * \text{Post Height}$
	Void Vol.	ml	$\text{Void Vol.} = \text{Pre Vol.} - \text{Post Vol.}$
Prostate Vol.	Length	cm	Refer to "Distance Measurement (Distance)"
	Width	cm	Refer to "Distance Measurement (Distance)"
	Height	cm	Refer to "Distance Measurement (Distance)"
	Volume	ml	$V = (\pi/6) * \text{PV Length} * \text{PV Width} * \text{PV Height}$
	PPSA	ml	$\text{PPSA} = 0.12 * \text{PV}$
	PSAD	ng/ ml	$\text{PSAD} = \text{SPSA}[\text{ng}]/\text{PV}[\text{ml}]$ Note: SPSA must be input by manual in patient interface when you select the urology exam type
Kid Volume	Length	cm	Refer to "Distance Measurement (Distance)"
	Width	cm	Refer to "Distance Measurement (Distance)"
	Height	cm	Refer to "Distance Measurement (Distance)"
	Renal Pelvis	cm	Refer to "Distance Measurement (Distance)"
T-Zone	Length	cm	Refer to "Distance Measurement (Distance)"
	Width	cm	Refer to "Distance Measurement (Distance)"
	Height	cm	Refer to "Distance Measurement (Distance)"
	Volume	ml	$V = (\pi/6) * \text{length} * \text{Width} * \text{Height}$
Bladder Vol	Length	cm	Refer to "Distance Measurement (Distance)"
	Width	cm	Refer to "Distance Measurement (Distance)"
	Height	cm	Refer to "Distance Measurement (Distance)"

	Volume	ml	$V = (\pi/6) * \text{length} * \text{Width} * \text{Height}$
StA%	A Out A In StA%	cm ² cm ² %	Refer to "Area Measurement (Area)"
StD%	D Out D In StD%	cm ² cm ² %	Refer to "Area Measurement (Area)"
Vessel Area	Vessel Area	cm ²	Refer to "Area Measurement (Area)"
Vessel Dis	Vessel Dis	cm	Refer to "Distance Measurement (Distance)"

5.3.1.4 Pediatric Measurement in B mode

Hip

Hip is used for evaluating the abnormality of the cotyle. In order to make measurement, user has to draw three lines on the image responding to the anatomy. The system will calculate the angles automatically.

Operation Method:

1. Select HIP from Pediatric submenu on CALC menu. Press ENTER-key to select it.
2. The first line D1 will be displayed on the screen and use the Trackball to move the line.
3. The second line D2 will be displayed on the screen, repeat the above procedure to fix the second line and third line D3. The angles between these three lines will be displayed on the screen.




Fig.5-1



NOTE: α is the angle between the D1 and D2 (acute angle). β is the angle between D1 and D3 (acute angle).

5.3.1.5 Vascular Measurement in B mode

For CCA and ICA, user also can measure the thickness of intima.

IMT: measure the thickness of intima manual.

IMT Auto: get a good B image for intima, and then click this measurement. Set start position and end position, be sure to contain the intima in the box, the system will calculate the thickness of

intimae.




Fig.5-2




Fig.5-3

Meas. item name	mark	unit	Meas. Method and calc. formular
Prox CCA (Rt) / (Lt)	A Out	cm²	Refer to "Area/Circumference Measurement (Area)"
	A In	cm²	Refer to "Area/Circumference Measurement (Area)"
	StA%	%	StA%= (A Out- A In) / A Out *100%
	D Out	cm	Refer to "Distance Measurement (Distance)"
	D In	cm	Refer to "Distance Measurement (Distance)"
	StD%	%	StD%= (D Out- D In) / D Out *100%

	Vessel Area Vessel Dis	cm ² cm	Refer to Area(Ellipse) Refer to "Distance Measurement (Distance)
Mid CCA (Rt) / (Lt)	The same as above	The same as above	The same as above
Distal CCA (Rt) / (Lt)	The same as above	The same as above	The same as above
Prox ICA (Rt) / (Lt)	The same as above	The same as above	The same as above
Mid ICA (Rt) / (Lt)	The same as above	The same as above	The same as above
Distal ICA (Rt) / (Lt)	The same as above	The same as above	The same as above
ECA (Rt) / (Lt)	The same as above	The same as above	The same as above
Vertebral A (Rt) / (Lt)	The same as above	The same as above	The same as above
INT IIL (Rt) / (Lt)	The same as above	The same as above	The same as above
EXT IIL (Rt) / (Lt)	The same as above	The same as above	The same as above
ILIAC (Rt) / (Lt)	The same as above	The same as above	The same as above
CFA (Rt) / (Lt)	The same as above	The same as above	The same as above
ProFun (Rt) / (Lt)	The same as above	The same as above	The same as above
LTCIR (Rt) / (Lt)	The same as above	The same as above	The same as above
SFA (Rt) / (Lt)	The same as above	The same as above	The same as above
Pop A (Rt) / (Lt)	The same as above	The same as above	The same as above
ATA (Rt) / (Lt)	The same as above	The same as above	The same as above
PTA (Rt) / (Lt)	The same as above	The same as above	The same as above
PERON	The same	The same	The same as above

(Rt) / (Lt)	as above	as above	
DRPED (Rt) / (Lt)	The same as above	The same as above	The same as above

5.3.1.6 GYN Measurement in B mode

Meas. item name	Mark	Unit	Meas. Method and calc. formular
UT	UT_L UT_W UT_H UT_D CX_L	cm cm cm cm cm	Refer to "Distance Measurement (Distance) " Refer to "Distance Measurement (Distance) " "Refer to "Distance Measurement (Distance) " UT_D= UT_L +UT_H + UT_W Refer to "Distance Measurement (Distance) "
Cervix Vol.	Length Height Width Volume	cm cm cm cm	Refer to "Distance Measurement (Distance) " Refer to "Distance Measurement (Distance) " Refer to "Distance Measurement (Distance) " $PV = (\pi/6) * L * W * H$
ENDO	endometrium	cm	Refer to "Distance Measurement (Distance) "
OV_Volume	Length Width Height Volume	cm cm cm ml	Refer to "Distance Measurement (Distance) " Refer to "Distance Measurement (Distance) " Refer to "Distance Measurement (Distance) " $Volume = (\pi/6) * L * W * H$
FO_D	Length Width Height	cm cm cm	Refer to "Distance Measurement (Distance) " Refer to "Distance Measurement (Distance) " Refer to "Distance Measurement (Distance) "
FO Auto			Auto calculate Area and Length
Uterine Artery	A Out A In STA% D Out D In STD% Vessel Area Vessel Dis	cm ² cm ² % cm cm % cm ² cm	Refer to "Refer to Area(Ellipse)" Refer to "Refer to Area(Ellipse)" $StA\% = (A \text{ Out} - A \text{ In}) / A \text{ Out} * 100\%$ Refer to "Distance Measurement (Distance) " Refer to "Distance Measurement (Distance) " $StD\% = (D \text{ Out} - D \text{ In}) / D \text{ Out} * 100\%$ Refer to "Refer to Area(Ellipse)" Refer to "Distance Measurement (Distance) "

5.3.1.7 OB Measurement in B mode

SonoCNS

SonoCNS is to extract the features of key anatomical tissues and organs from the three-dimensional intracranial data of the fetus. It is a modal technology to detect the fetal median sagittal plane, transverse thalamus, lateral ventricle plane and transverse cerebellar plane, which helps the doctor find the sagittal plane quickly so as to observe the anatomical structure in the sagittal plane. Get a good 3D image and click [SonoCNS] on the touchscreen to get into the function.

Meas. item name	Mark	Unit	Meas. Method and calc. formula
GS	GS	cm	Refer to "Distance Measurement (Distance)" EDD = current date + (280days – average UGA)
	GA	week/day	
	EDD	mm/dd/yyyy	
CRL	CRL	cm	Refer to "Distance Measurement (Distance)" EDD = current date + (280days – average UGA)
	GA	week/day	
	EDD	mm/dd/yyyy	
YS	YS	cm	Refer to "Distance Measurement (Distance)"
BPD	BPD	cm	Refer to "Distance Measurement (Distance)"
	GA	week/day	
	EDD	mm/dd/yyyy	
	EFW		
OFD	OFD	cm	Refer to "Distance Measurement (Distance)"
	GA	week/day	
	EDD	mm/dd/yyyy	
HC (ellipse)	HC	cm	Refer to "Distance Measurement (Distance)"
	GA	week/day	
	EDD	mm/dd/yyyy	
APD	APD	cm	Refer to "Distance Measurement (Distance)"
	GA	week/day	
	EDD	mm/dd/yyyy	
TAD	TAD	cm	Refer to "Distance Measurement (Distance)"
	GA	week/day	
	EDD	mm/dd/yyyy	
AC (ellipse)	AC	cm	Refer to "Distance Measurement (Distance)"
	GA	week/day	
	EDD	mm/dd/yyyy	
	EFW	g	
FTA	FTA	cm	Refer to "Distance Measurement (Distance)"
	GA	week/day	
	EDD	mm/dd/yyyy	

FL	FL GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
SL	SL	cm	Refer to "Distance Measurement (Distance)"
APTD	APTD GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
TTD	TTD GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
ThC	ThC	cm	Refer to "Distance Measurement (Distance)"
Humerus	Humerus GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
ULNA	ULNA GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
Tibian	Tibian GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
RAD	RAD	cm	Refer to "Distance Measurement (Distance)"
FIB	FIB	cm	Refer to "Distance Measurement (Distance)"
CLAV	CLAV GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
CER	CER GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
CM	CM	cm	Refer to "Distance Measurement (Distance)"
NF	NF	cm	Refer to "Distance Measurement (Distance)"
OOD	OOD GA EDD	cm week/day mm/dd/yyyy	Refer to "Distance Measurement (Distance)"
IOD	IOD	cm	Refer to "Distance Measurement (Distance)"
NB	NB	cm	Refer to "Distance Measurement (Distance)"
Lvent	Lvent GA EDD	cm	Refer to "Distance Measurement (Distance)"

HW	HW	cm	Refer to "Distance Measurement (Distance)"
LtKid	LtKid	cm	Refer to "Distance Measurement (Distance)"
RtKid	RtKid	cm	Refer to "Distance Measurement (Distance)"
LtRenalAP	LtRenalAP	cm	Refer to "Distance Measurement (Distance)"
RtRenalAP	RtRenalAP	cm	Refer to "Distance Measurement (Distance)"
LVWrHEM	LVWrHEM	cm	Refer to "Distance Measurement (Distance)"
MAD	MAD	cm	Refer to "Distance Measurement (Distance)"
AFI	AF1	cm	Refer to "Distance Measurement (Distance)"
	AF2	cm	Refer to "Distance Measurement (Distance)"
	AF3	cm	Refer to "Distance Measurement (Distance)"
	AF4	cm	Refer to "Distance Measurement (Distance)"
	AFI	cm	AFI=AF1+AF2+AF3+AF4
FBP	AF	cm	Refer to "Distance Measurement (Distance)"
	Result	Scores	Give result according to AF

5.3.1.8 Carotid Measurement in B mode

Meas. item name	Mark	Unit	Meas. Method and calc. formular
Subclavian A (Lt/Rt)	A Out	cm ²	Refer to "Area Measurement (Ellipse)"
	A In	cm ²	Refer to "Area Measurement (Ellipse)"
	StA%	%	StA % = (A Out - A In) / A Out * 100%
	D Out	cm	Refer to "Distance Measurement (Distance)"
	D In	cm	Refer to "Distance Measurement (Distance)"
	StD%	%	StD % = (D Out - D In) / D Out * 100%
	Vessel Area	cm ²	Refer to "Area Measurement (Ellipse)"
	Vessel Dis	cm	Refer to "Distance Measurement (Distance)"
Prox CCA (Lt/Rt)	The same as above	The same as above	The same as above
Mid CCA (Lt/Rt)	The same as above	The same as above	The same as above
Distal CCA (Lt/Rt)	The same as above	The same as above	The same as above
Bulb (Lt/Rt)	The same as above	The same as above	The same as above
Prox ICA (Lt/Rt)	The same as above	The same as above	The same as above

Mid ICA (Lt/Rt)	The same as above	The same as above	The same as above
Distal ICA (Lt/Rt)	The same as above	The same as above	The same as above
ECA (Lt/Rt)	The same as above	The same as above	The same as above
Vertebral (Lt/Rt)	The same as above	The same as above	The same as above
General Measurement	The same as above	The same as above	The same as above
Flow Volume	Vessel Dis Vessel Area	cm cm ²	Refer to "Distance Measurement (Distance)" Refer to "Area Measurement (Ellipse)"

5.3.1.9 Abdomen Measurement in B mode

Meas. item name	Mark	Unit	Meas. Method and calc. formular
CBD	CBD	cm	Refer to "Distance Measurement (Distance)"
GB Wall	GB Wall	cm	Refer to "Distance Measurement (Distance)"
Liver Length	Liver Length	cm	Refer to "Distance Measurement (Distance)"
Artery	Height Width A Out A In StA% D Out D In StD% Vessel Area Vessel Dis	cm cm cm cm ² % cm cm % cm ² cm	Refer to "Distance Measurement (Distance)" Refer to "Distance Measurement (Distance)" Refer to "Area Measurement (Trace)" Refer to "Area Measurement (Trace)" $StA\% = (A\ Out - A\ In) / A\ Out * 100\%$ Refer to "Distance Measurement (Distance)" Refer to "Distance Measurement (Distance)" $StD\ \% = (D\ Out - D\ In) / D\ Out * 100\%$ Refer to "Area Measurement (Trace)" Refer to "Distance Measurement (Distance)"
Spleen	Length Height Width Volume	cm cm cm ml	Refer to "Distance Measurement (Distance)" Refer to "Distance Measurement (Distance)" Refer to "Distance Measurement (Distance)" $Volume = (\pi/6) * Length * Width * Height$
Renal Vol.	Length Height Width	cm cm cm	Refer to "Distance Measurement (Distance)" Refer to "Distance Measurement (Distance)" Refer to "Distance Measurement (Distance)"

GB Volume	Length	cm	Refer to "Distance Measurement (Distance)"
	Height	cm	Refer to "Distance Measurement (Distance)"
	Width	cm	Refer to "Distance Measurement (Distance)"
Lliac	Height	cm	Refer to "Distance Measurement (Distance)"
	Width	cm	Refer to "Distance Measurement (Distance)"
	A Out	cm ²	Refer to "Area Measurement (Trace)"
	A In	cm ²	Refer to "Area Measurement (Trace)"
	StA%	%	$StA\% = (A\ Out - A\ In) / A\ Out * 100\%$
	D Out	cm	Refer to "Distance Measurement (Distance)"
	D In	cm	Refer to "Distance Measurement (Distance)"
	StD%	%	$StD\% = (D\ Out - D\ In) / D\ Out * 100\%$
	Vessel Area	cm ²	Refer to "Area Measurement (Trace)"
	Vessel Dis	cm	Refer to "Distance Measurement (Distance)"

5.3.1.10 Small Parts Measurement in B mode

Breast

Measure the vertical diameter, transverse diameter and anteroposterior diameter of the breast and calculate the volume.

SonoBreast

Get a good breast image in B mode, click [SonoBreast], the system will automatically trace the breast and calculate the circumference and area of the breast.

Thyroid

Measure the vertical diameter, transverse diameter and anteroposterior diameter of the thyroid and calculate the volume.

SonoThyroid

Get a good thyroid image in B mode, click [SonoThyroid], the system will automatically trace the thyroid and calculate the circumference and area of the thyroid.

As for other measurement items, please refer to general measurement in B mode.

5.3.1.11 TCD Measurement in B mode

Refer to vascular measurement in B mode.

5.3.2 Measurement in M mode

5.3.2.1 General Measurement in M mode

1. M distance

This feature allows the measurement of the distance between two points. It is a measurement between the two horizontal lines that lean on the two cursors. The position of the vertical time line does not affect the distance measurement

Operation:

Rotate the trackball to select the “M distance” item in the menu. A start cursor “+” will appear on the screen. Move the cursor through rotating the trackball and press ENTER-key to fix the first point. The second cursor will appear. Move the second cursor to the end point and press ENTER-key to fix it. The measurement result will appear on the right side of the screen.

2. M time

Time is the measurement between the two vertical time lines created by two cursors. The position of the horizontal distance line does not affect time measurements.

3. Velocity

Velocity is the measurement between the intersections of the two cursors. Velocity can be positive or negative, and is measured as the rate of change between the two points defined by the intersections of the cursors in cm/sec.

4. HR

HR is the measurement between the two vertical lines that are created by two cursors in beat per minute (BPM). The position of the horizontal distance line does not affect HR.

5.3.2.2 Cardiac Measurement in M mode

1. Distance

Refer to distance measurement in general measurement in M mode.

2. Time

Refer to time measurement in general measurement in M mode.

3. Slope

Refer to slope measurement in general measurement in M mode.

4. HR

Refer to HR measurement in general measurement in M mode.

5. Left Ventricle

Cube

Calculation formula for volume:

$$EDV = LVIDd^3$$

$$ESV = LVIDs^3$$

<< Items to be measured >>

Meas. item	Description	Meas. Tool
------------	-------------	------------

IVSd[cm]	Interventricular septal thickness at end diastole	M Line
LVIDd[cm]	Left ventricular short-axis diameter at end diastole	M Line
LVPWd[cm]	Left ventricular posterior wall thickness at end diastole	M Line
IVSs[cm]	Interventricular septal thickness at end systole	M Line
LVIDs[cm]	Left ventricular short-axis diameter at end systole	M Line
LVPWs[cm]	Left ventricular posterior wall thickness at end systole	M Line

<<Items to be calculated>>

Calc. item	Description	Calc. formula
EDV[ml]	End-diastolic left ventricular volume	Refer to << Calculation formula for volume>>
ESV[ml]	End-systolic left ventricular volume	Refer to << Calculation formula for volume>>
SV [ml]	Stroke volume	SV=EDV – ESV
CO	Cardiac output [L/min]	CO=SV x HR /1000
EF	Ejection fraction	EF=SV/ EDV
FS	Fractional shortening	FS=(LVIDd-LVIDs)/LVIDd
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

Teichholz

<<Calculation formula for volume>>

$$EDV = \frac{7 \times LVIDd^3}{2.4 + LVIDd}$$

$$ESV = \frac{7 \times LVIDs^3}{2.4 + LVIDs}$$

<<Items to be measured>>

Meas. item	Description	Meas. Tool
LVIDd [cm]	Left ventricular internal diameter at end diastole	M Line
LVIDs [cm]	Left ventricular internal diameter at end systole	M Line

<< Items to be calculated >>

Calc. item	Description	Calc. formula
EDV [ml]	End-diastolic left ventricular Volume	Please refer to <<Calculation formula for volume>>
ESV [ml]	End-systolic left ventricular Volume	Please refer to <<Calculation formula for volume>>
SV [ml]	Stroke Volume	SV=EDV – ESV

CO [l/min]	Cardiac Output	CO=SV*HR/1000
EF [%]	Ejection Fraction	EF=SV/EDV
FS [%]	Fractional Shortening	FS=(LVIDd-LVIDs)/LVIDd
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

Gibson

Calculation formula for volume:

$$EDV = \frac{\pi}{6} \times (0.98 \times LVIDd + 0.59) \times LVIDd^2$$

$$ESV = \frac{\pi}{6} \times (0.98 \times LVIDs + 0.59) \times LVIDs^2$$

<<Items to be measured>>

Meas. item	Description	Meas. Tool
LVLd[cm]	Left ventricular long-axis length at end diastole.	M Line
LVLs[cm]	Left ventricular long-axis length at end systole.	M Line

<<Items to be calculated>>

Calc. item	Description	Calc. formula
EDV[ml]	End-diastolic volume	Please refer to <<Calculation formula for volume>>
ESV [ml]	End-systolic volume	Please refer to <<Calculation formula for volume>>
SV[ml]	Stroke Volume	SV=EDV – ESV
CO[L/min]	Cardiac output	CO=SV * HR /1000
EF	Ejection Fraction	EF=SV/EDV
SI	Stroke Volume Index	SI=SV/BSA
CI	Cardiac Output Index	CI=CO/BSA

IVSd/LVPWd

<<Items to be measured>>

Meas. item	Description	Meas. Tool
IVSd[cm]	Interventricular septal thickness at end diastole	M Line
LVPWd[cm]	Left ventricular posterior wall thickness at end diastole	M Line

6. Mitral Valve

<<Items to be measured>>

Meas. item	Description [Unit]
------------	--------------------

EPSS[cm]	Distance between the mitral valve and interventricular septum[cm]
MV E Amp[cm]	Mitral Valve E wave amplitude
MV F Amp[cm]	Mitral Valve E wave amplitude
MV D-E Exc Dist	D-E Slope (cm/s)
	D-E Exc Time(s)
	D-E Exc Distance(cm)
MV E-F slope	E-F Slope (cm/s)
	E-F Time(s)
	E-F Distance(cm)
MV D-E slope	D-E Slope (cm/s)
	E-F Time(s)
	E-F Distance(cm)
MV A-C Int time	A-C Int slope (cm/s)
	A-C Int Time(s)
	A-C Int Distance(cm)
E Duration	E Duration
A Duration	A Duration

7. Aortic Valve

<Items to be measured>>

Meas. item name	Description	Meas. Method
AOD[cm]	Aorta diameter at diastole	Refer to distance measurement in M mode
Ao Sinus Diam[cm]	Aortic sinus diameter	Refer to distance measurement in M mode
Ao Asc Diam[cm]	Aorta ascendens diameter	Refer to distance measurement in M mode
Ao Arch Diam[cm]	Aortic arch diameter	Refer to distance measurement in M mode
Ao Desc Diam[cm]	Descending artery diameter	Refer to distance measurement in M mode
LVOT Diam[cm]	Left ventricular outflow tract diameter	Refer to distance measurement in M mode
LAD[cm]	Left atrium diameter	Refer to distance measurement in M mode
AA[cm]	Aorta amplitude	Refer to distance measurement in M mode
LVEP	Left ventricular pre-ejection phase	Refer to time measurement in M mode
LVET	Left ventricular ejection time	Refer to time measurement in M mode

8. Tricuspid Valve

<<Items to be measured>>

Meas. item name	Description [Unit]
RVOT Diam[cm]	Right ventricular outflow tract diameter
RA Diam[cm]	Right atrium diameter
D-E Exc Dist	D-E Slope (cm/s)
	D-E Exc Time(s)
	D-E Exc Distance(cm)
E-F slope	E-F Slope (cm/s)
	E-F Time(s)
	E-F Distance(cm)
MV A-C Int time	A-C Int slope (cm/s)
	A-C Int Time(s)
	A-C Int Distance(cm)

9. Pulmonary Valve

<<Items to be measured>>

Meas. item name	Description [Unit]
RVPEP[s]	Right ventricular pre-ejection phase
RVET[s]	Right ventricular ejection time
A wave Amp (cm)	Atrium wave amplitude
B-C slope	B-C Slope (cm/s)
	B-C Time(s)
	B-C Distance(cm)

10. RV/LV

<<Items to be measured>>

Meas. item	Description	Meas. Tool
RVIDd	Right ventricular short-axis diameter at end diastole	M Line
LVIDd	Left ventricular short-axis diameter at end diastole	M Line

<<Items to be calculated>>

Meas. item	Description	Calc. formula
RV/LV	The ratio of RV and LV	RVIDd/LVIDd

11. LV Mass

<<Items to be measured>>

Meas. item name	Description
-----------------	-------------

IVSd	Interventricular septal thickness at end diastole
LVIDd	Left ventricular short-axis diameter at end diastole
LVPWd	Left ventricular posterior wall thickness at end diastole

12. TAPSE

<<Items to be measured>>

Meas. item name	Description
TAPSE	Tricuspid annular plane systolic excursion

5.3.3 Measurement in PW mode

5.3.3 1 General Measurement in PW mode

1. Velocity

- 1) Rotate the trackball to move the cursor to the velocity item in the menu and press ENTER-key to select it. A cursor "+" will appear on the screen.
- 2) Move the cursor "+" to the place where need to be measured and press ENTER-key to fix it.
- 3) The value of velocity and pressure will appear on the screen.
- 4) Repeat step 1) to 3) to measure the next point.

2. Distance

The measurement method is the same as distance in general measurement in B mode.

3. Peak

Calculate the velocity over one cardiac cycle. The velocity, slope, RI and SD ratio are calculated.

- 1) Scan the object area in PW mode and freeze the image.
- 2) Rotate the trackball to move the cursor to the Peak item in the menu and press ENTER-key to select.
- 3) A cursor "+" will appear on the screen. Move the cursor "+" to the peak point where the cardiac systole and press ENTER-key to fix it.
- 4) A second cursor "+" will appear on the screen again. Fix the second cursor to the end point where the cardiac diastole.
- 5) When the two points are all fixed, the value of Vs,Vd,RI,SD(Vs/Vd) will appear on the right side of the screen.

4. Auto Trace

- 1) Rotate the trackball to move the cursor to the auto trace item in the menu.
- 2) Press ENTER-key to select it. A cursor "]" will appear on the screen.
- 3) Move the cursor by rotating the trackball to the start point of the one cycle and press ENTER-

- key to fix it. A second cursor "I" will appear.
- 4) Rotate the trackball the end point of the cycle and press ENTER-key to fix it.
- 5) After the measurement, the value of Vs,Vd,RI,SD ratio,PI will appear on the screen.




Fig.5-4

- Peak Velocity or Doppler frequency

Peak velocity over one cardiac cycle (Vpk)

- Doppler Time Distance

Time distance between two cursors in ms. The invert is the heart beat rate if the two cursors are at one cardiac cycle period. (T)

- Spectral Velocity Time Integral (VTI)

$VTI = \sum V_{pk} * (\Delta t)$. Where Δt is T/N , N is the data point over one cardiac cycle.

- Pulsatility Index (PI)

It can be used to represent the degree of pulse-wave damping at different arterial sites; the smaller the PI, the greater the degree of damping. Typical value for CCA is 1.90 ± 0.5 .

$PI = |(A-B)/TimeAvgPk|$, $TimeAvgPk = \sum V_{pk} / N$

- Resistive Index (RI)

Vary from 0 to 1. It is an indicator of the circulatory resistance. Typical value for CCA is 0.75 ± 0.05 .

$RI = |(A-B) / A|$



NOTE:

- a) In order to get accurate result, the PW image must be clear and high quality.
- b) Insure you fix the cursor at the exact place of cardiac systole and diastole.

5. Manual trace

Automatically measure the Peak Systolic Velocity, VTI, end diastolic velocity, HR, Time, minimum velocity, PI, and RI after manually tracing of the curve is finished.



NOTE: Manual trace requires the user to trace two peak points of two

cycles.

6. STD%

Measure the diameter reduction ratio of vessel in B mode.

- 1) Move the cursor to the “STD” item and press ENTER-key to select it .A cursor “+” will appear on the screen.
- 2) Move the cursor to the point of outside wall of the vessel and press ENTER-key to fix it. The method to measure the diameter of the outside wall of the vessel is the same as “distance” in general measurement in B mode.
- 3) When the diameter of outside wall of the vessel is finished, the cursor “+” will appear on the screen again. Measure the diameter of the stenosis area.
- 4) The value of diameter and STD will be displayed in the result window.

7. STA%

Measure the area reduction ratio of vessel in B mode.

- 1) Move the cursor to the “STA” item and press ENTER-key to select it .A cursor “+” will appear on the screen.
- 2) Move the cursor to the point of outside wall of the vessel and press ENTER-key to fix it. The method to measure the area of the outside of the vessel is the same as “area-ellipse” in general measurement in B mode.
- 3) When the area of outside of the vessel is finished , the cursor “+” will appear on the screen again. Measure the area of the stenosis area.
- 4) The value of area and STD will be displayed in the result window.

8. Area

The same as “area-trace” in general measurement in B mode.

9. ICA/CCA

Measure ICA and CCA separately and get the ratio value of ICA/CCA

10. Flow volume

<<Items to be measured>>

Meas. item name	Description [Unit]	Meas. Method
Diam	Diameter of vessel [mm]	Refer to “distance” measurement
VTI	Velocity time integral	Refer to “area-trace” measurement
Time	Time	Refer to “time” measurement in M mode

<<Items to be calculated>>

Calc. item name	Description [Unit]	Calc. formula
SV	Stroke volume[ml]	$SV = 0.785 * \text{diameter} * \text{diameter} * \text{VTI} $

CO	Cardiac output[L/min]	$CO = SV \times \text{heart rate} / 1000$
HR	Heart rate	Heart rate = 60 / time

11. HR

The measurement is the same as “heart rate” in general measurement in M mode.

5.3.3.2 Cardiology Measurement in PW mode

1. MV

Measurement of Mitral valve function is performed in frozen Doppler spectral trace.

<<Measurement and calculation items>>

Meas. item name	Description [Unit]	Meas. procedure / Calc. formula
E Vel	E-wave flow velocity [cm/s]	Refer to "Velocity Measurement".
A Vel	A-wave flow velocity [cm/s]	Refer to "Velocity Measurement".
PHT	Pressure half time [ms]	Refer to "Velocity Measurement".
E Dur	E-wave duration [ms]	Refer to "Time Measurement (Time)".
A Dur	A-wave duration [ms]	Refer to "Time Measurement (Time)".
IRT	Isovelocity relaxation time [ms]	Refer to "Time Measurement (Time)".
MV VTI	Mitral valve VTI [m]	Refer to "Full trace".
MV VM	Mitral valve mean velocity [cm/s]	Refer to "Full trace"
MV VP	Mitral valve maximum velocity [cm/s]	Refer to "Full trace"
MV MPG	Mitral valve mean pressure gradient [mmHg]	Refer to "Full trace"
MV PPG	Mitral valve maximum pressure gradient [mmHg]	Refer to "Full trace"
MV Diam	Mitral valve diameter [mm]	Refer to "Distance Measurement in B mode"
HR	Heart rate [bpm]	Refer to "Heart Rate Measurement (HR)"
E/A	[no unit]	E Vel / A Vel
A/E	[no unit]	A Vel / E Vel
SV	Stroke volume [mL]	$0.785 \times (MVDiam)^2 \times \text{abs} VTI $
CO	Cardiac output [L/min]	$SV \times HR / 1000$
MV Area	Mitral valve area [mm ²]	220 / PHT

Note: [MV Diam] can only be performed in 2-D mode.

<<Measurement points>>

- 1) Measure [E Vel] and [A Vel]. Measure [PHT]. Measure [E-Dur], [A-Dur], and [IRT] Measure

[HR].




Fig.5-5

2) Measure [MV V Trace].




Fig.5-6

3) Measure [MV Diameter]




Fig.5-7

2. AV

Measurement of aortic valve function is performed by using 2-D-mode and Doppler images.

<<Measurement items>>

Meas. Item name	Description [Unit]	Meas. method
LVOT Vel	PW-mode LV out tract flow velocity [cm/s]	Refer to "Velocity Measurement".
LVOT PG	PW-mode LV out tract pressure gradient [mmHg]	Refer to "Velocity Measurement".

LVOT V General

Meas. Item name	Description [Unit]	Meas. method
LVOT VTI	Aortic valve VTI [m]	Refer to "Full trace"
LVOT VM	Aortic valve mean velocity [cm/s]	Refer to "Full trace"
LVOT VP	Aortic valve maximum velocity [cm/s]	Refer to "Full trace"

LVOT MPG	Aortic valve mean pressure gradient [mmHg]	Refer to "Full trace"
LVOT PPG	Aortic valve maximum pressure gradient[mmHg]	Refer to "Full trace"

AoV VM

Meas. Item name	Description [Unit]	Meas. method
AoV MPG	Aortic valve mean pressure gradient [mmHg]	Refer to "Full trace"
AoV PPG	Aortic valve maximum pressure gradient [mmHg]	Refer to "Full trace"
LVOT Diam	Left ventricular outflow tract diameter [mm]	Refer to "Distance Measurement in B mode".
HR	Heart rate [bpm]	Refer to "Heart Rate Measurement (HR)".
SV	Stroke volume [mL]	$0.785 \times (\text{LVOT Diam})^2 \times \text{abs VTI }$
CO	Cardiac output [L/min]	$\text{SV} \times \text{HR} / 1000$
AV Area	AoV area [cm ²]	$\text{SV} / \text{AoV VTI} / 100$

Note: [LVOT Diam] can only be performed in 2D mode

<<Measurement points>>

- 1) Measure [LVOT Vel]. Measure [HR].




Fig.5-8

- 2) Measure [LVOT Diam]




Fig.5-9

3. TV

Measurement of tricuspid valve function is performed by using Doppler images.

<<Measurement items>>

Meas. Item name	Contents	Meas. Method
TV VTI	Tricuspid valve VTI [m]	Refer to "Manual Trace Measurement "
TV VM	Tricuspid valve mean velocity [cm/s]	Refer to " Manual Trace Measurement "
TV VP	Tricuspid valve maximum velocity [cm/s]	Refer to " Manual Trace Measurement "
TV MPG	Tricuspid valve mean pressure gradient [mmHg]	Refer to " Manual Trace Measurement "
TV PPG	Tricuspid valve maximum pressure gradient [mmHg]	Refer to "Manual Trace Measurement "
TV Vel	Tricuspid valve velocity [cm/s]	Refer to "Velocity Measurement"
TV PG	Tricuspid valve pressure gradient [mmHg]	Refer to " Velocity Measurement'

<<Measurement points>>

1) Measure [TV VEL Trace]




Fig.5-10

2) Measure [TV VEL]




Fig.5-11

4. PV

Measurement of pulmonary valve function is performed by using a Doppler image.

<<Measurement items>>

Meas. item name	Description [Unit]	Meas. method
PV VTI	Pulmonary valve VTI [m]	Refer to "Full Trace Measurement ".

PV VM	Pulmonary valve mean velocity [cm/s]	Refer to "Full Trace Measurement ".
PV VP	Pulmonary valve maximum velocity [cm/s]	Refer to "Full Trace Measurement ".
PV MPG	Pulmonary valve mean pressure gradient [mmHg]	Refer to "Full Trace Measurement ".
PV PPG	Pulmonary valve maximum pressure gradient [mmHg]	Refer to " Velocity Measurement"
PV Max Vel	Maximum velocity [cm/s]	Refer to " Velocity Measurement"
PV Max PG	Maximum pressure gradient [mmHg]	Refer to "Velocity Measurement"
PV Diam	Pulmonary artery diameter [mm]	Refer to "Distance Measurement in B mode"
HR	Heart rate [bpm]	Refer to "Heart Rate Measurement (HR)"
RV ET	Ejection time [ms]	Refer to "Time Measurement (Time)"
RV AcT	Acceleration time [ms]	Refer to "Time Measurement (Time)"
RV PEP	Pre-ejection period [ms]	Refer to "Time Measurement (Time)"

 **NOTE:** [PV Diam] can only be performed in 2D mode.

<<Calculation items>>

Calc. item name	Description [Unit]	Calc. formula
SV	Stroke volume [mL]	$0.785 \times (\text{PV Diam})^2 \times \text{abs VTI }$
CO	Cardiac output [L/min]	$\text{SV} \times \text{HR} / 1000$
RV AcT/ET	Ratio of AcT to ET	$\text{RV ACT} / \text{RV ET}$
RV STI	Systolic time interval	$\text{RV REP} / \text{RV ET}$

<<Measurement points>>

1) Measure [PV Vel]




Fig.5-12

2) Measure [PVV Trace]




Fig.5-13

3) Measure [ET], [ACT], and [PEP]




Fig.5-14

4) Measure [PV Diam]




Fig.5-15

5.3.3.3 Urology Measurement in PW mode

The measurement method and calculation items are the same as general measurement in PW mode.

5.3.3.4 Pediatric Measurement in PW mode

The measurement method and calculation items are the same as general measurement in PW mode.

5.3.3.5 Small parts Measurement in PW mode

The measurement method and calculation items are the same as general measurement in PW mode.

5.3.3.6 Vascular Measurement in PW mode

The measurement method and calculation items are the same as “manual trace”.

5.3.3.7 GYN Measurement in PW mode

The measurement method and calculation items are the same as “manual trace”.

5.3.3.8 OB Measurement in PW mode

The measurement method and calculation items are the same as “manual trace”.

5.4 Edit measurement results

After the user finishes the measurements, this system allows users to move the position of measurement results, or change the font size of measurement results.

To move result position, operations are as follows:

1. In the measurement status, rotate SK2 key anticlockwise to activate “Move Result Pos.”.
2. Use trackball to move the position of measurement results.
3. Press 【ENTER】 key to confirm it.

To reset result position, operations are as follows:

1. In the measurement status, rotate SK2 key clockwise to activate “Reset Result Pos.”.
2. The position of current measurement result will be reset back to default.

To change the font size of measurement results, operations are as follows:


In the measurement status, rotate SK1 key to adjust the font size of measurement results.

To reset the font size of measurement results, operations are as follows:

In the measurement status, press SK1 key, the font size of measurement results will be reset back to default. The system default font size can be set in system setting, refer to 7.2.

5.5 Report

After finishing the exam, press the



key on the touch screen to pop up report interface,

edit and print the report.

Report Interface:




Fig.5-16

Choose the image: Click the image can add the image to the image area in the report, and click the image in the report to remove it.

Preview: preview the whole report and its format before printing.

Print: print the current report, please make sure that the printer is normal working.

Save: store the report on the disk.

OK: confirm the operation and exit the interface.

Undo: cancel the operation and exit the interface.

Edit Meas: Press this icon to enter into Measurement Edit Item. Choose the measurement result, which you want to display in the report. Customer also can edit the result.




Fig.5-17

Graph: Under OB report you can press this icon to enter in to see the fetal production graph.




Fig.5-18

Input Template: Input annotation from templates.






Fig.5-19

Chapter 6 Preset

6.1 Recall Preset



- a) Press  button on the touch screen to choose the probe type.
- b) Choose a clinical application.
- c) Choose a preset.

The default choice of the system is DEFAULT preset.


After entering into the system, if you want to change preset but do not want to quit the current interface, operate as follows:


1. Press Utility menu on the touch screen.
2. Now SK shortcut keys and touch screen are corresponding with following operation:
 - SK1—rotate SK1 to choose the clinical application.
 - SK2—rotate SK2 to choose the preset.
 - SK4—press SK4 to edit preset name.
 - SK6—press SK6 to save preset.
 - Load Preset on touch screen—press to load preset.

6.2 Save User Defined Preset

- 1) Choose and recall a preset first.
- 2) Adjust parameters to current preset.
- 3) Press Utility menu on the touch screen.
- 4) Rotate SK1 to choose clinical application name (skip this step if do not need adjustment).
- 5) Rotate SK2 to choose from User 1-User 5.
- 6) Press SK4 to edit preset name.
- 7) Press SK6 to save preset.

6.3 Manage Preset



Press  key on the touch screen to enter into system setting interface. Select preset page. Add or delete the current preset.




Fig.6-1

Name: input Application/Preset/User Defined name.

Exam Type: select clinical application.

Add: add Application/Preset.

Delete: delete an Application/Preset.

Modify: revise Application/Preset/User Defined name.

SetDefault: user can set the default preset of the probe on the first connector.

Reset: restore the preset to factory setting.

Import: import preset to system. Plug in U disk which has preset. Press Import key, preset can be imported into the system.

Export: export system preset. Plug in USB, press the Export key, system automatically derive the preset to U disk.

OK: exit system setting interface.

Chapter 7 System Setting

Press the  **System** key on the touch screen to enter system setting interface. User can do user-defined setting.

7.1 General

General Settings:

Set the hospital information, date-time, language etc.




Fig.7-1

Hospital Information	Hospital; Department;TelePhone	Input hospital name, department name, phone number
Name Format	To set the patient name format	
Screen Saver Setting	To set Screen Saver ON or OFF and Screen Saver wait time and import screen saver figures. Auto standby in low battery	
Fast Storage Setting	3 seconds/5 seconds/10 seconds fast storage Customize storage seconds	Please refer to 4.4.6.

	Manual storage Export to removable media Export to network media Export to PACS Number of frames saved after freezing Image Quality	
Image Option	Always show STC Curve Always show LGC Curve	Select it, it will always show STC Curve. Select it, it will always show LGC Curve.
Date-Time	Set Date Time Set Regional	Set Date and time (need password) Set Time Zone, Data Format, Time Format
Color Map	Set Color	Click the Configure key to set the color map.
Export Image Compensation	Normal Contrast Soft User Define	Set image parameters when exporting images
Language	Chinese;English; French; Spanish; Russian; Polish; Portuguese; Danish; German; Italian; Turkish; Romanian	Switch language interface.

Keyboard Settings




Fig.7-2

Function Setting	User1	Choose the function for User1.
	User3	Choose the function for User3.
	P1 Key	Choose the function for P1-key.
End Exam Option	Back to real-time scan; back to worklist; back to patient information	Set the system status after ending the exam.
Trackball Option	Sensitivity(1~20)	Set the sensitivity of the trackball option.
Touch screen options	Allow gestures sliding	
Print key: Print 1, Print 2, Touch Screen Print Button Option	Video Print; PC Print	Set the print-key function.
PC Print & Image Save Selection	Picture and info or Only Picture	Please refer to 3.6.3.
Print De-Identification		Tick to not to show the identification of the patient when printing the image.
Video Config	NTSC, PALD VIDEOEnable VGA Enable	Set the video standard. Select "Video Enable" to enable video port Select "VGA Enable" to enable VGA port
Angle Step		
Cine Store Option		
Video Print & Output Option	Only Picture Picture and Info FullScreen	Please refer to 3.6.1.
Helper Key Configuration		

Key Config:





Fig.7-3

Hot Key Config

Set the function for the hotkey 0~9 of Alphanumeric Keyboard. Select the Key Function that want to set, then select the Function and Measure for the selected key.



Function Config:




Fig.7-4

Items	Description
TIB in B/C Mode	Select it to display TIB in B/C mode
Changing the focus by trackball in B Mode	Select it to change the focus by trackball in B Mode
Show ParaSettingName	Select to show parasetting name
Keyboard Light	Set the brightness of the keyboard light.

UnFreezing Setting	Set the Unfreezing setting: Select to Clear BodyMark, Clear Comment, and Clear Arrow when Unfreezing the image.
Roi Setting	Select moving up to enlarge or minify ROI
Report Setting	Select to display Doctor Signature and Report Date in the report.
Dicom Setting	Select “Compressed single frame image”
Focus Setting	Select “FocusPosition changed by Depth”
Other	Select to enable environment light
Measure Related Settings	Select “Realtime Measurement”

4.3.70 Electronic Signature

This function is used to add electronic signature after the users finish editing the report.

Click [Signature] on the touchscreen to open the function.

7.2 Measurement

Set the measurement formula of measurement.

Edit Settings:

Click the measurement page, and click on the Edit page, then users can edit custom configuration measurement.




Fig.7-4

1. Package Name: Choose the different examinations.

2. Choose the different exam modes.
3. Choose the measurement item and click "UP" or "Down" button to move its position.
4. Add Folder: Click this button to add a new item.
5. Clicking "Add Meas." can add a new measurement.

General Settings:

Click the General page, and you can do some general settings of measurement items.




Fig.7-5

Using ellipse method to make the area measurement (only for fast Measure): choose to select the menu or not.

Font size: Set to default font size of the measurement results.

HR cycle options:

One: one cycle to measure heart rate in the Doppler envelope calculation.

Two: two cycles to measure heart rate in the Doppler envelope calculated, the result will be more accurate. 1~6 can be chosen.

Unit: set the unit of measurement results.

Always show measure result: select the menu to always show measure result on the screen until press CLEAR key.

Follicle: choose 3 Distances or 2 Distances.

OB Table Setting:

Click OB Table page, you can set the formula of the OB measurement items for measuring fetal

weight formula.




Fig.7-6

Set measurement Item: choose a measurement project, gestational age table or developmental table, and then choose a formula, click the information button, you can view the current formula.

Click New, you can create a new formula in the current directory.

Fetal Weight: Select the EFW equation, or EFW development table, and then choose a formula, click on the information button, you can view the information of current fetal weight formula. Click New, you can create a new fetal weight formula in the current directory.

7.3 Comment

Set the comment function, manage the comment Library.

Options Page:




Fig.7-7

Search option	Search comment or Abbreviations	Input search function Select the comment item, it will search notes when searching Select abbreviations, it will search abbreviations when searching
Font	Font size	Set the font size of notes

Library Page:

Edit and delete the annotation in the comment library.




Fig.7-8

Comment Libs	Provide the comment notes in different exam types.
Comment UserLibs	List the comment notes have been chosen from Libs and added by user.
Exam type in UserLibs	Establish the new exam type: click New near the drop-down list on the right to establish the new exam type.
	Edit the name of the exam type: click Edit near the drop-down list on the right to edit the name of current exam type displayed in the drop-down list box.
	Delete the exam type: click Delete near the drop-down list on the right to delete the current exam type displayed in the drop-down list box.
Comment notes	Add comment notes from Libs: select the comment note on the left, click “->” to add it on the right; click “=>” to add all comment notes on the left to the UserLibs.
	Add comment notes by user-defined: input the comment note in boxes after Abbreviation and Comment, and click Add to create the new comment note.
	Click delete, delete all to delete the selected note or all the notes on the right.
	Edit the comment note: select the comment note on the right, and input comment notes in boxes after Abbreviation and Comment, then click Edit to edit the selected comment note.
	Click Up, Down, Top, Bottm to adjust the sequence of the comment note selected.
OK	Click OK to save the edit and exit the system settings interface.

7.4 BodyMark



Fig.7-90

This function is to preset body marks in each preset or application of probes.

Libs: provide different body marks of different presets or applications of probes

UserLibs: list the body mark chosen from the Libs.

Establish the new category in the UserLibs:

Click the New icon near the drop-down list in the right to create a new category of body marks.

Click the Edit icon near the drop-down list in the right to change the name of current category displayed in the drop-down list box.

Click the Delete icon near the drop-down list in the right to delete the current category displayed in the drop-down list box.

Add or delete body mark:

Click  to add the selected item in the body mark library into User Body Mark.

Click  to add all items in the body mark library into User Body Mark.

Click Delete icon to delete the selected item on the right.

Click DeleteAll icon to delete all items on the right.

Click  to adjust the sequence of the selected items.

7.5 Report

Including the report design and report template

Customize settings:

Set the design of the report.




Fig.7-10

Report Name: select report type.

Left: Set left margins.

Right: Set right margins.

Top: Set top margins.

Bottom: Set bottom margins.

Sections: list all the sections which can be added.

Selected sections: Section used in current report.

All: Click this button to add "available section " to " selected section.

Add: Select the section that need to be added, then click this button, add selected section "selected section".

Delete: Select the section that need to be deleted, then click this button, delete the selected section from the "selected section".

Delete all: Click this button to empty the content of "selected section".

Section attribute: Set the display of section, you can select arrange horizontally or vertically.

Set rows and columns of the arrangement.

Select the "Ultrasound image", set the width and height of the image.

Save: After setting the current report, click the button, save the settings.

OK: Save and exit the page.

Template Page:




Fig.7-11

Modify template note: Select the template note which you want to change, input modification content in the right box

Add Folder: Press this icon to add a new Folder and name it.

Add Item: Press this icon to a new Item under current folder, and name it.

Delete: Press this icon to delete the template, which you selected.

Save: Save the changes.

General Page:




Fig.7-12

Logo Setting: Upload Logo and set logo size.

Image Layout: Choose different styles of image layout.

Header&Footer: Set the format of the header and footer.

OK: Save the changes and exit the system setting interface.



NOTE: the LOGO should be named with 'hLogo' in PNG format, which resolution is 168x169. And put LOGO file into the folder named 'hLogo'.

7.6 Admin




Fig.7-134



NOTE: If the HIPPA function is not open, the system setting interface will not display Admin page. Please make sure the HIPPA function is open before you use this function.

Enable Account Control: Select the "Enable Account Control" to enable Account Control. Unselect the "Enable Account Control" if the user doesn't want to use the Account Control.

After enable Account Control, the user must use user name and password to login the system every time the system turns on. The login interface see as below.




Fig.7-145

User Name: Select the user name which want to login.

Password: Input the password for user name.

Login: Click the icon to login the system.

Emergency: Login as “Emergency User”, no password required. The Emergency user only can review and manager the patient information that was created by Emergency User.

User can select the “Admin” to login first time, **the default password for Admin is 123456.**

The authority of login account decides the operation authority of the patient information.

Administrator user can review and manager all the patient information. Administrator user also cans manager the user account such as add or delete the accounts and change password for user accounts.

Customer user only can review and manager the patient that was created by user himself.

SonoFIS

SonoFIS is fingerprint recognition. It is used for user management.

7.7 Remote Service

7.8 Scan Code

Item	Start	End	Test
Patient ID	1	2	
First Name	3	5	
Last Name	6	7	
Middle Name	8	9	
Birthday	10	12	
Age(Year)	13	14	
Age(Month)	3	6	
Age(Day)	6	8	
Gender	4	4	Male 0 Female 1

The barcode reader helps users fill in the related patient information quickly in both Patient Information interface and Archive information interface. Connect the USB port of the barcode reader with the main unit to activate the code scanning function. Scan the barcode or the QR code. In the System Setting, users can set the scanning item in the Scan Code interface.

7.9 Network

Click the page of Network Settings to enter into the network settings interface.

Before setting up the network settings, plug the cable into the LAN port of the system.




Fig.7-15

Network Settings:

Network Adapte	Display the existing network cards on this system. The user can select the network card for current use.
Obtain an IP address automatically	When this option is selected, the system will obtain an IP address automatically.
Use the following IP address (recommended)	Set the IP address of the system manually. After setting, click Apply to confirm the application.
IP address	Input the IP address of the system manually, and make sure the IP address of the network printer is in the same network segment.
Subnet Mask	Input the subnet mask manually.
Default Gateway	Input the gateway manually.
DNS server	Input DNS server manually.
Network server Address	
Use Proxy Server	Select to use Proxy Server.
Address	Input the address information.
Port	Input the port information.
Network Connection Status	Check the network connection status.
Connected	The system is connected to the network successfully.
Unconnected	The system is not connected to the network.

Network Information	Display the current network information.
Config	Configuration information: Click this button to display the network configuration information.
Config All	Click this button to display network details.
Release	Release of the connection: Click this button to disconnect the current connection.
Renew	Renew of the connection: Click this button to update the current connection status.
Network test	In the case of network connection status is connected, enter the IP address of the test, click the test, you can test the network the connection status between the IP of current system and the IP that needs to be tested.

Network Storage:

With this feature, you can save the image to PC by network.

To use the network storage function, the user should connect the target computer to the network first, and build a new folder on the computer, right-click it to open the attribute, select shared page, set to share this folder on the network and allow network users to change the file (as shown below).



NOTE: the folder built on the computer is not allowed to sit on the desktop, otherwise create an error 67.




Fig.7-16

Click "Advanced Sharing"




Fig.7-17

Select "Share this folder", and click "Permissions" to modify user permissions.




Fig.7-18

Select "Full Control", "Change", "Read" for "Allow", click "OK" to complete modification.

Click Network Storage page, enter the network storage settings page. Then input shared directory name and IP address, and click Add.




Fig.7-20

Shared Dir: The shared directory name should be same as the shared folder name in target computer.

IP address: the IP address of the computer with the shared directory.

User name: input the name of computer account.

Password: input the password of computer account.

Ping: Click it test the communication between the shared PC and the system.

Add: add a network storage service.

Delete: can delete the selected network storage service.



NOTE: You can add multi-numbers of network storage service to realize the transmission among multi systems.

HL7:

HL7, Health Information Exchange Standard (Health Level 7), is a standardized health information transmission protocol, which is a protocol for electronic transmission between different applications in the medical field. HL7 brings together a standard format used by different vendors to design interfaces between applications, which will allow healthcare organizations to interact data between heterogeneous systems.

In the System Setting, choose [Network], and click [HL7] to make settings. Users can turn on/off HL7 function.

7.10 System Settings




Fig.7-21

Display the settings of system configuration information and functions

System information:

System SN: display the unit SN.

Build time: The build time of software.

Version: display the current software version and hardware version.

SW Upgrade: When the U-disk with the upgrade software is inserted to the system, click this button to upgrade software.

FW Upgrade: When the U-disk with the upgrade FPGA is inserted to the system, click this button to upgrade FPGA.

Remote Update

Remote Update function realizes a relatively convenient way to upgrade products in a timely manner through mobile communication networks (4G/5G or Wifi).

In brief, three steps are needed:

1. Establish the network connection
2. Download the upgrade package
3. Update

Function settings

Display the status of current function, click the “open” or “close” and input password to turn on or turn down the function.



NOTE: DICOM is option. If need to turn on this function, the user need to

input the password. Please contact the CHISON's authorized service engineer to get the corresponding password.

Software Lease: Key Input

It is time to the data time, the system cannot be entered for use. User can decode by pressing Key Input key and input the code which gets from CHISON's authorized service engineer.

Log: Export Log

Plug in U disk to export log files which saved on system automatically for engineers to analyze.

License: Export License

Plug in U disk to export license to U Disk.

Printer Manager: Users can manage the printers. Click "Open" to into the printer management screen.




Fig.7-22

Operating methods: insert U disk with installation driver to the system, click "Install Driver".




Fig.7-19

Select the driver, click “Install”, the system will jump out installation guide. Complete the installation by following up the guide.

The steps are as below:




Fig.7-20

Click “Yes”.




Fig.7-21

Select “Dynamic Mode”, click “Install”.




Fig.7-22

Be installed automatically.




Fig.7-23

Click "Finish" to finish the printer installation.

Add: Click "Add", complete adding printer according to operation guide.



Fig.7-24

Click "NEXT".




Fig.7-25

Select "Local printer attached to this computer", and Click "Next".




Fig.7-30

Select "Use the following port", and click "Next".




Fig.7-31

Select the printer, and click "Next".




Fig.7-32

Select "Keep existing driver (recommended)", and Click "Next".




Fig.7-26

Select "Yes", and Click "Next".




Fig.7-27

Select "Yes" to print a test page, and click "Next" to continue.




Fig.7-28

Click "Finish".

Chapter 8 Probes

8.1 General Description




Fig.8-1: Convex Probe Overview

Scan Orientation:

A logo symbol is used to depict the orientation of the probe corresponding to a physical mark on the probe. With image parameter Left-Right Inversion false (the default) the logo is at the far upper left of the image. With Left-Right Inversion true, the logo is at the far upper right.

The probes provide high spatial and contrast ultrasound imaging. These probes operate by pulsing sound waves into the body and listening to the returning echoes to produce high-resolution brightness mode, and a real time display.

8.2 Care and Maintenance

The probes that come with the system are designed to be durable and dependable. These precision instruments should be inspected daily and handled with care. Please observe the following precautions:

- Do not drop the transducer on hard surface. This can damage the transducer elements and compromise the electrical safety of the transducer.
- Avoid kinking or pinching the transducer cable.
- Use only approved ultrasonic coupling gels.
- Follow the instructions for cleaning and disinfecting that come with each probe.



- *Only person received professional training can use the probes.*
- *Probes can't receive pressure sterilizer, when operation in sterile area, disposable sterile probe hood should be applied.*
- *Make sure not to drop the transducer on hard surface. This can damage the transducer elements and compromise the electrical safety of the transducer.*
- *Be careful when doing operation. Make sure not to scratch the probe surface.*
- *Avoid kinking or pinching the transducer cable.*
- *Make sure not to connect the probe to plug or put adjacent cable into any kind of liquid.*
- *Keep the probe clean and dry. Power off or freeze when fixing or dismantling the probe.*
- *Make sure not to use or deposit the probe in the environment above 50 degree.*
- *If any abnormal phenomena of probe is found, immediately stop operation and contact with Sale Office, Customer Service department or Agents of manufacturer.*
- *No matter which type of examination is performed, please always try to reduce the unnecessary radiation of ultrasound wave to the patient during the ultrasound examination.*

8.2.1 Inspecting Probes

Before and after each use, inspect carefully the probe's lens, cable, casing, and connector. Look for any damage that would allow liquid to enter the probe. If any damage is suspected, do not use the probe until it has been inspected and repaired/replaced by a CHISON's authorized service engineer.



NOTE: *Keep a log of all probe maintenance, along with a picture of any probe malfunction.*



WARNING: *The probes are designed to be used only with this ultrasound system. Use of these probes on any other system or a non-qualified probe may cause electric shock or damage on the system/ transducer.*

8.2.2 Preparation at the Point of Use

Remove gross soiling incl. gel on the device using a paper towel wetted with cold water (<40°C) immediately after use, if applicable. Don't use a fixating detergent or hot water (>40°C) as this can cause the fixation of residuals which may influence the result of the reprocessing process.

Safely store the device in a humid surrounding and transport it to the reprocessing area to avoid any damage and contamination to the environment.

8.3 Cleaning and Disinfecting

Definitions

Cleaning removes visible soil (for example, organic and inorganic material) from the probe surfaces and normally is accomplished manually or mechanically using water with detergents or enzymatic products. Thorough cleaning is essential before high-level disinfection because inorganic and organic materials that remain on the surfaces of probes interfere with the effectiveness of these processes.

Disinfection describes a process that eliminates many or all pathogenic microorganisms, except bacterial spores.

Low-Level Disinfection (LLD) destroys most bacteria, some viruses, and some fungi. Low-level disinfection will not necessarily inactivate *Mycobacterium tuberculosis* or bacterial spores.

Intermediate-Level Disinfection (ILD) inactivates *Mycobacterium tuberculosis*, bacteria, most viruses, most fungi, and some bacterial spores.

High-Level Disinfection (HLD) destroys or removes all microorganisms except bacterial spores.

To reduce the risk of contamination and infection, it is important to choose the appropriate level of disinfection, based on prior exam usage and whether the use is classified as non-critical or semi-critical. Use table below named “Probe Disinfection Class, Use, and Method” to determine the appropriate class and then follow the appropriate intermediate-level or high-level disinfection procedure.

Probe Disinfection Class, Use, and Method

Class	Use	Method
Non-Critical Class	Touches intact skin	Cleaning followed by intermediate-level disinfection (ILD)
Semi-Critical Class	Touches mucous membranes and non-intact skin	Cleaning followed by high-level disinfection (HLD)

 WARNING

After each use of the probe, it must be cleaned. Cleaning the probe is an important step before effective disinfection. While using detergents and disinfectants, be sure to follow the manufacturer's label instructions.

 WARNING

When cleaning and disinfecting any instrument, be sure to wear protective glasses and gloves.

 CAUTION

Using non-recommended disinfectants, inappropriate disinfectant concentration, or the immersion

depth or immersion time of the probe exceeding the recommended value will cause damage or discoloration of the probe, and invalidate the probe warranty.



Do not use a brush when cleaning transducers. Even the use of soft brushes can damage transducers.



Do not leave the probe in contact with cleaning agents and disinfectants for a long time. Limit the time the probe is exposed to cleaners and disinfectants to the shortest time recommended by the manufacturer.



Before cleaning, check the probe first. If any particles or body fluids adhere to the probe or cable, you can wipe it off with a soft cloth dipped with 75% ethanol.

8.3.1 Surface probe as Non-critical

1. Soak a dry soft cloth with 75% ethanol or other disinfectants which are approved for its efficacy by FDA and/or Health Canada approval, VAH/DGHM-listing, CE marking, etc.
2. Wipe all the surfaces of the device, until the surfaces of the device are visually clean.
3. Repeat this step with a new soft cloth soaked in above-mentioned disinfectant.
4. Further repetition may be performed, if necessary.

Manual cleaning and immersion disinfecting method may also be used for the heads of these types, if required.

8.3.2 TV, TR probe as Semi-Critical

Manual Cleaning:

Following conditions of manual cleaning can be performed for the products: 0.5%-1% neodisher MediClean forte, 10-30 minutes immersion time at 30 °C.

If other cleaners are used, select immersion times, concentrations and temperatures of the solutions as instructed by the manufacturer of the cleaner.

1. Wear disposable protective gloves.
2. Put the device in a suitable cleaning basin.
3. Create 0.5%-1% neodisher MediClean forte.
4. Immerse the transducer head to the specified maximal depth in the detergent solution. Make sure that there are no air bubbles on the surfaces during immersion.
5. The transducer head is wiped under the detergent solution with a soft cloth for 30 seconds until all visible residues are removed. Then flush the transducer head with the detergent solution by using a

syringe. Repeat the step for 5 times. The cleaning process takes about 10 minutes.

6.Remove the device from the cleaning basin. Rinse the transducer head under running water for 30 seconds at room temperature.

7.Put the cleaned and rinsed device into a suitable clean basin with fresh de-ionized water. Thoroughly rinse and flush the device with the water for 5 minutes.

8.Dry the device through lint free towel.

Immersion Disinfecting:

It is recommended that you use Cidex® OPA Solution.

1. After cleaning the probe, you must disinfect the probe. It is recommended that you use Cidex® OPA high-level disinfection solution.

2. Prepare Cidex® OPA high-level disinfection solution for use per the manufacturer's instructions. Fill a tray or basin with the disinfectant solution at room temperature to a level allowing immersion of the probe up to the immersion line (As shown Fig.8-2).

3. Immerse the probe in Cidex® OPA solution up to the immersion line and ensure no air or bubbles are trapped at least 5 minutes.

4. Thoroughly rinse the probe with sterile water for at least one 1 minute

5. Thoroughly dry all surfaces of the device using a sterile lint-free cloth, changing wipes/cloths when necessary to ensure the device is completely dry. Inspect the device to ensure all surfaces are clean and dry. Repeat the drying steps if any moisture is visible.

6. Once clean and disinfected, inspect the probe, sheath, cable, and connector for signs of damage or wear.

7. Always store probes in the probe case when you are not using them.

The disinfection conditions described above are only recommended parameters, please refer to instructions for use of the disinfectant agent for actual use parameters and precautions.

8.3.3 TEE probe as Semi-Critical

Cleaning

Sterile purified lukewarm water shall be used (maximum 26°C), unless otherwise stated by the cleaner manufacturer.

Perform the following steps, unless otherwise stated by the manufacturer of the selected cleaning agent:

Rinse the soiled probe (1 minute).

Soak the probe as prescribed by the Cidezyme manufacturer of the cleaning agent. Take care that only the distal end up to the 100cm marker is immersed (70cm for MT4-6 probe).

- Gently scrub the probe using a soft-bristled brush (or wipes).
- Rinse the probe for at least one 1 minute until there is no sign of blood or soil.
- Dry the probe by means of a dry sterile wipe.

Disinfecting


Sterile purified lukewarm water shall be used (maximum 26°C), unless otherwise stated by the

manufacturer of the disinfection agent.

Perform the following steps, unless otherwise stated by the manufacturer of the disinfection agent:

- Rinse the contaminated probe for at least one 1 minute.
- Soak the probe for 12 minutes in Cidex OPA. Take care that only the distal end up to the 100cm marker is immersed (70cm for MT4-6 probe).
- Gently wipe the probe.
- Rinse the probe for at least one 1 minute
- Dry the probe by means of a dry sterile wipe.

Probe Immersion Levels






Fig.8-2: Probe Immersion Levels

8.4 Using protective sheaths

The use of market cleared probe sheaths is recommended for clinical applications.

Protective sheaths may be required to minimize disease transmission. Probe sheaths are available for use with all clinical situations where infection is a concern. Use of legally marketed, sterile probe sheaths is strongly recommended for endo-cavitary procedures.

DO NOT use pre-lubricated condoms as a sheath. In some cases, they can damage the probe. Lubricants in these condoms may not be compatible with probe construction.

DO NOT use an expired probe sheath. Before using a sheath, verify if it has expired.

Chapter 9 System Maintenance and Troubleshooting

9.1 Back up information



CAUTION: All patient data created is NOT backed-up! It is highly recommended to create a full system backup of patient data regularly and empty the hard disk (HDD), to ensure the hard disk (HDD) has never reached its maximum capacity.



NOTE: To Backup exams to USB DVD/CD+(R) W disk, please confirm that the USB DVD/CD+(R) W storage medium used is clean and not scratched!



WARNING: Do not disconnect an external USB pen drive without stopping it. Disconnecting without stopping can lead to data loss on the external device.

9.2 System Care and Maintenance

The system is a precise electrical device. To ensure the best performance and operation of the system, observe proper maintenance procedures. Contact the local CHISON's Authorized Service Representative for parts or periodic maintenance inspections.

Inspecting the System

Examine the following on a monthly basis:

- Connectors on cables for any mechanical defects.
- Entire length of electrical and power cords for cuts or abrasions.
- Equipment for loose or missing hardware.
- Control panel and keyboard for defects.

To avoid electric shock hazard, do not remove panels or covers from console. This servicing must be performed by CHISON's authorized service engineer. Failure to do so could serious injury.

If any defect is observed or malfunctions occur, do not operate the equipment but inform CHISON's authorized service engineer for information.

Weekly Maintenance

The system requires weekly care and maintenance to function safely and properly. Clean the following:

- LED monitor
- Operator control panel
- Printer

Cleaning the System

Prior to cleaning any part of the system, turn off the system power and disconnect the power cord.

See Section 3.4.4“Power Off” in Chapter 3 for more information.

Cleaning Method

- Moisten a soft, non-abrasive folded cloth.
- Wipe down the top, front, back, and both sides of the system.



Do not spray any liquid directly into the unit.

Do not use acetone/alcohol or abrasives on painted or plastic surfaces.

Cleaning LED Monitor

To clean the monitor face:

- Use a soft, folded cloth. Gently wipe the monitor face.
- Do NOT use a glass cleaner that has a hydrocarbon base (such as Benzene, Methyl Alcohol or Methyl Ethyl Ketone) on monitors with the filter (anti-glare shield). Hard rubbing will also damage the filter.



Cleaning Control Panel

- Moisten a soft, non-abrasive folded cloth with a mild, general purpose, non-abrasive soap and water solution.
- Wipe down operator control panel.
- Use a cotton swab to clean around keys or controls. Use a toothpick to remove solids from between keys and controls.



- *When cleaning the operator control panel, make sure not to spill or spray any liquid on the controls, into the system cabinet, or in the probe connection receptacle.*
- *DO NOT use Tspray or Sani Wipes on the control panel.*

Cleaning Printer

- Turn off the power. If possible, disconnect the power cord.
- Wipe the external surfaces of the unit with a soft, clean, dry cloth.
- Remove stubborn stains with a cloth lightly dampened with a mild detergent solution.



NOTE:

- ***Never use strong solvents, such as thinner or benzine, or abrasive cleansers because they will damage the cabinet.***
- ***No further maintenance, such as lubrication, is required.***
- ***For more information, see the Printer's Operation Manual.***

Inspection process in daily use

The SonoMax series have self-checking function to inspect the probe element. The users can use this function to test a regularly scheduled transducer quality assurance program or if users suspect that the transducer has been damaged.

Users can go to setup interface, choose transducer test, click the transducer name, click start test. The result will display on the screen.

“Test Passed” means all the elements work normally.

“Needs Evaluation” means that issues are detected.

Contact your local CHISON representative if you are unable to correct the problem.

9.3 Safety Check

To ensure the system work normally, please make a maintenance plan, check the safety of the system periodically. If there is any abnormal phenomenon with the machine, please contact our authorized agent in your country as soon as possible.

If there is no image or menu on the screen or other phenomenon appears after switching on the machine, please do troubleshooting first according to the following check list. If the trouble is still not solved, please contact our authorized agent in your country as soon as possible.

9.4 Troubleshooting

It is necessary to maintain the system regularly, as it can ensure the system being operated under safe state by eliminating possible trouble, and it can shorten the checking and repair period, lower the service costs and reduce the operation danger.

If you have any difficulty with the system, use the following information for your reference to help correct the problem. For a problem not covered here, contact your local distributor or Manufacturer.

Symptom	Solution
---------	----------

The system can't power on	1) Check power connections, e.g. power cord connection on the rear panel; 2) Check the fuse: if it is burnt due to mains fluctuation, use spare fuse for replacement.
When starting the system, the monitor has signal but no ultrasound image	Switch off the system, and check probe connection.
System image quality is not good	1) Adjust the LED monitor position for a better viewing angle; 2) Adjust the brightness and contrast of LED monitor; 3) Adjust the image parameters, e.g. Gain, Dynamic range.
No OB calculation package menu	Select the OB application before scanning.
PRINT-key doesn't work	1) Check if the approved printer is connected; 2) Check if the printer power is on; 3) Check printer connection; 4) Check printer setting in system setup.
External monitor doesn't work	1) Check the monitor connections; 2) Check if the monitor power is on and is set up correctly.
CFM or PW Doppler image has noise	1) Adjust CFM or PW gain value properly; 2) Check if there is appliance or equipment resulting in strong electromagnetic interference
Image has interference	1) Move or avoid interference source; 2) Use separate power outlet; 3) Perform good ground protection
The gray scale is S-twisted in the image area	Adjust the power supply to normal voltage or use a voltage stabilizer
The date and time on the screen is not correct	Press Setup-key to display General Setting screen, and correct time and date.
The Video printer is not work	1) Please confirm the signal cable, Remote cable are connected well. 2) Please make sure you have finished the setting for Video printer at system interface. 3) If you can't change the setting at system interface, please check whether the video printer is turned on and connected well with the main unit. 4) Please make sure the switch on the rear panel of printer is "on" status.

9.5 Service Responsibility

If users install, use and maintain the system fully according to CHISON exchange the momanual, operation manual and service manual, then main unit has a life time of 10 years and probes have life time of 10 years after ex-work (except that the 4D probe life time is 18 months).

The warranty of the system and probes after ex-work is as the time in the warranty card.

The system is a precise electronic system. Only the CHISON's authorized service engineer could replace the defective parts. Any assembly, disassembly, handling, repair, or replacement by any other people may have adverse impact on the safety and effectiveness of the systems and transducers, and thus will reduce the life time of the system and probes, and such systems and probes will not be covered by CHISON warranty after the above improper handling. Standard maintenance must be performed by CHISON's authorized service engineer during the life time of the product.



CAUTION: When the above life time is expired, the effectiveness and safety of system and probes maybe greatly affected, so it's NOT suggested to continue using the system and probes even the system and probes seem work properly. But if user still wants to continue using the system and probes, user should first contact CHISON service center at CHISON headquarter to arrange the necessary safety check and calibration by CHISON's authorized service engineer. If CHISON headquarter service center provides the calibration certificate for the related system or probe, then user could continue use the system or probes according to the calibration certificate. However, if CHISON headquarter service center concludes that the system or probe is no longer complied to the safety and effectiveness standard, then user should immediately stop using the system or probe. User understands that such check and calibration cost will be born by the user.

Systems and probes keep on using after the life time may also be difficult to repair and maintain, so it's suggested to renew the product after the life time.

REFERENCE:

1. AIUM: Medical Ultrasound Safety, American Institute of Ultrasound in Medicine, Laurel, MD, 1994. NOTE: As of this writing an update of this publication is being finalized by the AIUM. AIUM: Acoustic Output Labeling Standard for Diagnostic Ultrasound Equipment: A Standard for How Manufacturers Should Specify Acoustic Output Data, Revision 1, American Institute of Ultrasound in Medicine, Laurel, MD, 2008.
2. AIUM/NEMA: Standard For Real-Time Display of Thermal and Mechanical Acoustic Output Indices On Diagnostic Ultrasound Equipment, Revision 2. NEMA Standards Publication UD 3-2004; American Institute of Ultrasound in Medicine, Laurel MD; National Electrical Manufacturers Association, Rosslyn, VA; 2004a.
3. AIUM/NEMA: Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment, Revision 3. NEMA Standards Publication UD 2-2004; American Institute of Ultrasound in Medicine, Laurel, MD; National Electrical Manufacturers Association, Rosslyn, VA; 2004b.
4. Implementation of the Principle of As Reasonably Achievable (ALARA) for Medical and Dental Personnel, National Council on Radiation Protection and Measurements (NCRP), report NO.107, December 31, 1990
5. FDA Center for Devices and radiological Health (CDRH), 510(K) Guidance for Diagnostic Ultrasound and Fetal Doppler Ultrasound Medical Devices, September 8 1989 draft
6. FDA/CDRH,510(K) Diagnostic Ultrasound Guidance Update of 1991, April 26, 1991 draft
7. Biological Effects of Ultrasound: Mechanisms and Clinical Implications, NCRP Report No. 74, December 30,1983
8. Exposure Criteria for Medical Diagnostic Ultrasound: I. Criteria Based on Thermal Mechanisms, NCRP Report No.113, June 1,1992
9. Bioeffects Considerations for the safety of Diagnostic Ultrasound, Journal of Ultrasound in Medicine, AIUM, September1988
10. Geneva Report on Safety and Standardization in Medical Ultrasound, WFUMB, May 1990 Medical Ultrasound Safety, AIUM, 1994
11. Medical Electrical Equipment standard IEC 60601-1, IEC60601-1-2, IEC 60601-2-37
12. Diagnostic Ultrasound Physics and Equipment, edit by P. R. Hoskins, in 2003

APPENDIX A The Information of EU Representative

R Sight B.V.

Address: Roald Dahlilaan 47, 5629 MC, Eindhoven. The Netherlands

SRN: NL-AR-000010445

[Tel:0031640845545](tel:0031640845545)

CCI No.:76704726


E-mail:info@rsight.nl

APPENDIX B System one-key-recovery Function

This system has one-key-recovery function. Use can use this function to recovery system when the system has problems.

Detailed operations are as follows:

1. Press  button to boot the system.
2. Press <F3> button when the system enter into the following boot screen.



3. When the recovery system starts, the following interface will pop up.



Export System ID: export system ID to the removeable storage device.

Recovery system: restore machine to the factory version.


Export Liense: export license to the removeable storage device.

Recover Ultrasound: restore ultrasound software to the factory verison.

Reboot: click to boot the system again.

Shutdown: click to close the system.


4. After click the "Recover System", the following interface will pop up:



Click “Yes” to confirm and start system recovery. The system will start to backup ultrasound system data, after finishing the backup, the system will transfer to the recovery procedure automatically.

Click “No” to cancel system recovery.

5. After click the “Recover Ultrasound”, the following interface will pop up:



Click “OK” to confirm and start ultrasound software recovery.

Click “Cancel” to cancel ultrasound software recovery.

6. After finishing the recovery, the system will pop the CMD interface. Input EXIT in the command line to reboot the system will reboot.

APPENDIX C Maximun Acoustic Output Report

Transducer Model: **C1-5**Operating Mode: **B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	1.0		1.5		–
Index component value				1.0	1.0	1.5	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.31					
	P	mW		149.35		149.35		–
	$P_{1 \times 1}$	mW		59.32		59.32		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	6.57					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	3.52	3.54		3.54		–
Other Information	prr	Hz	9685					
	srr	Hz	75.66					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	13.33					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	65.86					
	p_r at Z_{pii}	Mpa	–					
Operating conditions	Focus	cm	7.0	7.0		7.0		–
	Depth	cm	20.45	20.45		20.45		–
	Scanning width	%	100	100		100		–
	Freq	MHz	3.5	3.5		3.5		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-5**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.6		1.4		—
Index component value				0.6	0.6	1.4	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.12					
	P	mW		145.78		145.78		—
	$P_{1 \times 1}$	mW		35.69		35.69		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.25					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.50	3.53		3.53		—
Other Information	p_{rr}	Hz	12000					
	s_{rr}	Hz	93.75					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.54					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	70.42					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	6.0	8.0		8.0		—
	Depth	cm	11.82	9.36		9.36		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-5

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	1.0		2.0		–
Index component value				1.0	0.7	2.0	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.75					
	P	mW		205.55		199.52		–
	$P_{1 \times 1}$	mW		58.76		55.49		
	Z_s	cm			4.12			
	Z_b	cm					6.40	
	Z_{MI}	cm	4.45					
	$z_{pii,\alpha}$	cm	4.45					
	f_{awf}	MHz	3.51	3.53		3.54		–
Other Information	pr	Hz	12000					
	srr	Hz	–					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	131.52					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	125.54					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	369.29					
	p_r at z_{pii}	Mpa	1.56					
Operating conditions	Focus	cm	6.0	7.0		10.0		–
	Depth	cm	9.36	8.13		11.82		–
	Scanning width	%	100	100		100		–
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-5**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.3		0.7		—
Index component value				0.6	1.3	0.6	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.50					
	P	mW		155.49		155.49		—
	$P_{1 \times 1}$	mW		38.44		38.44		
	Z_s	cm			4.23			
	Z_b	cm					5.14	
	Z_{MI}	cm	5.56					
	$Z_{pii,\alpha}$	cm	5.56					
	f_{awf}	MHz	3.52	3.55		3.56		—
Other Information	p_{rr}	Hz	4846					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	118.25					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	115.54					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	446.55					
Operating conditions	p_r at Z_{pii}	Mpa	2.64					
	Focus	cm	7.0	6.0		6.0		—
	Depth	cm	8.13	6.89		6.89		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-5

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.2		1.6		—
Index component value				1.2	1.2	1.6	1.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.51					
	P	mW		162.33		162.33		—
	$P_{1 \times 1}$	mW		70.05		70.05		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.62					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.58	3.60		3.60		—
Other Information	p_{rr}	Hz	9685					
	s_{rr}	Hz	75.66					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	13.33					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	65.86					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	7.0	7.0		7.0		—
	Depth	cm	20.45	20.45		20.45		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-6**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	1.1		1.5		—
Index component value				1.1	1.1	1.5	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.66					
	P	mW		152.20		152.20		—
	$P_{1 \times 1}$	mW		66.96		66.96		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.43					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.42	3.45		3.45		—
Other Information	p_{rr}	Hz	9762					
	s_{rr}	Hz	76.27					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	13.44					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	61.39					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	7.0	7.0		7.0		—
	Depth	cm	20.45	20.45		20.45		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-6**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.6		1.4		—
Index component value				0.6	0.6	1.4	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.93					
	P	mW		146.52		146.52		—
	$P_{1 \times 1}$	mW		36.42		36.42		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.20					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.43	3.46		3.46		—
Other Information	pr	Hz	12000					
	srr	Hz	93.75					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.45					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	67.13					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	6.0	8.0		8.0		—
	Depth	cm	11.82	9.36		9.36		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-6**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.9		1.9		—
Index component value				0.9	0.8	1.9	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.74					
	P	mW		200.14		197.66		—
	$P_{1 \times 1}$	mW		56.41		55.42		
	Z_s	cm			3.99			
	Z_b	cm					6.41	
	Z_{MI}	cm	4.38					
	$z_{pii,\alpha}$	cm	4.38					
	f_{awf}	MHz	3.42	3.46		3.47		—
Other Information	pr	Hz	12000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	136.26					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	124.76					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	351.13					
	p_r at z_{pii}	Mpa	1.56					
Operating conditions	Focus	cm	6.0	7.0		10.0		—
	Depth	cm	9.36	8.13		11.82		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-6**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.0	1.2		0.8		—
Index component value				0.7	1.2	0.7	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.86					
	P	mW		168.75		168.75		—
	$P_{1 \times 1}$	mW		42.24		42.24		
	Z_s	cm			4.19			
	Z_b	cm					5.16	
	Z_{MI}	cm	5.63					
	$Z_{pii,\alpha}$	cm	5.63					
	f_{awf}	MHz	3.46	3.48		3.49		—
Other Information	p_{rr}	Hz	4872					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	119.52					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	116.10					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	445.89					
	p_r at Z_{pii}	<i>Mpa</i>	2.56					
Operating conditions	Focus	cm	7.0	6.0		6.0		—
	Depth	cm	8.13	6.89		6.89		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-6**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.9		1.6		—
Index component value				0.9	0.9	1.6	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.49					
	P	mW		165.23		165.23		—
	$P_{1 \times 1}$	mW		54.31		54.31		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.51					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.47	3.48		3.48		—
Other Information	p_{rr}	Hz	9856					
	s_{rr}	Hz	77.01					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	13.44					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	61.39					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	7.0	7.0		7.0		—
	Depth	cm	20.45	20.45		20.45		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.9		1.5		—
Index component value				0.9	0.9	1.5	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.49					
	P	mW		153.29		153.29		—
	$P_{1 \times 1}$	mW		54.11		54.11		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.63					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.48	3.50		3.50		—
Other Information	p_{rr}	Hz	9721					
	s_{rr}	Hz	75.95					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	9.88					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	48.63					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	7.0	7.0		7.0		—
	Depth	cm	20.45	20.45		20.45		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.8		1.4		—
Index component value				0.8	0.8	1.4	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.12					
	P	mW		146.52		146.52		—
	$P_{1 \times 1}$	mW		48.14		48.14		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.20					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.47	3.49		3.49		—
Other Information	pr	Hz	12000					
	srr	Hz	93.75					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	10.08					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	44.55					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	6.0	8.0		8.0		—
	Depth	cm	11.82	9.36		9.36		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.0		2.0		–
Index component value				1.0	0.7	2.0	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.11					
	P	mW		207.43		201.12		–
	$P_{1 \times 1}$	mW		58.44		55.43		
	Z_s	cm			4.11			
	Z_b	cm					6.30	
	Z_{MI}	cm	4.51					
	$z_{pii,\alpha}$	cm	4.51					
	f_{awf}	MHz	3.45	3.47		3.49		–
Other Information	p_{rr}	Hz	12000					
	s_{rr}	Hz	–					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	130.25					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	117.22					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	343.39					
	p_r at z_{pii}	<i>Mpa</i>	1.43					
Operating conditions	Focus	cm	6.0	7.0		10.0		–
	Depth	cm	9.36	8.13		11.82		–
	Scanning width	%	100	100		100		–
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.7		1.1		—
Index component value				1.0	1.7	1.0	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.49					
	P	mW		240.16		246.16		—
	$P_{1 \times 1}$	mW		60.69		60.69		
	Z_s	cm			4.20			
	Z_b	cm					5.05	
	Z_{MI}	cm	5.57					
	$Z_{pii,\alpha}$	cm	5.57					
	f_{awf}	MHz	3.45	3.46		3.48		—
Other Information	p_{rr}	Hz	4820					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	117.62					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	108.34					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	408.59					
	p_r at Z_{pii}	Mpa	2.66					
Operating conditions	Focus	cm	7.0	6.0		6.0		—
	Depth	cm	8.13	6.89		6.89		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.0		1.6		—
Index component value				1.0	1.0	1.6	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.49					
	P	mW		162.30		162.30		—
	$P_{1 \times 1}$	mW		60.11		60.11		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.53					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.48	3.50		3.50		—
Other Information	pr	Hz	9785					
	srr	Hz	76.44					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	9.88					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	48.63					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	7.0	7.0		7.0		—
	Depth	cm	20.45	20.45		20.45		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7P

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.9		1.5		—
Index component value				0.9	0.9	1.5	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.49					
	P	mW		148.52		148.52		—
	$P_{1 \times 1}$	mW		53.69		53.69		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.57					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.49	3.52		3.52		—
Other Information	p_{rr}	Hz	9812					
	s_{rr}	Hz	76.66					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	8.89					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	43.33					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	7.0	7.0		7.0		—
	Depth	cm	20.45	20.45		20.45		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7P**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.8		1.4		—
Index component value				0.8	0.8	1.4	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.12					
	P	mW		146.55		146.55		—
	$P_{1 \times 1}$	mW		47.73		47.73		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.23					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.50	3.52		3.52		—
Other Information	pr	Hz	12000					
	srr	Hz	93.75					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	18.44					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	83.16					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	6.0	8.0		8.0		—
	Depth	cm	11.82	9.36		9.36		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7P**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.9		1.9		–
Index component value				0.9	0.9	1.9	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.13					
	P	mW		200.01		196.52		–
	$P_{1 \times 1}$	mW		53.24		51.24		
	Z_s	cm			4.13			
	Z_b	cm					6.40	
	Z_{MI}	cm	4.43					
	$z_{pii,\alpha}$	cm	4.43					
	f_{awf}	MHz	3.52	3.54		3.55		–
Other Information	pr	Hz	12000					
	srr	Hz	–					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	132.42					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	130.11					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	382.05					
	p_r at z_{pii}	Mpa	1.54					
Operating conditions	Focus	cm	6.0	7.0		10.0		–
	Depth	cm	9.36	8.13		11.82		–
	Scanning width	%	100	100		100		–
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7P

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.6		0.7		—
Index component value				0.6	1.6	0.6	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.50					
	P	mW		162.44		162.44		—
	$P_{1 \times 1}$	mW		35.49		35.49		
	Z_s	cm			4.26			
	Z_b	cm					5.18	
	Z_{MI}	cm	5.64					
	$Z_{pii,\alpha}$	cm	5.64					
	f_{awf}	MHz	3.52	3.55		3.54		—
Other Information	p_{rr}	Hz	4778					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	120.12					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	107.46					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	423.48					
	p_r at Z_{pii}	<i>Mpa</i>	2.64					
Operating conditions	Focus	cm	7.0	6.0		6.0		—
	Depth	cm	8.13	6.89		6.89		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C1-7P**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.9		1.6		–
Index component value				0.9	0.9	1.6	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.50					
	P	mW		162.24		162.24		–
	$P_{1 \times 1}$	mW		53.24		53.24		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	6.61					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	3.53	3.55		3.55		–
Other Information	p_{rr}	Hz	9862					
	s_{rr}	Hz	77.06					
	n_{pps}		–	–				
	$l_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$l_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	8.89					
	l_{spta} at z_{pii} or z_{sii}	mW/cm ²	43.33					
	p_r at z_{pii}	Mpa	–					
Operating conditions	Focus	cm	7.0	7.0		7.0		–
	Depth	cm	20.45	20.45		20.45		–
	Scanning width	%	100	100		100		–
	Freq	MHz	3.5	3.5		3.5		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-10**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.6		0.7		—
Index component value				0.6	0.6	0.7	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.46					
	P	mW		31.26		31.26		—
	$P_{1 \times 1}$	mW		16.84		16.84		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.26					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.47	7.48		7.48		—
Other Information	p_{rr}	Hz	2953					
	s_{rr}	Hz	23.07					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	18.41					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	59.08					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-10**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.9		1.1		—
Index component value				0.9	0.9	1.1	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.78					
	P	mW		48.12		48.12		—
	$P_{1 \times 1}$	mW		29.21		29.21		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.63					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.45	6.47		6.47		—
Other Information	pr	Hz	20000					
	srr	Hz	156.25					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	22.10					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	45.69					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	2.5	1.0		1.0		—
	Depth	cm	4.93	3.70		3.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/6.5	7.5/6.5		7.5/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-10**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.6		1.1		—
Index component value				0.6	0.5	1.1	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.27					
	P	mW		35.81		47.18		—
	$P_{1 \times 1}$	mW		20.22		24.12		
	Z_s	cm						
	Z_b	cm						
	Z_{MI}	cm	1.75					
	$Z_{pii,\alpha}$	cm	1.75					
	f_{awf}	MHz	6.44	6.45		6.47		—
Other Information	p_{rr}	Hz	20000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	180.15					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	205.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	446.77					
	p_r at Z_{pii}	<i>Mpa</i>	2.26					
Operating conditions	Focus	cm	4.0	2.5		2.5		—
	Depth	cm	7.39	3.70		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/6.5	7.5/6.5		7.5/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-10

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.3		1.0		—
Index component value				0.9	1.3	0.9	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	2.18					
	P	mW		50.12		52.11		—
	$P_{1 \times 1}$	mW		25.43		26.41		
	Z_s	cm			1.80			
	Z_b	cm					1.86	
	Z_{MI}	cm	1.68					
	$Z_{pii,\alpha}$	cm	1.68					
	f_{awf}	MHz	7.42	7.43		7.45		—
Other Information	p_{rr}	Hz	2915					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	215.39					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	233.15					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	551.57					
	p_r at Z_{pii}	<i>Mpa</i>	2.62					
Operating conditions	Focus	cm	2.0	1.5		2.0		—
	Depth	cm	3.70	3.70		3.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-10

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.7		0.9		—
Index component value				0.7	0.7	0.9	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.45					
	P	mW		38.14		38.14		—
	$P_{1 \times 1}$	mW		19.76		19.76		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.95					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.43	7.44		7.44		—
Other Information	p_{rr}	Hz	3012					
	s_{rr}	Hz	23.53					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	18.41					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	59.08					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-18

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.2		0.4		–
Index component value				0.2	0.2	0.4	0.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.20					
	P	mW		13.52		13.52		–
	$P_{1 \times 1}$	mW		5.58		5.58		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	1.85					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	9.85	9.85		9.85		–
Other Information	p_{rr}	Hz	5420					
	s_{rr}	Hz	42.34					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	22.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	78.20					
	p_r at Z_{pii}	Mpa	–					
Operating conditions	Focus	cm	2.5	2.5		2.5		–
	Depth	cm	4.93	4.93		4.93		–
	Scanning width	%	100	100		100		–
	Freq	MHz	10.0	10.0		10.0		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-18

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.5		0.9		–
Index component value				0.5	0.5	0.9	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.26					
	P	mW		33.15		33.15		–
	$P_{1 \times 1}$	mW		12.22		12.22		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	1.85					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	9.88	9.90		9.90		–
Other Information	p_{rr}	Hz	20000					
	s_{rr}	Hz	156.25					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	24.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	85.22					
	p_r at Z_{pii}	Mpa	–					
Operating conditions	Focus	cm	2.5	2.5		2.5		–
	Depth	cm	9.86	8.62		8.62		–
	Scanning width	%	100	100		100		–
	Freq	MHz	10.0/10.0	10.0/10.0		10.0/10.0		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-18

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.3	0.7		0.9		—
Index component value				0.6	0.7	0.9	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.94					
	P	mW		25.53		35.52		—
	$P_{1 \times 1}$	mW		12.22		18.22		
	Z_s	cm			1.82			
	Z_b	cm					1.82	
	Z_{MI}	cm	3.15					
	$Z_{pii,\alpha}$	cm	3.15					
	f_{awf}	MHz	9.86	9.87		9.87		—
Other Information	pr	Hz	20000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	118.45					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	115.41					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	986.31					
	p_r at Z_{pii}	<i>Mpa</i>	0.77					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	6.16	6.16		6.16		—
	Scanning width	%	100	100		100		—
	Freq	MHz	10.0/10.0	10.0/10.0		10.0/10.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-18

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.3		1.5		—
Index component value				1.1	1.3	1.1	1.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.89					
	P	mW		48.55		6.21		—
	$P_{1 \times 1}$	mW		23.28		20.24		
	Z_s	cm			1.85			
	Z_b	cm					1.85	
	Z_{MI}	cm	1.75					
	$Z_{pii,\alpha}$	cm	1.75					
	f_{awf}	MHz	9.90	9.92		9.93		—
Other Information	p_{rr}	Hz	5422					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	172.45					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	122.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	404.44					
	p_r at Z_{pii}	Mpa	1.77					
Operating conditions	Focus	cm	2.5	2.5		2.0		—
	Depth	cm	4.93	4.93		6.16		—
	Scanning width	%	100	100		100		—
	Freq	MHz	10.0	10.0		10.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-18

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.3		0.5		—
Index component value				0.3	0.3	0.5	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.26					
	P	mW		18.45		18.45		—
	$P_{1 \times 1}$	mW		7.88		7.88		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.83					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	9.93	9.93		9.93		—
Other Information	p_{rr}	Hz	5462					
	s_{rr}	Hz	42.67					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	22.21					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	78.20					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	10.0	10.0		10.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L11-20**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.2		0.4		—
Index component value				0.2	0.2	0.4	0.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.78					
	P	mW		14.51		14.51		—
	$P_{1 \times 1}$	mW		3.67		3.67		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.83					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	15.80	15.80		15.80		—
Other Information	p_{rr}	Hz	5420					
	s_{rr}	Hz	42.34					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.44					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	113.78					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	18.0	18.0		18.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L11-20

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.4		0.9		—
Index component value				0.4	0.4	0.9	0.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.39					
	P	mW		33.14		33.14		—
	$P_{1 \times 1}$	mW		6.48		6.48		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.95					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	15.85	15.88		15.88		—
Other Information	pr	Hz	20000					
	srr	Hz	156.25					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	22.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	186.99					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	9.86	8.62		8.62		—
	Scanning width	%	100	100		100		—
	Freq	MHz	18.0/18.0	18.0/18.0		18.0/18.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L11-20**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.2	1.1		0.8		—
Index component value				0.9	1.1	0.6	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.82					
	P	mW		17.25		20.22		—
	$P_{1 \times 1}$	mW		11.48		12.24		
	Z_s	cm			1.94			
	Z_b	cm					1.94	
	Z_{MI}	cm	3.45					
	$Z_{pii,\alpha}$	cm	3.45					
	f_{awf}	MHz	16.91	16.92		16.92		—
Other Information	pr	Hz	20000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	120.45					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	102.10					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	5743.95					
	p_r at Z_{pii}	Mpa	0.85					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	6.16	6.16		6.16		—
	Scanning width	%	100	100		100		—
	Freq	MHz	18.0/18.0	18.0/18.0		18.0/18.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L11-20

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	1.3		1.7		—
Index component value				1.2	1.3	1.2	1.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	2.06					
	P	mW		28.99		30.24		—
	$P_{1 \times 1}$	mW		14.87		16.01		
	Z_s	cm			1.83			
	Z_b	cm					1.83	
	Z_{MI}	cm	1.96					
	$Z_{pii,\alpha}$	cm	1.96					
	f_{awf}	MHz	16.92	16.94		16.95		—
Other Information	p_{rr}	Hz	5462					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	185.12					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	101.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	998.45					
	p_r at Z_{pii}	<i>Mpa</i>	1.77					
Operating conditions	Focus	cm	2.5	2.5		2.0		—
	Depth	cm	4.93	4.93		6.16		—
	Scanning width	%	100	100		100		—
	Freq	MHz	18.0	18.0		18.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L11-20**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.3		0.5		—
Index component value				0.3	0.3	0.5	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.64					
	P	mW		18.20		18.20		—
	$P_{1 \times 1}$	mW		4.88		4.88		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.86					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	16.89	16.89		16.89		—
Other Information	pr	Hz	5422					
	srr	Hz	42.35					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.44					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	113.78					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	18.0	18.0		18.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15B

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.6		0.8		—
Index component value				0.6	0.6	0.8	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.64					
	P	mW		32.56		32.56		—
	$P_{1 \times 1}$	mW		14.63		14.63		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.65					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	8.59	8.61		8.61		—
Other Information	p_{rr}	Hz	3256					
	s_{rr}	Hz	25.44					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	10.02					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	48.28					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	9.0	9.0		9.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15B**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.9		1.1		—
Index component value				0.9	0.9	1.1	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.30					
	P	mW		48.52		48.52		—
	$P_{1 \times 1}$	mW		27.79		27.79		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.68					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.78	6.80		6.80		—
Other Information	pr	Hz	20000					
	srr	Hz	156.25					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	20.21					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	44.49					
	p_r at z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	2.5	1.0		1.0		—
	Depth	cm	4.93	3.70		3.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	9.0/6.5	9.0/6.5		9.0/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15B**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.7		1.1		—
Index component value				0.7	0.6	1.1	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.83					
	P	mW		33.25		48.25		—
	$P_{1 \times 1}$	mW		20.12		24.22		
	Z_s	cm			1.90			
	Z_b	cm					1.90	
	Z_{MI}	cm	1.86					
	$z_{pii,\alpha}$	cm	1.86					
	f_{awf}	MHz	6.81	6.83		6.85		—
Other Information	p_{rr}	Hz	20000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	182.30					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	181.24					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	434.76					
	p_r at z_{pii}	<i>Mpa</i>	2.26					
Operating conditions	Focus	cm	4.0	2.5		2.5		—
	Depth	cm	7.39	3.70		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	9.0/6.5	9.0/6.5		9.0/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15B

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	1.8		1.5		—
Index component value				1.4	1.8	1.4	1.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.65					
	P	mW		66.45		68.54		—
	$P_{1 \times 1}$	mW		33.87		34.52		
	Z_s	cm			1.86			
	Z_b	cm					1.86	
	Z_{MI}	cm	1.72					
	$Z_{pii,\alpha}$	cm	1.72					
	f_{awf}	MHz	8.67	8.68		8.69		—
Other Information	p_{rr}	Hz	3015					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	215.63					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	222.24					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	622.58					
	p_r at Z_{pii}	Mpa	3.05					
Operating conditions	Focus	cm	2.0	1.5		2.0		—
	Depth	cm	3.70	3.70		3.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	9.0	9.0		9.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15B

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.7		0.9		—
Index component value				0.7	0.7	0.9	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.37					
	P	mW		38.56		38.56		—
	$P_{1 \times 1}$	mW		16.70		16.70		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.12					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	8.78	8.80		8.80		—
Other Information	p_{rr}	Hz	3120					
	s_{rr}	Hz	23.60					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	10.02					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	48.28					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	9.0	9.0		9.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-14**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.5		0.9		—
Index component value				0.5	0.5	0.9	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.51					
	P	mW		37.18		37.18		—
	$P_{1 \times 1}$	mW		15.41		15.41		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.23					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.78	7.80		7.80		—
Other Information	p_{rr}	Hz	2915					
	s_{rr}	Hz	22.77					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.55					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	51.55					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	3.0		3.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-14**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.7		1.4		—
Index component value				0.7	0.7	1.4	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.04					
	P	mW		55.24		55.24		—
	$P_{1 \times 1}$	mW		22.54		22.54		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.60					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.80	6.82		6.82		—
Other Information	p_{rr}	Hz	20000					
	s_{rr}	Hz	156.25					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	20.21					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	42.85					
	p_r at z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	2.0	1.5		1.5		—
	Depth	cm	4.93	3.70		3.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0/6.5	8.0/6.5		8.0/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-14**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.8		1.0		—
Index component value				0.8	0.7	1.0	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.55					
	P	mW		40.12		40.12		—
	$P_{1 \times 1}$	mW		22.10		22.10		
	Z_s	cm			1.74			
	Z_b	cm					1.80	
	Z_{MI}	cm	1.73					
	$Z_{pii,\alpha}$	cm	1.73					
	f_{awf}	MHz	6.68	6.70		6.72		—
Other Information	p_{rr}	Hz	20000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	234.10					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	182.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	404.82					
	p_r at Z_{pii}	<i>Mpa</i>	2.26					
Operating conditions	Focus	cm	3.5	3.5		3.5		—
	Depth	cm	7.39	7.39		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0/6.5	8.0/6.5		8.0/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-14

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	1.1		0.7		—
Index component value				0.6	1.1	0.6	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.94					
	P	mW		38.41		38.41		—
	$P_{1 \times 1}$	mW		18.41		18.41		
	Z_s	cm			1.88			
	Z_b	cm					1.92	
	Z_{MI}	cm	2.01					
	$Z_{pii,\alpha}$	cm	2.01					
	f_{awf}	MHz	7.67	7.68		7.70		—
Other Information	p_{rr}	Hz	2952					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	250.46					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	150.22					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	435.34					
	p_r at Z_{pii}	<i>Mpa</i>	2.91					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	11.09	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-14

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.4		0.8		—
Index component value				0.4	0.4	0.8	0.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.24					
	P	mW		31.26		31.26		—
	$P_{1 \times 1}$	mW		12.24		12.24		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.13					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.83	7.85		7.85		—
Other Information	p_{rr}	Hz	3025					
	s_{rr}	Hz	23.63					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.55					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	51.55					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	3.0		3.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.0	1.1		2.6		2.6
Index component value				1.1	1.1	2.6	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.55					
	P	mW		182.45		182.45		182.45
	$P_{1 \times 1}$	mW		95.45		95.45		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.58					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	2.40	2.42		2.42		2.42
Other Information	p_{rr}	Hz	3067					
	s_{rr}	Hz	23.96					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	15.22					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	32.52					
Operating conditions	p_r at z_{pii}	<i>Mpa</i>	–					
	Focus	cm	9.0	4.0		4.0		4.0
	Depth	cm	9.86	13.55		13.55		13.55
	Scanning width	%	100	100		100		100
	Freq	MHz	2.5	2.5		2.5		2.5
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.6		1.7		1.7
Index component value				0.6	0.6	1.7	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.10					
	P	mW		120.15		120.15		120.15
	$P_{1 \times 1}$	mW		54.81		54.81		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.65					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	2.46	2.48		2.48		2.48
Other Information	pr	Hz	9000					
	srr	Hz	70.31					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	20.21					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	44.54					
	p_r at z_{pii}	Mpa	–					
Operating conditions	Focus	cm	4.0	3.0		3.0		3.0
	Depth	cm	9.86	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	2.5/2.5	2.5/2.5		2.5/2.5		2.5/2.5
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.8		1.6		1.6
Index component value				0.8	0.5	1.6	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.79					
	P	mW		108.49		112.41		112.41
	$P_{1 \times 1}$	mW		70.28		73.44		
	Z_s	cm			2.45			
	Z_b	cm					3.85	
	Z_{MI}	cm	5.70					
	$z_{pii,\alpha}$	cm	5.70					
	f_{awf}	MHz	2.50	2.51		2.52		2.52
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	157.80					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	145.25					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	231.53					
	p_r at z_{pii}	<i>Mpa</i>	1.63					
Operating conditions	Focus	cm	1.0	5.0		5.0		5.0
	Depth	cm	7.39	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	2.5/2.5	2.5/2.5		2.5/2.5		2.5/2.5
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.3		0.8		1.5
Index component value				0.8	1.3	0.8	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.26					
	P	mW		101.21		103.22		103.22
	$P_{1 \times 1}$	mW		67.46		68.45		
	Z_s	cm			2.90			
	Z_b	cm					2.95	
	Z_{MI}	cm	2.67					
	$Z_{pii,\alpha}$	cm	2.67					
	f_{awf}	MHz	2.48	2.49		2.50		2.50
Other Information	p_{rr}	Hz	5460					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	194.50					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	165.32					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	262.20					
	p_r at Z_{pii}	<i>Mpa</i>	1.63					
Operating conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	8.62	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	2.5	2.5		2.5		2.5
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5**Operating Mode: CW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.2	0.8		1.7		1.3
Index component value				0.6	0.8	1.3	1.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.31					
	P	mW		88.46		92.50		92.50
	$P_{1 \times 1}$	mW		53.22		55.46		
	Z_s	cm			5.52			
	Z_b	cm					6.91	
	Z_{MI}	cm	6.92					
	$Z_{pii,\alpha}$	cm	6.92					
	f_{awf}	MHz	2.44	2.45		2.47		2.47
Other Information	p_{rr}	Hz	0					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	50.00					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	105.42					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	338.43					
	p_r at Z_{pii}	<i>Mpa</i>	0.85					
Operating conditions	Focus	cm	3.0	4.0		4.0		4.0
	Depth	cm	6.16	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	2.5/2.5	2.5/2.5		2.5/2.5		2.5/2.5
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5**Operating Mode: CFM-M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.0	0.2		0.5		0.5
Index component value				0.2	0.1	0.5	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.57					
	P	mW		32.16		34.15		34.15
	$P_{1 \times 1}$	mW		20.09		22.11		
	Z_s	cm			6.20			
	Z_b	cm					6.15	
	Z_{MI}	cm	6.17					
	$Z_{pii,\alpha}$	cm	6.17					
	f_{awf}	MHz	2.45	2.46		2.47		2.47
Other Information	p_{rr}	Hz	200					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	100.05					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	105.24					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	299.00					
	p_r at Z_{pii}	Mpa	2.65					
Operating conditions	Focus	cm	5.0	4.0		6.0		6.0
	Depth	cm	6.16	8.62		8.62		8.62
	Scanning width	%	100	100		100		100
	Freq	MHz	2.5	2.5		2.5		2.5
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5

Operating Mode: TDI

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.7		1.8		1.8
Index component value				0.7	0.7	1.8	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.11					
	P	mW		125.14		125.14		125.14
	$P_{1 \times 1}$	mW		58.10		58.10		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	4.61					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	2.51	2.53		2.53		2.53
Other Information	prr	Hz	9000					
	srr	Hz	70.31					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.58					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	34.65					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	4.0	3.0		3.0		3.0
	Depth	cm	9.86	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	2.5/2.5	2.5/2.5		2.5/2.5		2.5/2.5
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S2-8

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.3		1.8		1.8
Index component value				1.3	1.3	1.8	1.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.73					
	P	mW		60.22		60.22		60.22
	$P_{1 \times 1}$	mW		58.21		58.21		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	2.25					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.67	4.69		4.69		4.69
Other Information	p_{rr}	Hz	4187					
	s_{rr}	Hz	32.71					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	16.87					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	34.86					
	p_r at Z_{pii}	Mpa	–					
Operating conditions	Focus	cm	5.0	8.0		8.0		8.0
	Depth	cm	23.32	9.86		9.86		9.86
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3	5.3		5.3		5.3
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S2-8**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.8		1.2		1.2
Index component value				0.8	0.8	1.2	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.21					
	P	mW		42.22		42.22		42.22
	$P_{1 \times 1}$	mW		41.08		41.08		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	2.28					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.08	4.09		4.09		4.09
Other Information	p_{rr}	Hz	10000					
	s_{rr}	Hz	78.13					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	25.55					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	48.58					
	p_r at Z_{pii}	Mpa	–					
Operating conditions	Focus	cm	6.0	7.0		7.0		7.0
	Depth	cm	19.71	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3/4.0	5.3/4.0		5.3/4.0		5.3/4.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S2-8**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.5		1.0		0.7
Index component value				0.4	0.5	0.7	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.23					
	P	mW		21.25		25.44		25.44
	$P_{1 \times 1}$	mW		20.22		23.39		
	Z_s	cm			1.75			
	Z_b	cm					1.88	
	Z_{MI}	cm	1.98					
	$z_{pii,\alpha}$	cm	1.98					
	f_{awf}	MHz	6.02	6.05		6.04		6.04
Other Information	pr	Hz	10000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	107.55					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	126.52					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	288.24					
	p_r at z_{pii}	Mpa	1.38					
Operating conditions	Focus	cm	7.0	5.0		5.0		5.0
	Depth	cm	7.39	11.09		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3/6.4	5.3/6.4		5.3/6.4		5.3/6.4
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S2-8

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.5		1.1		1.5
Index component value				1.0	1.5	1.0	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.76					
	P	mW		44.22		50.11		50.11
	$P_{1 \times 1}$	mW		43.12		48.89		
	Z_s	cm			1.98			
	Z_b	cm					2.05	
	Z_{MI}	cm	2.15					
	$Z_{pii,\alpha}$	cm	2.15					
	f_{awf}	MHz	4.85	4.87		4.88		4.88
Other Information	p_{rr}	Hz	6020					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	152.46					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	148.54					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	305.26					
	p_r at Z_{pii}	<i>Mpa</i>	2.35					
Operating conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	11.09	11.09		8.62		8.62
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3	5.3		5.3		5.3
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S2-8**Operating Mode: CW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.8		1.3		1.2
Index component value				0.8	0.7	1.2	1.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.85					
	P	mW		40.22		41.16		41.16
	$P_{1 \times 1}$	mW		39.45		40.65		
	Z_s	cm			3.92			
	Z_b	cm					4.10	
	Z_{MI}	cm	3.94					
	$Z_{pii,\alpha}$	cm	3.94					
	f_{awf}	MHz	4.53	4.72		4.75		4.72
Other Information	p_{rr}	Hz	0					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	28.44					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	102.28					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	350.94					
	p_r at Z_{pii}	Mpa	0.08					
Operating conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	6.16	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3/5.3	5.3/5.3		5.3/5.3		5.3/5.3
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S2-8**Operating Mode: CFM-M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.4		0.5		0.5
Index component value				0.3	0.4	0.5	0.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.28					
	P	mW		14.45		18.88		18.88
	$P_{1 \times 1}$	mW		13.76		17.59		
	Z_s	cm			2.92			
	Z_b	cm					2.99	
	Z_{MI}	cm	3.20					
	$Z_{pii,\alpha}$	cm	3.20					
	f_{awf}	MHz	4.55	4.58		4.62		4.58
Other Information	p_{rr}	Hz	198					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	53.28					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	106.25					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	290.49					
	p_r at Z_{pii}	<i>Mpa</i>	2.33					
Operating conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	6.16	8.62		8.62		8.62
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3	5.3		5.3		5.3
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S2-8**Operating Mode: TDI**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.7		1.1		1.1
Index component value				0.7	0.7	1.1	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.41					
	P	mW		36.66		36.66		36.66
	$P_{1 \times 1}$	mW		35.94		35.94		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.22					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.06	4.09		4.09		4.09
Other Information	pr	Hz	10000					
	srr	Hz	78.13					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	22.38					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	41.71					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	6.0	7.0		7.0		7.0
	Depth	cm	19.71	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3/4.0	5.3/4.0		5.3/4.0		5.3/4.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5P

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.0	1.1		1.5		1.5
Index component value				1.1	1.1	1.5	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.61					
	P	mW		122.55		122.55		122.55
	$P_{1 \times 1}$	mW		86.84		86.84		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.48					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	2.58	2.66		2.66		2.66
Other Information	p_{rr}	Hz	3325					
	s_{rr}	Hz	25.98					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.88					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	35.29					
	p_r at Z_{pii}	Mpa	–					
Operating conditions	Focus	cm	4.0	1.0		1.0		1.0
	Depth	cm	25.87	29.57		29.57		29.57
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0	3.0		3.0		3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5P

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.6		0.9		0.9
Index component value				0.6	0.6	0.9	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.28					
	P	mW		71.22		71.22		71.22
	$P_{1 \times 1}$	mW		47.91		47.91		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	3.42					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	2.55	2.63		2.63		2.63
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	70.31					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	20.25					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	36.99					
	p_r at z_{pii}	Mpa	–					
Operating conditions	Focus	cm	4.0	1.0		1.0		1.0
	Depth	cm	20.94	7.39		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		3.0/3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5P

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.8		1.0		1.0
Index component value				0.7	0.8	1.0	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.28					
	P	mW		84.55		84.55		84.55
	$P_{1 \times 1}$	mW		56.98		56.98		
	Z_s	cm			2.98			
	Z_b	cm					3.16	
	Z_{MI}	cm	5.88					
	$Z_{pii,\alpha}$	cm	5.88					
	f_{awf}	MHz	2.58	2.58		2.58		2.58
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	148.22					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	153.62					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	438.09					
	p_r at Z_{pii}	<i>Mpa</i>	1.16					
Operating conditions	Focus	cm	1.0	1.0		1.0		1.0
	Depth	cm	7.39	7.39		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		3.0/3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5P**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.3	1.2		1.1		1.8
Index component value				1.0	1.2	1.0	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.13					
	P	mW		122.24		123.41		123.41
	$P_{1 \times 1}$	mW		77.78		78.24		
	Z_s	cm			2.90			
	Z_b	cm					2.95	
	Z_{MI}	cm	2.60					
	$Z_{pii,\alpha}$	cm	2.60					
	f_{awf}	MHz	2.68	2.70		2.72		2.72
Other Information	p_{rr}	Hz	5690					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	135.60					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	150.22					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	243.09					
	p_r at Z_{pii}	Mpa	1.62					
Operating conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	8.62	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0	3.0		3.0		3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5P**Operating Mode: CW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.2	0.7		1.6		0.8
Index component value				0.6	0.7	0.8	1.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.34					
	P	mW		65.25		68.44		68.44
	$P_{1 \times 1}$	mW		45.32		47.26		
	Z_s	cm			5.33			
	Z_b	cm					5.74	
	Z_{MI}	cm	6.68					
	$Z_{pii,\alpha}$	cm	6.68					
	f_{awf}	MHz	2.82	2.78		2.82		2.82
Other Information	p_{rr}	Hz	0					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	45.22					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	125.58					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	461.37					
	p_r at Z_{pii}	<i>Mpa</i>	0.15					
Operating conditions	Focus	cm	5.0	4.0		6.0		6.0
	Depth	cm	6.16	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		3.0/3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5P**Operating Mode: CFM-M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.4		0.6		0.6
Index component value				0.4	0.2	0.6	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.46					
	P	mW		48.13		50.22		50.22
	$P_{1 \times 1}$	mW		31.58		33.24		
	Z_s	cm			5.58			
	Z_b	cm					5.62	
	Z_{MI}	cm	6.08					
	$z_{pii,\alpha}$	cm	6.08					
	f_{awf}	MHz	2.64	2.66		2.64		2.64
Other Information	p_{rr}	Hz	249					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	81.22					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	115.18					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	349.07					
	p_r at z_{pii}	<i>Mpa</i>	2.59					
Operating conditions	Focus	cm	5.0	3.0		6.0		6.0
	Depth	cm	6.16	8.62		8.62		8.62
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0	3.0		3.0		3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S1-5P

Operating Mode: TDI

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.5		0.7		0.7
Index component value				0.5	0.5	0.7	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.12					
	P	mW		60.21		60.21		60.21
	$P_{1 \times 1}$	mW		39.47		39.47		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	3.38					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	2.56	2.66		2.66		2.66
Other Information	pr	Hz	9000					
	srr	Hz	70.31					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	25.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	45.83					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	4.0	1.0		1.0		1.0
	Depth	cm	20.94	7.39		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		3.0/3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-10

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.5		0.8		–
Index component value				0.5	0.5	0.8	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.38					
	P	mW		32.15		32.15		–
	$P_{1 \times 1}$	mW		19.72		19.72		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	2.48					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	5.28	5.33		5.33		–
Other Information	pr	Hz	7268					
	srr	Hz	56.78					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	18.84					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	46.55					
	p_r at Z_{pii}	Mpa	–					
Operating conditions	Focus	cm	3.0	2.0		2.0		–
	Depth	cm	4.29	3.06		3.06		–
	Scanning width	%	100	100		100		–
	Freq	MHz	6.0	6.0		6.0		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-10

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.9		1.6		—
Index component value				0.9	0.9	1.6	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.21					
	P	mW		68.44		68.44		—
	$P_{1 \times 1}$	mW		41.45		41.45		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	3.38					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.08	4.56		4.56		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	46.88					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	20.16					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	52.26					
	p_r at z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	4.0	2.5		2.5		—
	Depth	cm	6.75	3.06		3.06		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0/4.0	6.0/4.0		6.0/4.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-10**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.8		1.1		—
Index component value				0.6	0.8	1.1	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.27					
	P	mW		44.21		46.69		—
	$P_{1 \times 1}$	mW		27.94		28.55		
	Z_s	cm			1.82			
	Z_b	cm					1.95	
	Z_{MI}	cm	2.44					
	$Z_{pii,\alpha}$	cm	2.44					
	f_{awf}	MHz	4.46	4.51		4.48		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	138.55					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	115.72					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	245.40					
	p_r at Z_{pii}	Mpa	1.52					
Operating conditions	Focus	cm	3.0	3.5		3.5		—
	Depth	cm	3.06	9.22		5.52		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0/4.0	6.0/4.0		6.0/4.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-10**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.5		1.2		—
Index component value				0.4	0.5	0.4	1.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.54					
	P	mW		20.14		21.44		—
	$P_{1 \times 1}$	mW		14.33		15.24		
	Z_s	cm			1.82			
	Z_b	cm					2.15	
	Z_{MI}	cm	2.36					
	$z_{pii,\alpha}$	cm	2.36					
	f_{awf}	MHz	5.83	5.86		5.83		—
Other Information	p_{rr}	Hz	8120					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	125.66					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	122.14					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	315.95					
	p_r at z_{pii}	Mpa	2.30					
Operating conditions	Focus	cm	3.0	3.0		3.0		—
	Depth	cm	4.29	10.45		5.22		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0	6.0		6.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-10**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.4		0.7		—
Index component value				0.4	0.4	0.7	0.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.61					
	P	mW		30.22		30.22		—
	$P_{1 \times 1}$	mW		15.76		15.76		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.48					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	5.28	5.33		5.33		—
Other Information	p_{rr}	Hz	7268					
	s_{rr}	Hz	56.78					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	18.84					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	46.55					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	3.0	2.0		2.0		—
	Depth	cm	4.29	3.06		3.06		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0	6.0		6.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: VE4-10

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.5		0.6		—
Index component value				0.5	0.5	0.6	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.60					
	P	mW		35.22		35.22		—
	$P_{1 \times 1}$	mW		19.27		19.27		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.58					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	5.25	5.45		5.45		—
Other Information	p_{rr}	Hz	7143					
	s_{rr}	Hz	55.80					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	15.41					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	39.28					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	3.0	2.0		2.0		—
	Depth	cm	3.06	3.06		3.06		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0	6.0		6.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: VE4-10**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.7		1.1		—
Index component value				0.7	0.7	1.1	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.00					
	P	mW		60.21		60.21		—
	$P_{1 \times 1}$	mW		32.31		32.31		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	3.08					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.02	4.55		4.55		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	46.88					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	16.47					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	38.74					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	4.0	3.0		3.0		—
	Depth	cm	10.45	4.29		4.29		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0/4.0	6.0/4.0		6.0/4.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: VE4-10**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.9		1.2		—
Index component value				0.8	0.9	1.2	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.26					
	P	mW		64.41		67.24		—
	$P_{1 \times 1}$	mW		37.75		40.01		
	Z_s	cm			1.74			
	Z_b	cm					1.88	
	Z_{MI}	cm	4.48					
	$z_{pii,\alpha}$	cm	4.48					
	f_{awf}	MHz	4.44	4.45		4.46		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	138.44					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	133.24					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	526.47					
	p_r at z_{pii}	<i>Mpa</i>	1.57					
Operating conditions	Focus	cm	4.0	4.0		4.0		—
	Depth	cm	4.29	4.29		5.22		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0/4.0	6.0/4.0		6.0/4.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: VE4-10**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.3		0.9		—
Index component value				0.2	0.3	0.2	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.69					
	P	mW		12.58		15.54		—
	$P_{1 \times 1}$	mW		7.16		9.87		
	Z_s	cm			1.89			
	Z_b	cm					2.06	
	Z_{MI}	cm	2.85					
	$z_{pii,\alpha}$	cm	2.85					
	f_{awf}	MHz	5.85	5.87		5.87		—
Other Information	p_{rr}	Hz	7000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	123.50					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	118.54					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	375.00					
	p_r at z_{pii}	<i>Mpa</i>	2.38					
Operating conditions	Focus	cm	3.0	3.0		3.0		—
	Depth	cm	7.98	5.22		5.22		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0	6.0		6.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: VE4-10**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.6		0.8		—
Index component value				0.6	0.6	0.8	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.60					
	P	mW		48.24		48.24		—
	$P_{1 \times 1}$	mW		23.12		23.12		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.58					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	5.25	5.45		5.45		—
Other Information	p_{rr}	Hz	7143					
	s_{rr}	Hz	55.80					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	15.41					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	39.28					
	p_r at z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	3.0	2.0		2.0		—
	Depth	cm	3.06	3.06		3.06		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0	6.0		6.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: BL3-12**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.3		0.4		—
Index component value				0.3	0.3	0.4	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.34					
	P	mW		14.34		14.34		—
	$P_{1 \times 1}$	mW		8.76		8.76		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.25					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.19	7.19		7.19		—
Other Information	pr	Hz	4344					
	srr	Hz	33.94					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.47					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	47.30					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	5.16	5.16		5.16		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: BL3-12**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.7		1.3		—
Index component value				0.7	0.7	1.3	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.38					
	P	mW		50.18		50.18		—
	$P_{1 \times 1}$	mW		27.79		27.79		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.48					
	$Z_{pii,\alpha}$	Cm	—					
	f_{awf}	MHz	5.29	5.29		5.29		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	46.88					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	18.84					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	46.63					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	6.39	6.39		6.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/5.3	7.5/5.3		7.5/5.3		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: BL3-12

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.3		1.8		–
Index component value				1.3	1.0	1.8	1.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.39					
	P	mW		72.22		72.22		–
	$P_{1 \times 1}$	mW		50.84		50.84		
	Z_s	cm			2.31			
	Z_b	Cm					2.35	
	Z_{MI}	cm	2.24					
	$z_{pii,\alpha}$	cm	2.24					
	f_{awf}	MHz	5.37	5.37		5.37		–
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	–					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	144.55					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	114.28					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	262.32					
	p_r at z_{pii}	<i>Mpa</i>	1.49					
Operating conditions	Focus	cm	2.5	2.5		2.5		–
	Depth	cm	8.86	8.86		8.86		–
	Scanning width	%	100	100		100		–
	Freq	MHz	7.5/5.3	7.5/5.3		7.5/5.3		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: BL3-12

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.5		1.8		—
Index component value				1.4	1.5	1.4	1.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.61					
	P	mW		60.22		62.24		—
	$P_{1 \times 1}$	mW		40.72		41.12		
	Z_s	cm			2.38			
	Z_b	cm					2.44	
	Z_{MI}	cm	2.50					
	$Z_{pii,\alpha}$	cm	2.50					
	f_{awf}	MHz	7.22	7.22		7.23		—
Other Information	p_{rr}	Hz	5420					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	142.30					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	123.22					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	428.73					
	p_r at Z_{pii}	Mpa	1.30					
Operating conditions	Focus	cm	3.0	3.0		3.0		—
	Depth	cm	5.16	5.16		6.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: BL3-12**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.2		0.3		—
Index component value				0.2	0.2	0.3	0.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.61					
	P	mW		10.04		10.04		—
	$P_{1 \times 1}$	mW		5.84		5.84		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.25					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.19	7.19		7.19		—
Other Information	p_{rr}	Hz	4344					
	s_{rr}	Hz	33.94					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.47					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	47.30					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	5.16	5.16		5.16		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: V2-6**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.2		1.5		—
Index component value				1.2	1.2	1.5	1.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.24					
	P	mW		115.22		115.22		—
	$P_{1 \times 1}$	mW		57.93		57.93		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	3.22					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.26	4.35		4.35		—
Other Information	p_{rr}	Hz	9254					
	s_{rr}	Hz	72.30					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	18.87					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	48.67					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	4.0	2.0		2.0		—
	Depth	cm	6.61	5.40		5.40		—
	Scanning width	%	100	100		100		—
	Freq	MHz	4.5	4.5		4.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: V2-6**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.6		0.8		—
Index component value				0.6	0.6	0.8	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.00					
	P	mW		65.44		65.44		—
	$P_{1 \times 1}$	mW		32.47		32.47		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	5.25					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.97	3.88		3.88		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	46.88					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.48					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	65.32					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	6.0	6.0		6.0		—
	Depth	cm	10.59	11.82		11.82		—
	Scanning width	%	100	100		100		—
	Freq	MHz	4.5/4.5	4.5/4.5		4.5/4.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: V2-6**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	1.1		1.2		—
Index component value				1.0	1.1	1.2	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.98					
	P	mW		94.78		95.55		—
	$P_{1 \times 1}$	mW		51.98		52.24		
	Z_s	cm			4.11			
	Z_b	cm					4.08	
	Z_{MI}	cm	3.24					
	$z_{pii,\alpha}$	cm	3.24					
	f_{awf}	MHz	3.88	4.04		4.08		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	125.14					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	125.49					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	299.05					
	p_r at z_{pii}	<i>Mpa</i>	1.18					
Operating conditions	Focus	cm	6.0	7.0		6.0		—
	Depth	cm	8.13	8.13		8.13		—
	Scanning width	%	100	100		100		—
	Freq	MHz	4.5/4.5	4.5/4.5		4.5/4.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: V2-6

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.5		1.7		—
Index component value				1.4	1.5	1.4	1.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.24					
	P	mW		134.45		138.52		—
	$P_{1 \times 1}$	mW		68.85		70.25		
	Z_s	cm			3.88			
	Z_b	cm					5.08	
	Z_{MI}	cm	4.15					
	$z_{pii,\alpha}$	cm	4.15					
	f_{awf}	MHz	4.26	4.27		4.29		—
Other Information	pr	Hz	4825					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	202.50					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	130.24					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	441.69					
	p_r at z_{pii}	Mpa	2.41					
Operating conditions	Focus	cm	7.0	4.0		5.0		—
	Depth	cm	8.13	5.40		6.61		—
	Scanning width	%	100	100		100		—
	Freq	MHz	4.5	4.5		4.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: V2-6

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	1.3		1.3		—
Index component value				1.3	1.3	1.2	1.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.44					
	P	mW		92.55		92.55		—
	$P_{1 \times 1}$	mW		62.76		62.76		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	3.22					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.26	4.35		4.35		—
Other Information	p_{rr}	Hz	9254					
	s_{rr}	Hz	72.30					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	18.87					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	48.67					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	4.0	2.0		2.0		—
	Depth	cm	6.61	5.40		5.40		—
	Scanning width	%	100	100		100		—
	Freq	MHz	4.5	4.5		4.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C2-6**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.1		1.5		—
Index component value				1.1	1.1	1.5	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.09					
	P	mW		102.28		102.28		—
	$P_{1 \times 1}$	mW		67.15		67.15		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	4.98					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.28	3.44		3.44		—
Other Information	p_{rr}	Hz	3837					
	s_{rr}	Hz	29.98					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	14.59					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	45.09					
	p_r at z_{pii}	<i>Mpa</i>	—					
Operating conditions	Focus	cm	5.0	7.0		7.0		—
	Depth	cm	8.84	10.70		10.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.0	3.0		3.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C2-6**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.6		1.1		—
Index component value				0.6	0.6	1.1	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.67					
	P	mW		75.43		75.43		—
	$P_{1 \times 1}$	mW		44.06		44.06		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	3.87					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	2.84	2.86		2.86		—
Other Information	pr	Hz	6000					
	srr	Hz	46.88					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.45					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	33.01					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	5.0	5.0		5.0		—
	Depth	cm	7.61	7.61		7.61		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C2-6

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.7		1.4		—
Index component value				0.7	0.6	1.4	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.84					
	P	mW		93.28		94.11		—
	$P_{1 \times 1}$	mW		51.04		52.13		
	Z_s	cm			3.92			
	Z_b	cm					3.58	
	Z_{MI}	cm	5.58					
	$z_{pii,\alpha}$	cm	5.58					
	f_{awf}	MHz	2.85	2.88		2.87		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	104.58					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	155.44					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	466.29					
	p_r at z_{pii}	Mpa	0.74					
Operating conditions	Focus	cm	5.0	5.0		5.0		—
	Depth	cm	8.84	8.84		17.46		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C2-6**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.3		0.5		—
Index component value				0.5	1.3	0.5	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.96					
	P	mW		100.22		103.45		—
	$P_{1 \times 1}$	mW		40.71		42.21		
	Z_s	cm			3.51			
	Z_b	cm					4.33	
	Z_{MI}	cm	4.68					
	$Z_{pii,\alpha}$	cm	4.68					
	f_{awf}	MHz	2.56	2.58		2.58		—
Other Information	pr	Hz	7620					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	110.22					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	158.54					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	362.71					
	p_r at Z_{pii}	Mpa	1.95					
Operating conditions	Focus	cm	5.0	5.0		5.0		—
	Depth	cm	6.37	6.37		6.37		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.0	3.0		3.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C2-6**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	1.2		1.6		–
Index component value				1.2	1.2	1.6	1.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.27					
	P	mW		110.28		110.28		–
	$P_{1 \times 1}$	mW		73.26		73.26		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.98					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	3.28	3.44		3.44		–
Other Information	p_{rr}	Hz	3837					
	s_{rr}	Hz	29.98					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	14.59					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	45.09					
	p_r at z_{pii}	Mpa	–					
Operating conditions	Focus	cm	5.0	7.0		7.0		–
	Depth	cm	8.84	10.70		10.70		–
	Scanning width	%	100	100		100		–
	Freq	MHz	3.0	3.0		3.0		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C4-11**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	1.3		1.4		—
Index component value				1.3	1.3	1.4	1.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.67					
	P	mW		60.28		60.28		—
	$P_{1 \times 1}$	mW		46.27		46.27		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.68					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	5.68	5.90		5.90		—
Other Information	p_{rr}	Hz	5918					
	s_{rr}	Hz	46.23					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.89					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	45.31					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.5	2.0		2.0		—
	Depth	cm	5.69	6.92		6.92		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0	6.0		6.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C4-11**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	1.0		1.2		–
Index component value				1.0	1.0	1.2	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.61					
	P	mW		48.29		48.29		–
	$P_{1 \times 1}$	mW		39.03		39.03		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	1.78					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	5.28	5.38		5.38		–
Other Information	pr	Hz	10000					
	srr	Hz	78.13					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	15.55					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	29.76					
	p_r at z_{pii}	Mpa	–					
Operating conditions	Focus	cm	2.5	1.0		1.0		–
	Depth	cm	11.85	3.22		3.22		–
	Scanning width	%	100	100		100		–
	Freq	MHz	6.0/5.3	6.0/5.3		6.0/5.3		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C4-11

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.3		1.4		—
Index component value				1.3	1.0	1.4	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.37					
	P	mW		60.28		60.28		—
	$P_{1 \times 1}$	mW		50.56		50.56		
	Z_s	cm			1.82			
	Z_b	cm					1.83	
	Z_{MI}	cm	2.88					
	$Z_{pii,\alpha}$	cm	2.88					
	f_{awf}	MHz	5.22	5.40		5.40		—
Other Information	pr	Hz	10000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	135.55					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	145.28					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	410.16					
	p_r at Z_{pii}	Mpa	2.42					
Operating conditions	Focus	cm	2.5	3.5		3.5		—
	Depth	cm	3.22	4.46		4.46		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0/5.3	6.0/5.3		6.0/5.3		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C4-11

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	2.0		1.4		—
Index component value				1.2	2.0	1.2	1.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.66					
	P	mW		89.54		92.55		—
	$P_{1 \times 1}$	mW		44.84		45.65		
	Z_s	cm			1.98			
	Z_b	cm					2.06	
	Z_{MI}	cm	3.56					
	$z_{pii,\alpha}$	cm	3.56					
	f_{awf}	MHz	5.60	5.62		5.64		—
Other Information	p_{rr}	Hz	12300					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	183.52					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	145.66					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	577.32					
	p_r at z_{pii}	Mpa	2.49					
Operating conditions	Focus	cm	4.0	2.0		3.5		—
	Depth	cm	10.62	4.46		8.15		—
	Scanning width	%	100	100		100		—
	Freq	MHz	6.0	6.0		6.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: C4-11

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.1		1.2		–
Index component value				1.1	1.1	1.2	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.91					
	P	mW		48.59		48.59		–
	$P_{1 \times 1}$	mW		39.15		39.15		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	2.68					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	5.68	5.90		5.90		–
Other Information	p_{rr}	Hz	5918					
	s_{rr}	Hz	46.23					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	15.89					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	45.31					
	p_r at z_{pii}	Mpa	–					
Operating conditions	Focus	cm	2.5	2.0		2.0		–
	Depth	cm	5.69	6.92		6.92		–
	Scanning width	%	100	100		100		–
	Freq	MHz	6.0	6.0		6.0		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: CW2

Operating Mode: CW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.6		0.9		0.8
Index component value				0.5	0.6	0.8	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.82					
	P	mW		98.84		98.84		98.84
	$P_{1 \times 1}$	mW		55.56		55.56		
	Z_s	cm			3.88			
	Z_b	cm					3.52	
	Z_{MI}	cm	4.23					
	$Z_{pii,\alpha}$	cm	4.23					
	f_{awf}	MHz	1.85	1.89		1.88		1.90
Other Information	pr	Hz	0					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	35.79					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	112.28					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	192.78					
	p_r at Z_{pii}	Mpa	0.68					
Operating conditions	Focus	cm	3.0	3.0		3.0		3.0
	Depth	cm	3.12	3.12		3.12		3.12
	Scanning width	%	100	100		100		100
	Freq	MHz	2.0	2.0		2.0		2.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: T4-6**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.9		1.1		1.1
Index component value				0.9	0.9	1.1	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.94					
	P	mW		80.29		80.29		80.29
	$P_{1 \times 1}$	mW		39.87		39.87		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.56					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.66	4.74		4.74		4.74
Other Information	pr	Hz	3114					
	srr	Hz	24.33					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	13.84					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	60.07					
	p_r at Z_{pii}	<i>Mpa</i>	–					
Operating conditions	Focus	cm	9.0	4.0		4.0		4.0
	Depth	cm	9.86	13.55		13.55		13.55
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0	5.0		5.0		5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: T4-6
Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.7		1.0		1.0
Index component value				0.7	0.7	1.0	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.52					
	P	mW		78.19		78.19		78.19
	$P_{1 \times 1}$	mW		32.45		32.45		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	4.18					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.72	4.53		4.53		4.53
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	70.31					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	12.29					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	48.02					
	p_r at z_{pii}	Mpa	—					
Operating conditions	Focus	cm	4.0	3.0		3.0		3.0
	Depth	cm	9.86	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0/5.0	5.0/5.0		5.0/5.0		5.0/5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: T4-6**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.8		1.1		1.1
Index component value				0.8	0.6	1.1	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.65					
	P	mW		78.88		82.19		82.19
	$P_{1 \times 1}$	mW		42.28		43.87		
	Z_s	cm			3.84			
	Z_b	cm					3.98	
	Z_{MI}	cm	4.66					
	$Z_{pii,\alpha}$	cm	4.66					
	f_{awf}	MHz	4.25	4.28		4.28		4.28
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	108.84					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	119.55					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	469.57					
	p_r at Z_{pii}	Mpa	1.44					
Operating conditions	Focus	cm	1.0	5.0		5.0		5.0
	Depth	cm	7.39	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0/5.0	5.0/5.0		5.0/5.0		5.0/5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: T4-6

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.7		1.4		1.0
Index component value				0.6	0.7	0.6	1.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.72					
	P	mW		64.45		68.87		68.87
	$P_{1 \times 1}$	mW		29.12		32.12		
	Z_s	cm			2.81			
	Z_b	cm					2.93	
	Z_{MI}	cm	2.59					
	$z_{pii,\alpha}$	cm	2.59					
	f_{awf}	MHz	4.62	4.62		4.62		4.62
Other Information	pr	Hz	5362					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	125.10					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	118.42					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	270.64					
Operating conditions	p_r at z_{pii}	Mpa	1.66					
	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	8.62	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0	5.0		5.0		5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: T4-6
Operating Mode: CW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.6		0.9		0.9
Index component value				0.5	0.6	0.9	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.24					
	P	mW		60.67		66.58		66.58
	$P_{1 \times 1}$	mW		28.49		30.22		
	Z_s	cm			3.78			
	Z_b	cm					3.92	
	Z_{MI}	cm	4.68					
	$z_{pii,\alpha}$	cm	4.68					
	f_{awf}	MHz	4.26	4.27		4.27		4.27
Other Information	p_{rr}	Hz	0					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	52.28					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	95.46					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	378.38					
	p_r at z_{pii}	Mpa	1.49					
Operating conditions	Focus	cm	3.0	5.0		4.0		4.0
	Depth	cm	7.39	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0/5.0	5.0/5.0		5.0/5.0		5.0/5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: T4-6**Operating Mode: CFM-M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.5		0.8		0.8
Index component value				0.5	0.4	0.8	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.65					
	P	mW		55.49		58.75		58.75
	$P_{1 \times 1}$	mW		27.78		28.84		
	Z_s	cm			2.59			
	Z_b	cm					2.77	
	Z_{MI}	cm	2.62					
	$Z_{pii,\alpha}$	cm	2.62					
	f_{awf}	MHz	4.25	4.25		4.27		4.27
Other Information	p_{rr}	Hz	5212					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	81.24					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	101.28					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	218.56					
	p_r at Z_{pii}	Mpa	1.58					
Operating conditions	Focus	cm	3.0	5.0		6.0		6.0
	Depth	cm	8.62	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0	5.0		5.0		5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: T4-6

Operating Mode: TDI

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.6		0.9		0.9
Index component value				0.6	0.6	0.9	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.31					
	P	mW		68.44		68.44		68.44
	$P_{1 \times 1}$	mW		30.25		30.25		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.18					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.74	4.55		4.55		4.55
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	70.31					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	12.58					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	49.44					
	p_r at z_{pii}	Mpa	–					
Operating conditions	Focus	cm	4.0	3.0		3.0		3.0
	Depth	cm	9.86	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0/5.0	5.0/5.0		5.0/5.0		5.0/5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-8

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.53	0.18		0.18		—
Index component value				0.18	0.18	0.18	0.18	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.10					
	P	mW		8.64		8.64		—
	$P_{1 \times 1}$	mW		59.32		59.32		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.35					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.26	4.29		4.29		—
Other Information	p_{rr}	Hz	6438.00					
	s_{rr}	Hz	75.66					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	43.09					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	65.86					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	8.0	8.0		8.0		—
	Depth	cm	8.59	8.59		8.59		—
	Scanning width	%	100	100		100		—
	Freq	MHz	5.5	5.5		5.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-8

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.72	0.30		0.30		—
Index component value				0.30	0.24	0.30	0.24	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.49					
	P	mW		16.38		16.38		—
	$P_{1 \times 1}$	mW		35.69		35.69		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.45					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.26	4.10		4.10		—
Other Information	pr	Hz	1165.00					
	sr	Hz	93.75					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	2.12					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	108.48					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	8.59	8.59		8.59		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.0	7.0		7.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-8

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.67	0.28		0.77		—
Index component value				0.28	0.21	0.77	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.75					
	P	mW		15.69		14.18		—
	$P_{1 \times 1}$	mW		15.69		14.18		
	Z_s	cm			-			
	Z_b	cm					-	
	Z_{MI}	cm	1.0					
	$Z_{pii,\alpha}$	cm	1.0					
	f_{awf}	MHz	5.04	4.99		4.97		—
Other Information	p_{rr}	Hz	4000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	77.59					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	78.56					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	79.39					
	p_r at Z_{pii}	Mpa	1.64					
Operating conditions	Focus	cm	8.0	8.0		8.0		—
	Depth	cm	8.59	8.59		8.59		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.0	7.0		7.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L3-8

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.58	0.17		0.17		—
Index component value				0.17	0.01	0.17	0.01	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.20					
	P	mW		8.31		8.31		—
	$P_{1 \times 1}$	mW		8.31		8.31		
	Z_s	cm			4.23			
	Z_b	cm					2.55	
	Z_{MI}	cm	2.10					
	$Z_{pii,\alpha}$	cm	2.10					
	f_{awf}	MHz	4.19	4.32		4.32		—
Other Information	pr	Hz	5255.00					
	sr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	47.52					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	47.50					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	48.30					
	p_r at Z_{pii}	Mpa	2.64					
Operating conditions	Focus	cm	8.0	8.0		8.0		—
	Depth	cm	8.59	8.59		8.59		—
	Scanning width	%	100	100		100		—
	Freq	MHz	5.5	5.5		5.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-10W**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.22	0.39		0.39		—
Index component value				0.39	0.32	0.39	0.32	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.10					
	P	mW		9.83		9.83		-
	$P_{1 \times 1}$	mW		9.83		9.83		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	0.95					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	8.06	8.42		8.42		—
Other Information	p_{rr}	Hz	5132.00					
	s_{rr}	Hz	75.66					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	43.09					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	299.39					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	22	22		22		—
	Scanning width	%	100	100		100		—
	Freq	MHz	-	-		-		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-10W**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.16	0.34		0.34		—
Index component value				0.34	0.24	0.34	0.24	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.65					
	P	mW		11.52		11.52		—
	$P_{1 \times 1}$	mW		11.52		11.52		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	0.90					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	5.28	5.21		5.21		—
Other Information	pr	Hz	2368.00					
	sr	Hz	93.75					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	3.08					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	312.57					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	5.0	5.0		5.0		—
	Depth	cm	28	28		28		—
	Scanning width	%	100	100		100		—
	Freq	MHz	-	-		-		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-10W

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.25	0.24		0.98		—
Index component value				0.24	0.21	0.98	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.75					
	P	mW		13.12		6.86		—
	$P_{1 \times 1}$	mW		15.69		14.18		
	Z_s	cm			-			
	Z_b	cm					-	
	Z_{MI}	cm	0.85					
	$Z_{pii,\alpha}$	cm	1.0					
	f_{awf}	MHz	5.11	5.06		5.12		—
Other Information	p_{rr}	Hz	1500					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	408.00					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	408.00					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	413.52					
Operating conditions	p_r at Z_{pii}	Mpa	3.19					
	Focus	cm	5.0	5.0		5.0		—
	Depth	cm	28	28		28		—
	Scanning width	%	100	100		100		—
	Freq	MHz	-	-		-		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-10W

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.09	0.15		0.15		—
Index component value				0.15	0.01	0.15	0.01	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	3.11					
	P	mW		3.79		0.23		—
	$P_{1 \times 1}$	mW		3.79		0.23		
	Z_s	cm			-			
	Z_b	cm					2.55	
	Z_{MI}	cm	0.95					
	$Z_{pii,\alpha}$	cm	2.10					
	f_{awf}	MHz	8.14	8.14		8.14		—
Other Information	pr	Hz	4254.00					
	sr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	47.52					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	256.00					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	275.16					
	p_r at Z_{pii}	Mpa	4.28					
Operating conditions	Focus	cm	5.0	5.0		5.0		—
	Depth	cm	28	28		28		—
	Scanning width	%	100	100		100		—
	Freq	MHz	-	-		-		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10R**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.64	0.12		0.12		—
Index component value				0.12	0.10	0.12	0.10	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.59					
	P	mW		3.99		3.99		-
	$P_{1 \times 1}$	mW		3.99		3.99		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.75					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.14	6.06		6.06		—
Other Information	p_{rr}	Hz	2754.00					
	s_{rr}	Hz	75.66					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	94.72					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	99.39					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	8.0	8.0		8.0		—
	Depth	cm	8.87	8.87		8.87		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10R

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.59	0.20		02.0		–
Index component value				0.20	0.14	0.20	0.14	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.45					
	P	mW		7.66		7.66		–
	$P_{1 \times 1}$	mW		7.66		7.66		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	2.85					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	6.17	4.97		4.97		–
Other Information	pr	Hz	1922.00					
	srr	Hz	93.75					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	3.08					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	82.03					
	p_r at Z_{pii}	Mpa	–					
Operating conditions	Focus	cm	1.5	1.5		1.5		–
	Depth	cm	8.87	8.87		8.87		–
	Scanning width	%	100	100		100		–
	Freq	MHz	4.0	4.0		4.0		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10R

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.17	0.13		0.34		—
Index component value				0.13	0.10	0.34	0.30	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.38					
	P	mW		7.33		7.33		—
	$P_{1 \times 1}$	mW		7.33		7.33		
	Z_s	cm			-			
	Z_b	cm					-	
	Z_{MI}	cm	1.30					
	$Z_{pii,\alpha}$	cm	1.0					
	f_{awf}	MHz	4.95	4.98		4.98		—
Other Information	pr	Hz	4000.00					
	sr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	408.00					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	408.00					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	413.52					
	p_r at Z_{pii}	Mpa	4.48					
Operating conditions	Focus	cm	8.0	8.0		8.0		—
	Depth	cm	8.87	8.87		8.87		—
	Scanning width	%	100	100		100		—
	Freq	MHz	4.0	4.0		4.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10R

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.72	0.12		0.12		—
Index component value				0.12	0.01	0.12	0.01	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.75					
	P	mW		4.24		4.24		—
	$P_{1 \times 1}$	mW		4.24		4.24		
	Z_s	cm			-			
	Z_b	cm					2.50	
	Z_{MI}	cm	2.50					
	$Z_{pii,\alpha}$	cm	2.10					
	f_{awf}	MHz	5.94	6.04		6.04		—
Other Information	pr	Hz	354.10					
	sr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	47.52					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	354.10					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	275.16					
	p_r at Z_{pii}	Mpa	2.89					
Operating conditions	Focus	cm	8.0	8.0		8.0		—
	Depth	cm	8.87	8.87		8.87		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-15i**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.51	0.13		0.13		—
Index component value				0.13	0.10	0.13	0.10	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.07					
	P	mW		6.34		6.34		-
	$P_{1 \times 1}$	mW		6.34		6.34		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.15					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.31	4.33		4.33		—
Other Information	p_{rr}	Hz	6438.00					
	s_{rr}	Hz	75.66					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	36.70					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	39.39					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	5.27	5.27		5.27		—
	Scanning width	%	100	100		100		—
	Freq	MHz	9.0	9.0		9.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-15i

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.76	0.24		0.24		—
Index component value				0.24	0.14	0.24	0.14	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.45					
	P	mW		13.33		13.33		—
	$P_{1 \times 1}$	mW		13.33		13.33		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.05					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.31	4.16		4.16		—
Other Information	pr	Hz	1165.00					
	sr	Hz	93.75					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	3.08					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	82.03					
	p_r at Z_{pii}	Mpa	—					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	5.27	5.27		5.27		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-15i

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.62	0.21		0.72		—
Index component value				0.21	0.10	0.72	0.30	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.40					
	P	mW		11.83		11.83		—
	$P_{1 \times 1}$	mW		11.83		11.83		
	Z_s	cm			-			
	Z_b	cm					-	
	Z_{MI}	cm	1.95					
	$Z_{pii,\alpha}$	cm	1.0					
	f_{awf}	MHz	5.01	5.02		5.03		—
Other Information	pr	Hz	3500.00					
	sr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	85.76					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	85.76					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	85.76					
	p_r at Z_{pii}	Mpa	4.48					
Operating conditions	Focus	cm	8.0	8.0		8.0		—
	Depth	cm	5.27	5.27		5.27		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L6-15i

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.60	0.14		0.14		—
Index component value				0.14	0.01	0.14	0.01	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.24					
	P	mW		6.76		6.76		—
	$P_{1 \times 1}$	mW		6.76		6.76		
	Z_s	cm			-			
	Z_b	cm					230	
	Z_{MI}	cm	2.25					
	$Z_{pii,\alpha}$	cm	2.10					
	f_{awf}	MHz	4.36	4.37		4.37		—
Other Information	pr	Hz	5255.00					
	sr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	47.52					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	354.10					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	275.16					
	p_r at Z_{pii}	Mpa	1.70					
Operating conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	5.27	5.27		5.27		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.6		0.8		—
Index component value				0.6	0.6	0.8	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.39					
	P	mW		32.56		32.56		—
	$P_{1 \times 1}$	mW		17.71		17.71		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.12					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.05	7.12		7.12		—
Other Information	p_{rr}	Hz	3015					
	s_{rr}	Hz	23.55					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	42.43					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.9		1.1		—
Index component value				0.9	0.9	1.1	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.75					
	P	mW		48.52		48.52		—
	$P_{1 \times 1}$	mW		28.14		28.14		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.68					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.22	6.25		6.25		—
Other Information	pr	Hz	20000					
	srr	Hz	156.25					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	20.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	41.60					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.5	1.0		1.0		—
	Depth	cm	4.93	3.70		3.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/6.5	7.5/6.5		7.5/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.6		1.1		—
Index component value				0.6	0.5	1.1	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.25					
	P	mW		36.91		48.12		—
	$P_{1 \times 1}$	mW		20.24		26.44		
	Z_s	cm			1.74			
	Z_b	cm					1.85	
	Z_{MI}	cm	1.73					
	$Z_{pii,\alpha}$	cm	1.73					
	f_{awf}	MHz	6.25	6.26		6.28		—
Other Information	pr	Hz	20000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	182.65					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	198.84					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	419.64					
	p_r at Z_{pii}	<i>Mpa</i>	2.20					
Operating control conditions	Focus	cm	4.0	2.5		2.5		—
	Depth	cm	7.39	3.70		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/6.5	7.5/6.5		7.5/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	1.3		1.0		—
Index component value				0.8	1.3	0.8	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	2.43					
	P	mW		48.51		50.24		—
	$P_{1 \times 1}$	mW		22.92		24.11		
	Z_s	cm			1.80			
	Z_b	cm					1.85	
	Z_{MI}	cm	1.67					
	$Z_{pii,\alpha}$	cm	1.67					
	f_{awf}	MHz	7.30	7.33		7.35		—
Other Information	pr	Hz	3012					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	210.25					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	235.66					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	547.03					
	p_r at Z_{pii}	<i>Mpa</i>	2.61					
Operating control conditions	Focus	cm	2.0	1.5		2.0		—
	Depth	cm	3.70	3.70		3.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-10**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.7		0.9		—
Index component value				0.7	0.7	0.9	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.17					
	P	mW		38.12		38.12		—
	$P_{1 \times 1}$	mW		19.95		19.95		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.15					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.35	7.37		7.37		—
Other Information	pr	Hz	2962					
	srr	Hz	23.14					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	42.43					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.0	1.0		1.1		—
Index component value				1.0	1.0	1.1	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.77					
	P	mW		45.65		45.65		—
	$P_{1 \times 1}$	mW		27.42		27.42		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	4.55					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.65	7.66		7.66		—
Other Information	p_{rr}	Hz	3057					
	s_{rr}	Hz	23.88					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	22.45					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	248.56					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	9.0	4.0		4.0		—
	Depth	cm	9.86	13.55		13.55		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.7		0.8		—
Index component value				0.7	0.7	0.8	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.66					
	P	mW		33.25		33.25		—
	$P_{1 \times 1}$	mW		19.07		19.07		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	4.52					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.68	7.71		7.71		—
Other Information	pr	Hz	16000					
	srr	Hz	125.00					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	20.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	221.22					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	4.0	3.0		3.0		—
	Depth	cm	9.86	11.09		11.09		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0/8.0	8.0/8.0		8.0/8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.5		0.8		—
Index component value				0.3	0.5	0.3	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.68					
	P	mW		12.56		13.45		—
	$P_{1 \times 1}$	mW		6.87		7.44		
	Z_s	cm			3.11			
	Z_b	cm					4.20	
	Z_{MI}	cm	5.61					
	$Z_{pii,\alpha}$	cm	5.61					
	f_{awf}	MHz	7.85	7.87		7.89		—
Other Information	pr	Hz	16000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	213.12					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	75.24					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	1576.11					
	p_r at Z_{pii}	<i>Mpa</i>	1.62					
Operating control conditions	Focus	cm	1.0	5.0		5.0		—
	Depth	cm	7.39	8.62		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0/8.0	8.0/8.0		8.0/8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.1		0.9		—
Index component value				0.8	1.1	0.8	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.24					
	P	mW		45.22		49.37		—
	$P_{1 \times 1}$	mW		21.46		23.27		
	Z_s	cm			2.61			
	Z_b	cm					3.05	
	Z_{MI}	cm	2.54					
	$Z_{pii,\alpha}$	cm	2.54					
	f_{awf}	MHz	7.81	7.83		7.84		—
Other Information	pr	Hz	5460					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	124.50					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	70.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	275.99					
	p_r at Z_{pii}	Mpa	1.62					
Operating control conditions	Focus	cm	5.0	5.0		6.0		—
	Depth	cm	8.62	8.62		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L4-15

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.7		0.9		—
Index component value				0.7	0.7	0.9	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.25					
	P	mW		40.15		40.15		—
	$P_{1 \times 1}$	mW		20.22		20.22		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	5.01					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.92	7.94		7.94		—
Other Information	pr	Hz	3186					
	srr	Hz	24.89					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	22.45					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	248.56					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	9.0	4.0		4.0		—
	Depth	cm	9.86	13.55		13.55		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-15P**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.2		0.4		—
Index component value				0.2	0.2	0.4	0.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.17					
	P	mW		14.56		14.56		—
	$P_{1 \times 1}$	mW		5.88		5.88		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.83					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	9.65	9.65		9.65		—
Other Information	p_{rr}	Hz	5420					
	s_{rr}	Hz	42.34					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	18.25					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	61.81					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	10.0	10.0		10.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-15P**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.3		0.4		—
Index component value				0.3	0.3	0.4	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.56					
	P	mW		16.28		16.28		—
	$P_{1 \times 1}$	mW		7.78		7.78		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.91					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	9.78	9.80		9.80		—
Other Information	pr	Hz	20000					
	srr	Hz	156.25					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	25.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	91.25					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	9.86	8.62		8.62		—
	Scanning width	%	100	100		100		—
	Freq	MHz	10.0/10.0	10.0/10.0		10.0/10.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-15P**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.3	0.7		0.8		—
Index component value				0.6	0.7	0.8	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.94					
	P	mW		27.88		30.22		—
	$P_{1 \times 1}$	mW		13.25		15.14		
	Z_s	cm			1.81			
	Z_b	cm					1.86	
	Z_{MI}	cm	3.45					
	$Z_{pii,\alpha}$	cm	3.45					
	f_{awf}	MHz	9.79	9.80		9.80		—
Other Information	p_{rr}	Hz	20000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	125.46					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	101.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	1039.98					
	p_r at Z_{pii}	Mpa	0.83					
Operating control conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	6.16	6.16		6.16		—
	Scanning width	%	100	100		100		—
	Freq	MHz	10.0/10.0	10.0/10.0		10.0/10.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-15P

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.9		1.2		—
Index component value				0.6	0.9	0.6	1.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.26					
	P	mW		28.34		30.12		—
	$P_{1 \times 1}$	mW		12.22		14.51		
	Z_s	cm			1.85			
	Z_b	cm					1.87	
	Z_{MI}	cm	1.77					
	$Z_{pii,\alpha}$	cm	1.77					
	f_{awf}	MHz	9.96	9.97		9.99		—
Other Information	p_{rr}	Hz	5579					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	181.45					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	98.44					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	332.71					
	p_r at Z_{pii}	<i>Mpa</i>	1.84					
Operating control conditions	Focus	cm	2.5	2.5		2.0		—
	Depth	cm	4.93	4.93		6.16		—
	Scanning width	%	100	100		100		—
	Freq	MHz	10.0	10.0		10.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: L5-15P

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.3		0.5		—
Index component value				0.3	0.3	0.5	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.57					
	P	mW		18.59		18.59		—
	$P_{1 \times 1}$	mW		7.14		7.14		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.84					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	9.89	9.92		9.92		—
Other Information	pr	Hz	5546					
	srr	Hz	43.33					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	18.25					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	61.81					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.5	2.5		2.5		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	10.0	10.0		10.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E3-14**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.3		0.4		—
Index component value				0.3	0.3	0.4	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.52					
	P	mW		14.66		14.66		—
	$P_{1 \times 1}$	mW		9.75		9.75		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.28					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.43	6.46		6.46		—
Other Information	p_{rr}	Hz	4366					
	s_{rr}	Hz	34.11					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	25.44					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	70.04					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating control conditions	Focus	cm	2.5	1.5		1.5		—
	Depth	cm	3.93	2.70		2.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E3-14**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.8		1.6		—
Index component value				0.8	0.8	1.6	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.15					
	P	mW		60.21		60.21		—
	$P_{1 \times 1}$	mW		31.81		31.81		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.48					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	5.28	5.28		5.28		—
Other Information	pr	Hz	6000					
	srr	Hz	46.88					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	32.25					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	79.41					
	p_r at z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.5	2.0		2.0		—
	Depth	cm	10.09	3.93		3.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/5.3	7.5/5.3		7.5/5.3		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E3-14

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.9		1.3		—
Index component value				0.8	0.9	1.3	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.38					
	P	mW		48.59		51.11		—
	$P_{1 \times 1}$	mW		31.52		33.20		
	Z_s	cm			1.82			
	Z_b	cm					1.93	
	Z_{MI}	cm	2.58					
	$Z_{pii,\alpha}$	cm	2.58					
	f_{awf}	MHz	5.31	5.33		5.32		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	138.54					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	149.62					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	385.46					
	p_r at Z_{pii}	<i>Mpa</i>	1.43					
Operating control conditions	Focus	cm	3.0	2.5		2.0		—
	Depth	cm	8.86	6.39		2.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/5.3	7.5/5.3		7.5/5.3		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E3-14

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.9		0.5		—
Index component value				0.4	0.9	0.4	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.05					
	P	mW		35.11		45.24		—
	$P_{1 \times 1}$	mW		13.16		16.48		
	Z_s	cm			2.22			
	Z_b	cm					2.31	
	Z_{MI}	cm	2.45					
	$z_{pii,\alpha}$	cm	2.45					
	f_{awf}	MHz	6.89	6.91		6.93		—
Other Information	pr	Hz	5623					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	152.23					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	148.46					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	476.46					
	p_r at z_{pii}	Mpa	1.32					
Operating control conditions	Focus	cm	3.0	2.0		3.0		—
	Depth	cm	10.09	5.16		10.09		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E3-14

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.4		0.6		—
Index component value				0.4	0.4	0.6	0.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.27					
	P	mW		25.16		25.16		—
	$P_{1 \times 1}$	mW		13.24		13.24		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.28					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.43	6.46		6.46		—
Other Information	p_{rr}	Hz	4366					
	s_{rr}	Hz	34.11					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	25.44					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	70.04					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.5	1.5		1.5		—
	Depth	cm	3.93	2.70		2.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-13**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.3		0.5		—
Index component value				0.3	0.3	0.5	0.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.05					
	P	mW		19.24		19.24		—
	$P_{1 \times 1}$	mW		9.62		9.62		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.48					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.88	6.55		6.55		—
Other Information	p_{rr}	Hz	4355					
	s_{rr}	Hz	34.02					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	25.55					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	83.04					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating control conditions	Focus	cm	3.0	3.0		3.0		—
	Depth	cm	6.39	3.93		3.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-13

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.8		1.3		—
Index component value				0.8	0.8	1.3	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.26					
	P	mW		49.24		49.24		—
	$P_{1 \times 1}$	mW		25.69		25.69		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.38					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.35	6.54		6.54		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	46.88					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	30.22					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	85.84					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating control conditions	Focus	cm	3.0	2.0		2.0		—
	Depth	cm	10.09	3.93		3.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/7.5	7.5/7.5		7.5/7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-13

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.9		1.3		—
Index component value				0.9	0.7	1.3	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.52					
	P	mW		45.33		48.54		—
	$P_{1 \times 1}$	mW		29.72		32.23		
	Z_s	cm			2.32			
	Z_b	cm					2.40	
	Z_{MI}	cm	2.22					
	$Z_{pii,\alpha}$	cm	2.22					
	f_{awf}	MHz	6.39	6.36		6.41		—
Other Information	p_{rr}	Hz	6000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	130.21					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	122.24					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	325.68					
	p_r at Z_{pii}	<i>Mpa</i>	1.55					
Operating control conditions	Focus	cm	2.5	2.0		3.0		—
	Depth	cm	8.86	6.39		8.86		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/7.5	7.5/7.5		7.5/7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-13**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.3	1.2		1.1		—
Index component value				1.0	1.2	1.0	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.79					
	P	mW		48.87		50.77		—
	$P_{1 \times 1}$	mW		30.47		32.22		
	Z_s	cm			2.22			
	Z_b	cm					2.35	
	Z_{MI}	cm	2.56					
	$Z_{pii,\alpha}$	cm	2.56					
	f_{awf}	MHz	6.88	6.89		6.92		—
Other Information	p_{rr}	Hz	5870					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	138.50					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	128.48					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	133.73					
	p_r at Z_{pii}	<i>Mpa</i>	1.29					
Operating control conditions	Focus	cm	3.0	2.0		3.0		—
	Depth	cm	6.39	3.93		6.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: E4-13**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.4		0.6		—
Index component value				0.4	0.4	0.6	0.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.31					
	P	mW		24.12		24.12		—
	$P_{1 \times 1}$	mW		12.82		12.82		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.48					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.88	6.55		6.55		—
Other Information	p_{rr}	Hz	4355					
	s_{rr}	Hz	34.02					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	25.55					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	83.04					
	p_r at z_{pii}	<i>Mpa</i>	—					
Operating control conditions	Focus	cm	3.0	3.0		3.0		—
	Depth	cm	6.39	3.93		3.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MT4-6**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	1.1		1.3		1.3
Index component value				1.1	1.1	1.3	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.92					
	P	mW		101.22		101.22		101.22
	$P_{1 \times 1}$	mW		53.97		53.97		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.56					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.26	4.28		4.28		4.28
Other Information	p_{rr}	Hz	3058					
	s_{rr}	Hz	23.89					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	11.89					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	45.49					
	p_r at z_{pii}	Mpa	–					
Operating control conditions	Focus	cm	9.0	4.0		4.0		4.0
	Depth	cm	9.86	13.55		13.55		13.55
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0	5.0		5.0		5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MT4-6

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.6		0.8		0.8
Index component value				0.6	0.6	0.8	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.71					
	P	mW		56.94		56.94		56.94
	$P_{1 \times 1}$	mW		28.77		28.77		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.19					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.59	4.38		4.38		4.38
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	70.31					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	11.59					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	43.76					
	p_r at Z_{pii}	<i>Mpa</i>	–					
Operating control conditions	Focus	cm	4.0	3.0		3.0		3.0
	Depth	cm	9.86	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0/5.0	5.0/5.0		5.0/5.0		5.0/5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MT4-6**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.6		0.8		0.8
Index component value				0.4	0.6	0.8	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.82					
	P	mW		55.33		58.46		58.46
	$P_{1 \times 1}$	mW		25.94		28.76		
	Z_s	cm			3.89			
	Z_b	cm					3.98	
	Z_{MI}	cm	4.59					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.19	4.20		4.20		4.20
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	99.87					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	101.25					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	382.26					
	p_r at Z_{pii}	Mpa	1.48					
Operating control conditions	Focus	cm	1.0	5.0		5.0		5.0
	Depth	cm	7.39	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0/5.0	5.0/5.0		5.0/5.0		5.0/5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MT4-6**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.8		1.1		1.0
Index component value				0.7	0.8	0.6	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.49					
	P	mW		70.22		72.54		72.54
	$P_{1 \times 1}$	mW		35.12		38.13		
	Z_s	cm			2.75			
	Z_b	cm					2.95	
	Z_{MI}	cm	2.58					
	$z_{pii,\alpha}$	cm	2.58					
	f_{awf}	MHz	4.55	4.57		4.57		4.57
Other Information	p_{rr}	Hz	5420					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	105.46					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	108.44					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	244.85					
	p_r at z_{pii}	<i>Mpa</i>	1.63					
Operating control conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	8.62	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0	5.0		5.0		5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MT4-6

Operating Mode: CW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.5		0.9		0.9
Index component value				0.5	0.4	0.9	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.82					
	P	mW		62.28		66.58		66.58
	$P_{1 \times 1}$	mW		28.55		31.19		
	Z_s	cm			3.55			
	Z_b	cm					3.72	
	Z_{MI}	cm	4.55					
	$Z_{pii,\alpha}$	cm	4.55					
	f_{awf}	MHz	4.19	4.20		4.22		4.22
Other Information	p_{rr}	Hz	0					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	85.45					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	88.46					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	330.13					
	p_r at Z_{pii}	<i>Mpa</i>	1.48					
Operating control conditions	Focus	cm	1.0	4.0		5.0		5.0
	Depth	cm	7.39	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0/5.0	5.0/5.0		5.0/5.0		5.0/5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MT4-6

Operating Mode: CFM-M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.6		0.8		0.8
Index component value				0.5	0.6	0.8	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.44					
	P	mW		55.48		56.82		56.82
	$P_{1 \times 1}$	mW		25.59		25.66		
	Z_s	cm			2.64			
	Z_b	cm					2.82	
	Z_{MI}	cm	2.59					
	$z_{pii,\alpha}$	cm	2.59					
	f_{awf}	MHz	4.23	4.26		4.25		4.25
Other Information	p_{rr}	Hz	5221					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	87.49					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	92.19					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	196.49					
	p_r at z_{pii}	Mpa	1.49					
Operating control conditions	Focus	cm	6.0	5.0		6.0		6.0
	Depth	cm	8.62	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0	5.0		5.0		5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MT4-6

Operating Mode: TDI

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.7		0.9		0.9
Index component value				0.7	0.7	0.9	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.93					
	P	mW		68.49		68.49		68.49
	$P_{1 \times 1}$	mW		33.64		33.64		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.22					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.62	4.37		4.37		4.37
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	70.31					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	12.24					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	47.06					
	p_r at z_{pii}	Mpa	–					
Operating control conditions	Focus	cm	4.0	3.0		3.0		3.0
	Depth	cm	9.86	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	5.0/5.0	5.0/5.0		5.0/5.0		5.0/5.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S4-12

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.3		1.8		1.8
Index component value				1.3	1.3	1.8	1.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.73					
	P	mW		60.22		60.22		60.22
	$P_{1 \times 1}$	mW		58.21		58.21		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	2.25					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.67	4.69		4.69		4.69
Other Information	p_{rr}	Hz	4187					
	s_{rr}	Hz	32.71					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	16.87					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	34.86					
	p_r at z_{pii}	<i>Mpa</i>	–					
Operating control conditions	Focus	cm	5.0	8.0		8.0		8.0
	Depth	cm	23.32	9.86		9.86		9.86
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3	5.3		5.3		5.3
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S4-12**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.8		1.2		1.2
Index component value				0.8	0.8	1.2	0.8	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.21					
	P	mW		42.22		42.22		42.22
	$P_{1 \times 1}$	mW		41.08		41.08		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	2.28					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	4.08	4.09		4.09		4.09
Other Information	p_{rr}	Hz	10000					
	s_{rr}	Hz	78.13					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	25.55					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	48.58					
	p_r at Z_{pii}	<i>Mpa</i>	–					
Operating control conditions	Focus	cm	6.0	7.0		7.0		7.0
	Depth	cm	19.71	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3/4.0	5.3/4.0		5.3/4.0		5.3/4.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S4-12

Operating Mode: PW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.5	0.5		1.0		0.7
Index component value				0.4	0.5	0.7	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.23					
	P	mW		21.25		25.44		25.44
	$P_{1 \times 1}$	mW		20.22		23.39		
	Z_s	cm			1.75			
	Z_b	cm					1.88	
	Z_{MI}	cm	1.98					
	$z_{pii,\alpha}$	cm	1.98					
	f_{awf}	MHz	6.02	6.05		6.04		6.04
Other Information	pr	Hz	10000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	107.55					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	126.52					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	288.24					
	p_r at z_{pii}	<i>Mpa</i>	1.38					
Operating control conditions	Focus	cm	7.0	5.0		5.0		5.0
	Depth	cm	7.39	11.09		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3/6.4	5.3/6.4		5.3/6.4		5.3/6.4
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S4-12**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.5		1.1		1.5
Index component value				1.0	1.5	1.0	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.76					
	P	mW		44.22		50.11		50.11
	$P_{1 \times 1}$	mW		43.12		48.89		
	Z_s	cm			1.98			
	Z_b	cm					2.05	
	Z_{MI}	cm	2.15					
	$Z_{pii,\alpha}$	cm	2.15					
	f_{awf}	MHz	4.85	4.87		4.88		4.88
Other Information	p_{rr}	Hz	6020					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	152.46					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	148.54					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	305.26					
	p_r at Z_{pii}	<i>Mpa</i>	2.35					
Operating control conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	11.09	11.09		8.62		8.62
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3	5.3		5.3		5.3
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S4-12**Operating Mode: CW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.8		1.3		1.2
Index component value				0.8	0.7	1.2	1.3	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	0.85					
	P	mW		40.22		41.16		41.16
	$P_{1 \times 1}$	mW		39.45		40.65		
	Z_s	cm			3.92			
	Z_b	cm					4.10	
	Z_{MI}	cm	3.94					
	$Z_{pii,\alpha}$	cm	3.94					
	f_{awf}	MHz	4.53	4.72		4.75		4.72
Other Information	p_{rr}	Hz	0					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	28.44					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	102.28					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	350.94					
	p_r at Z_{pii}	Mpa	0.08					
Operating control conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	6.16	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3/5.3	5.3/5.3		5.3/5.3		5.3/5.3
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S4-12**Operating Mode: CFM-M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.4		0.5		0.5
Index component value				0.3	0.4	0.5	0.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.28					
	P	mW		14.45		18.88		18.88
	$P_{1 \times 1}$	mW		13.76		17.59		
	Z_s	cm			2.92			
	Z_b	cm					2.99	
	Z_{MI}	cm	3.20					
	$Z_{pii,\alpha}$	cm	3.20					
	f_{awf}	MHz	4.55	4.58		4.62		4.58
Other Information	p_{rr}	Hz	198					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	53.28					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	106.25					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	290.49					
	p_r at Z_{pii}	<i>Mpa</i>	2.33					
Operating control conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	6.16	8.62		8.62		8.62
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3	5.3		5.3		5.3
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: S4-12

Operating Mode: TDI

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.7		1.1		1.1
Index component value				0.7	0.7	1.1	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.41					
	P	mW		36.66		36.66		36.66
	$P_{1 \times 1}$	mW		35.94		35.94		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.22					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	4.06	4.09		4.09		4.09
Other Information	pr	Hz	10000					
	srr	Hz	78.13					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	22.38					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	41.71					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	6.0	7.0		7.0		7.0
	Depth	cm	19.71	11.09		11.09		11.09
	Scanning width	%	100	100		100		100
	Freq	MHz	5.3/4.0	5.3/4.0		5.3/4.0		5.3/4.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: ML4-6**Operating Mode: B**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.5		0.9		—
Index component value				0.5	0.5	0.9	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	2.51					
	P	mW		37.18		37.18		—
	$P_{1 \times 1}$	mW		15.41		15.41		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.23					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.78	7.80		7.80		—
Other Information	p_{rr}	Hz	2915					
	s_{rr}	Hz	22.77					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.55					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	51.55					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating control conditions	Focus	cm	2.5	3.0		3.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: ML4-6**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	0.7		1.4		—
Index component value				0.7	0.7	1.4	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.04					
	P	mW		55.24		55.24		—
	$P_{1 \times 1}$	mW		22.54		22.54		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.60					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.80	6.82		6.82		—
Other Information	p_{rr}	Hz	20000					
	s_{rr}	Hz	156.25					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	20.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	42.85					
	p_r at Z_{pii}	<i>Mpa</i>	—					
Operating control conditions	Focus	cm	2.0	1.5		1.5		—
	Depth	cm	4.93	3.70		3.70		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0/6.5	8.0/6.5		8.0/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: ML4-6**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.8		1.0		—
Index component value				0.8	0.7	1.0	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.55					
	P	mW		40.12		40.12		—
	$P_{1 \times 1}$	mW		22.10		22.10		
	Z_s	cm			1.74			
	Z_b	cm					1.80	
	Z_{MI}	cm	1.73					
	$Z_{pii,\alpha}$	cm	1.73					
	f_{awf}	MHz	6.68	6.70		6.72		—
Other Information	pr	Hz	20000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	234.10					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	182.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	404.82					
	p_r at Z_{pii}	<i>Mpa</i>	2.26					
Operating control conditions	Focus	cm	3.5	3.5		3.5		—
	Depth	cm	7.39	7.39		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0/6.5	8.0/6.5		8.0/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: ML4-6**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	1.1		0.7		—
Index component value				0.6	1.1	0.6	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.94					
	P	mW		38.41		38.41		—
	$P_{1 \times 1}$	mW		18.41		18.41		
	Z_s	cm			1.88			
	Z_b	cm					1.92	
	Z_{MI}	cm	2.01					
	$z_{pii,\alpha}$	cm	2.01					
	f_{awf}	MHz	7.67	7.68		7.70		—
Other Information	p_{rr}	Hz	2952					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	250.46					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	150.22					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	435.34					
	p_r at z_{pii}	<i>Mpa</i>	2.91					
Operating control conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	11.09	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: ML4-6

Operating Mode: Elasticity imaging

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.4		0.8		—
Index component value				0.4	0.4	0.8	0.4	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	2.24					
	P	mW		31.26		31.26		—
	$P_{1 \times 1}$	mW		12.24		12.24		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	2.13					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	7.83	7.85		7.85		—
Other Information	p_{rr}	Hz	3025					
	s_{rr}	Hz	23.63					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	15.55					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	51.55					
	p_r at z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	2.5	3.0		3.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	8.0	8.0		8.0		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MC1-5

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	1.0		1.5		—
Index component value				1.0	1.0	1.5	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.31					
	P	mW		149.35		149.35		—
	$P_{1 \times 1}$	mW		59.32		59.32		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.57					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.52	3.54		3.54		—
Other Information	p_{rr}	Hz	9685					
	s_{rr}	Hz	75.66					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	13.33					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	65.86					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	7.0	7.0		7.0		—
	Depth	cm	20.45	20.45		20.45		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MC1-5**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.6		1.4		—
Index component value				0.6	0.6	1.4	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.12					
	P	mW		145.78		145.78		—
	$P_{1 \times 1}$	mW		35.69		35.69		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.25					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.50	3.53		3.53		—
Other Information	pr	Hz	12000					
	srr	Hz	93.75					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.54					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	70.42					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	6.0	8.0		8.0		—
	Depth	cm	11.82	9.36		9.36		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MC1-5**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.4	1.0		2.0		–
Index component value				1.0	0.7	2.0	1.0	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.75					
	P	mW		205.55		199.52		–
	$P_{1 \times 1}$	mW		58.76		55.49		
	Z_s	cm			4.12			
	Z_b	cm					6.40	
	Z_{MI}	cm	4.45					
	$Z_{pii,\alpha}$	cm	4.45					
	f_{awf}	MHz	3.51	3.53		3.54		–
Other Information	pr	Hz	12000					
	srr	Hz	–					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	131.52					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	125.54					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	369.29					
	p_r at Z_{pii}	<i>Mpa</i>	1.56					
Operating control conditions	Focus	cm	6.0	7.0		10.0		–
	Depth	cm	9.36	8.13		11.82		–
	Scanning width	%	100	100		100		–
	Freq	MHz	3.5/3.5	3.5/3.5		3.5/3.5		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MC1-5

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.3		0.7		—
Index component value				0.6	1.3	0.6	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.50					
	P	mW		155.49		155.49		—
	$P_{1 \times 1}$	mW		38.44		38.44		
	Z_s	cm			4.23			
	Z_b	cm					5.14	
	Z_{MI}	cm	5.56					
	$Z_{pii,\alpha}$	cm	5.56					
	f_{awf}	MHz	3.52	3.55		3.56		—
Other Information	p_{rr}	Hz	4846					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	118.25					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	115.54					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	446.55					
	p_r at Z_{pii}	Mpa	2.64					
Operating control conditions	Focus	cm	7.0	6.0		6.0		—
	Depth	cm	8.13	6.89		6.89		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MC1-5**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	1.2		1.6		—
Index component value				1.2	1.2	1.6	1.2	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.51					
	P	mW		162.33		162.33		—
	$P_{1 \times 1}$	mW		70.05		70.05		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	6.62					
	$z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	3.58	3.60		3.60		—
Other Information	pr	Hz	9685					
	srr	Hz	75.66					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	13.33					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	65.86					
	p_r at z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	7.0	7.0		7.0		—
	Depth	cm	20.45	20.45		20.45		—
	Scanning width	%	100	100		100		—
	Freq	MHz	3.5	3.5		3.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MS1-5

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.0	1.1		1.5		1.5
Index component value				1.1	1.1	1.5	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.61					
	P	mW		122.55		122.55		122.55
	$P_{1 \times 1}$	mW		86.84		86.84		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	4.48					
	$z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	2.58	2.66		2.66		2.66
Other Information	p_{rr}	Hz	3325					
	s_{rr}	Hz	25.98					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	15.88					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	35.29					
	p_r at z_{pii}	<i>Mpa</i>	–					
Operating control conditions	Focus	cm	4.0	1.0		1.0		1.0
	Depth	cm	25.87	29.57		29.57		29.57
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0	3.0		3.0		3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MS1-5**Operating Mode: B+C**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.6		0.9		0.9
Index component value				0.6	0.6	0.9	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.28					
	P	mW		71.22		71.22		71.22
	$P_{1 \times 1}$	mW		47.91		47.91		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	3.42					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	2.55	2.63		2.63		2.63
Other Information	pr	Hz	9000					
	srr	Hz	70.31					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	20.25					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	36.99					
	p_r at Z_{pii}	Mpa	–					
Operating control conditions	Focus	cm	4.0	1.0		1.0		1.0
	Depth	cm	20.94	7.39		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		3.0/3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MS1-5**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.8	0.8		1.0		1.0
Index component value				0.7	0.8	1.0	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.28					
	P	mW		84.55		84.55		84.55
	$P_{1 \times 1}$	mW		56.98		56.98		
	Z_s	cm			2.98			
	Z_b	cm					3.16	
	Z_{MI}	cm	5.88					
	$z_{pii,\alpha}$	cm	5.88					
	f_{awf}	MHz	2.58	2.58		2.58		2.58
Other Information	p_{rr}	Hz	9000					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	148.22					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	153.62					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	438.09					
	p_r at z_{pii}	<i>Mpa</i>	1.16					
Operating control conditions	Focus	cm	1.0	1.0		1.0		1.0
	Depth	cm	7.39	7.39		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		3.0/3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MS1-5**Operating Mode: B+M**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			1.3	1.2		1.1		1.8
Index component value				1.0	1.2	1.0	1.1	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	2.13					
	P	mW		122.24		123.41		123.41
	$P_{1 \times 1}$	mW		77.78		78.24		
	Z_s	cm			2.90			
	Z_b	cm					2.95	
	Z_{MI}	cm	2.60					
	$Z_{pii,\alpha}$	cm	2.60					
	f_{awf}	MHz	2.68	2.70		2.72		2.72
Other Information	p_{rr}	Hz	5690					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	135.60					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	150.22					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	243.09					
	p_r at Z_{pii}	<i>Mpa</i>	1.62					
Operating control conditions	Focus	cm	5.0	5.0		6.0		6.0
	Depth	cm	8.62	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0	3.0		3.0		3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MS1-5

Operating Mode: CW

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.2	0.7		1.6		0.8
Index component value				0.6	0.7	0.8	1.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	0.34					
	P	mW		65.25		68.44		68.44
	$P_{1 \times 1}$	mW		45.32		47.26		
	Z_s	cm			5.33			
	Z_b	cm					5.74	
	Z_{MI}	cm	6.68					
	$Z_{pii,\alpha}$	cm	6.68					
	f_{awf}	MHz	2.82	2.78		2.82		2.82
Other Information	p_{rr}	Hz	0					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	45.22					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	125.58					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	461.37					
	p_r at Z_{pii}	<i>Mpa</i>	0.15					
Operating control conditions	Focus	cm	5.0	4.0		6.0		6.0
	Depth	cm	6.16	8.62		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		3.0/3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MS1-5

Operating Mode: CFM-M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.9	0.4		0.6		0.6
Index component value				0.4	0.2	0.6	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	<i>Mpa</i>	1.46					
	P	mW		48.13		50.22		50.22
	$P_{1 \times 1}$	mW		31.58		33.24		
	Z_s	cm			5.58			
	Z_b	cm					5.62	
	Z_{MI}	cm	6.08					
	$z_{pii,\alpha}$	cm	6.08					
	f_{awf}	MHz	2.64	2.66		2.64		2.64
Other Information	p_{rr}	Hz	249					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $z_{pii,\alpha}$	W/cm ²	81.22					
	$I_{spta,\alpha}$ at $z_{pii,\alpha}$ or $z_{sii,\alpha}$	mW/cm ²	115.18					
	I_{spta} at z_{pii} or z_{sii}	mW/cm ²	349.07					
	p_r at z_{pii}	<i>Mpa</i>	2.59					
Operating control conditions	Focus	cm	5.0	3.0		6.0		6.0
	Depth	cm	6.16	8.62		8.62		8.62
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0	3.0		3.0		3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: MS1-5

Operating Mode: TDI

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.5		0.7		0.7
Index component value				0.5	0.5	0.7	0.5	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.12					
	P	mW		60.21		60.21		60.21
	$P_{1 \times 1}$	mW		39.47		39.47		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	3.38					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	2.56	2.66		2.66		2.66
Other Information	pr	Hz	9000					
	srr	Hz	70.31					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	25.21					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	45.83					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	4.0	1.0		1.0		1.0
	Depth	cm	20.94	7.39		7.39		7.39
	Scanning width	%	100	100		100		100
	Freq	MHz	3.0/3.0	3.0/3.0		3.0/3.0		3.0/3.0
	Power	%	100	100		100		100

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: LA1-15

Operating Mode: B

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.6		0.8		—
Index component value				0.6	0.6	0.8	0.6	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.52					
	P	mW		35.58		35.58		—
	$P_{1 \times 1}$	mW		19.57		19.57		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.92					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.44	6.44		6.44		—
Other Information	pr	Hz	2874					
	srr	Hz	22.45					
	n_{pps}		—	—				
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	16.47					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	38.69					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	3.0	3.0		3.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: LA1-15

Operating Mode: B+C

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	0.9		1.3		–
Index component value				0.9	0.9	1.3	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.52					
	P	mW		55.25		55.25		–
	$P_{1 \times 1}$	mW		29.35		29.35		
	Z_s	cm			–			
	Z_b	cm					–	
	Z_{MI}	cm	1.68					
	$Z_{pii,\alpha}$	cm	–					
	f_{awf}	MHz	6.44	6.42		6.42		–
Other Information	pr	Hz	20000					
	srr	Hz	156.25					
	n_{pps}		–					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	–					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	15.58					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	32.90					
	p_r at Z_{pii}	Mpa	–					
Operating control conditions	Focus	cm	2.0	2.5		2.5		–
	Depth	cm	4.93	3.70		3.70		–
	Scanning width	%	100	100		100		–
	Freq	MHz	7.5/6.5	7.5/6.5		7.5/6.5		–
	Power	%	100	100		100		–

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths z_{pii} and $z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths z_{sii} and $z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: LA1-15**Operating Mode: PW**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.9		1.1		—
Index component value				0.7	0.9	1.1	0.9	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.76					
	P	mW		44.25		45.84		—
	$P_{1 \times 1}$	mW		23.19		24.22		
	Z_s	cm			1.49			
	Z_b	cm					1.66	
	Z_{MI}	cm	1.68					
	$Z_{pii,\alpha}$	cm	1.68					
	f_{awf}	MHz	6.34	6.34		6.32		—
Other Information	pr	Hz	20000					
	srr	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	118.44					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	115.46					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	240.97					
	p_r at Z_{pii}	Mpa	1.88					
Operating control conditions	Focus	cm	2.0	2.0		3.0		—
	Depth	cm	7.39	7.39		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5/6.5	7.5/6.5		7.5/6.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: LA1-15

Operating Mode: B+M

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.6	1.3		0.7		—
Index component value				0.6	1.3	0.6	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.58					
	P	mW		36.15		40.12		—
	$P_{1 \times 1}$	mW		18.18		20.11		
	Z_s	cm			1.88			
	Z_b	cm					1.90	
	Z_{MI}	cm	2.50					
	$Z_{pii,\alpha}$	cm	2.50					
	f_{awf}	MHz	6.91	6.93		6.95		—
Other Information	p_{rr}	Hz	2788					
	s_{rr}	Hz	—					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	140.52					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	120.11					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	396.12					
	p_r at Z_{pii}	Mpa	2.56					
Operating control conditions	Focus	cm	2.0	2.0		2.0		—
	Depth	cm	7.39	3.70		7.39		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: LA1-15**Operating Mode: Elasticity imaging**

Index label			MI	TIS		TIB		TIC
				At surface	Below surface	At surface	Below surface	
Maximum index value			0.7	0.7		1.0		—
Index component value				0.7	0.7	1.0	0.7	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	Mpa	1.78					
	P	mW		44.69		44.69		—
	$P_{1 \times 1}$	mW		22.83		22.83		
	Z_s	cm			—			
	Z_b	cm					—	
	Z_{MI}	cm	1.92					
	$Z_{pii,\alpha}$	cm	—					
	f_{awf}	MHz	6.44	6.44		6.44		—
Other Information	pr	Hz	2874					
	srr	Hz	22.45					
	n_{pps}		—					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	W/cm ²	—					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	mW/cm ²	16.47					
	I_{spta} at Z_{pii} or Z_{sii}	mW/cm ²	38.69					
	p_r at Z_{pii}	Mpa	—					
Operating control conditions	Focus	cm	3.0	3.0		3.0		—
	Depth	cm	4.93	4.93		4.93		—
	Scanning width	%	100	100		100		—
	Freq	MHz	7.5	7.5		7.5		—
	Power	%	100	100		100		—

NOTE 1 Only one operating condition per index.

NOTE 2 Data should be entered for “at surface” and “below surface” both in the columns related to TIS or TIB.

NOTE 3 Information need not be provided regarding TIC for any TRANSDUCER ASSEMBLY not intended for transcranial or neonatal cephalic uses.

NOTE 4 If the requirements of 201.12.4.2a) are met, it is not required to enter any data in the columns related to TIS, TIB or TIC.

NOTE 5 If the requirements of 201.12.4.2b) are met, it is not required to enter any data in the column related to MI.

NOTE 6 The depths Z_{pii} and $Z_{pii,\alpha}$ apply to NON-SCANNING MODES, while the depths Z_{sii} and $Z_{sii,\alpha}$ apply to SCANNING MODES.

Transducer Model: **S1-5W**Operation Mode: **B**

Index Label			MI	TIS		TIB		TIC
				At Surface	Below Surface	At Surface	Below Surface	
Maximum Index Value			1.10	0.04		0.04		0.04
Index component Value				0.04	0.04	0.04	0.04	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	(MPa)	3.09					
	P	(mW)		0.92		0.92		0.67
	P_{1*1}	(mW)		0.92		0.92		
	Z_s	(cm)			0.90			
	Z_b	(cm)					0.90	
	Z_{MI}	(cm)	1.15					
	$Z_{pii,\alpha}$	(cm)	1.15					
	f_{awf}	(MHz)	7.93	8.20		8.20		10.65
Other Information	prr	(Hz)	-					
	srr	(Hz)	9.00					
	n_{pss}	NA	1.00					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	(W/cm ²)	437.39					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	(mW/cm ²)	0.09					
	I_{spta} at Z_{pii} or Z_{sii}	(mW/cm ²)	0.09					
	p_r at Z_{pii}	(MPa)	4.15					
Operating control conditions	Focus	(cm)	1.50	5.50		5.50		1.00
	Depth	(cm)	4.90	8.90		8.90		4.90
	Freq	MHz	9.00	9.00		9.00		11.00
	PRF	HZ	-	-		-		-

Transducer Model: S1-5W**Operation Mode: B+M**

Index Label			MI	TIS		TIB		TIC
				At Surface	Below Surface	At Surface	Below Surface	
Maximum Index Value			1.22	0.05		0.08		0.07
Index component Value				0.05	0.04	0.05	0.08	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	(MPa)	3.50					
	P	(mW)		2.68		2.68		2.68
	P_{1*1}	(mW)		2.68		2.68		
	Z_s	(cm)			0.95			
	Z_b	(cm)					0.95	
	Z_{MI}	(cm)	1.05					
	$Z_{pii,\alpha}$	(cm)	1.05					
	f_{awf}	(MHz)	8.24	8.20		8.20		8.20
Other Information	prr	(Hz)	245					
	srr	(Hz)	-					
	n_{pss}	NA	1.00					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	(W/cm ²)	450.24					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	(mW/cm ²)	10.26					
	I_{spta} at Z_{pii} or Z_{sii}	(mW/cm ²)	15.22					
	p_r at Z_{pii}	(MPa)	4.55					
Operating control conditions	Focus	(cm)	1.50	5.50		5.50		5.50
	Depth	(cm)	4.90	8.90		8.90		8.90
	Freq	MHz	9.00	9.00		9.00		9.00
	PRF	HZ	-	-		-		-

Transducer Model: S1-5W**Operation Mode: B+CFM**

Index Label			MI	TIS		TIB		TIC
				At Surface	Below Surface	At Surface	Below Surface	
Maximum Index Value			1.30	0.08		0.08		0.18
Index component Value				0.08	0.08	0.08	0.08	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	(MPa)	3.17					
	P	(mW)		2.81		2.81		2.81
	P_{1*1}	(mW)		2.81		2.81		
	Z_s	(cm)			0.75			
	Z_b	(cm)					0.75	
	Z_{MI}	(cm)	0.75					
	$Z_{pii,\alpha}$	(cm)	0.75					
	f_{awf}	(MHz)	5.93	5.93		5.93		5.93
Other Information	prr	(Hz)	-					
	srr	(Hz)	3.54					
	n_{pss}	NA	9.00					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	(W/cm ²)	499.13					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	(mW/cm ²)	2.74					
	I_{spta} at Z_{pii} or Z_{sii}	(mW/cm ²)	2.74					
	p_r at Z_{pii}	(MPa)	3.61					
Operating control conditions	Focus	(cm)	1.00	1.00		1.00		1.00
	Depth	(cm)	8.90	8.90		8.90		8.90
	Freq	MHz	6.00	6.00		6.00		6.00
	PRF	HZ	-	-		-		-

Transducer Model: **S1-5W**Operation Mode: **PW**

Index Label			MI	TIS		TIB		TIC
				At Surface	Below Surface	At Surface	Below Surface	
Maximum Index Value			1.19	0.23		0.83		0.33
Index component Value				0.23	0.18	0.23	0.83	
Acoustic Parameters	$p_{r,\alpha}$ at Z_{MI}	(MPa)	3.08					
	P	(mW)		7.11		7.11		7.11
	P_{1+1}	(mW)		7.11		7.11		
	Z_s	(cm)			0.50			
	Z_b	(cm)					0.85	
	Z_{MI}	(cm)	1.05					
	$Z_{pii,\alpha}$	(cm)	1.05					
	f_{awf}	(MHz)	6.67	6.66		6.66		6.66
Other Information	pr	(Hz)	4500					
	srr	(Hz)	-					
	n_{pss}	NA	1.00					
	$I_{pa,\alpha}$ at $Z_{pii,\alpha}$	(W/cm ²)	513.21					
	$I_{spta,\alpha}$ at $Z_{pii,\alpha}$ or $Z_{sii,\alpha}$	(mW/cm ²)	432.56					
	I_{spta} at Z_{pii} or Z_{sii}	(mW/cm ²)	710.03					
	p_r at Z_{pii}	(MPa)	3.82					
Operating control conditions	Focus	(cm)	1.50	7.00		7.00		7.00
	Depth	(cm)	8.90	8.90		8.90		8.90
	Freq	MHz	6.00	6.00		6.00		6.00
	PRF	HZ	4500	4500		4500		4500

APPENDIX D Display Accuracy and Acoustic Measurement Uncertainties

According to IEC60601-2-37 and NEMA UD-3 2004, the display accuracy and acoustic measurement uncertainties are summarized in the table below.

Display accuracy of MI is $\pm 20\%$, and TI is $\pm 40\%$ or < 0.2 , if MI, TI below 0.5.

Item	Measurement Uncertainty (Percentage, 95% Confidence Value)
Center Frequency	$\pm 15\%$
Acoustic Power	$\pm 30\%$
Acoustic Intensity	$\pm 30\%$
Peak Rarefactional Pressure	$\pm 15\%$

APPENDIX E Transducer Maximum Surface Temperature

According to the requirements of the section 201.11 in the standard IEC60601-2-37:2007/AMD1:2015, the transducer surface temperature has been tested in two kinds of conditions: the transducer suspended in still air or transducer contacting human-tissue mimicking material. The calculation of the expanded uncertainty is based on the ISO Guide to the Expression of uncertainty in measurement. Three transducer samples have been tested and the confidence coefficient is at 95%, the value of $t_{.975}$ is 4.30.

The measurement data were obtained under the test conditions employed at CHISON.

Transducer model	Maximum surface temperature (°C) Contacting human-tissue mimicking material	Maximum surface temperature (°C) Suspending in air	Transducer model	Maximum surface temperature(°C) Contacting human-tissue mimicking material	Maximum surface temperature (°C) Suspending in air
C1-5	≤40.5	≤40.9	S2-8	≤39.5	≤36.5
C1-6	≤40.5	≤41.6	S1-5P	≤40.7	≤37.7
C1-7	≤41.4	≤41.2	E4-10	≤40.3	≤32.6
C1-7P	≤41.5	≤41.8	VE4-10	≤40.9	≤37.2
L3-10	≤40.2	≤34.3	BL3-12	≤40.5	≤33.1
LA1-15	≤41.4	≤33.0	V2-6	≤40.5	≤40.0
L6-18	≤39.2	≤38.4	C2-6	≤39.7	≤40.5
L11-20	≤39.4	≤41.3	C4-11	≤42.5	≤39.9
L4-15B	≤39.7	≤42.2	CW2	≤38.7	≤38.6
L5-14	≤40.1	≤39.8	L6-15i	≤41.3	≤34.6
L4-10	≤38.5	≤39.5	T4-6	≤42.4	≤40.5
L5-10W	≤38.9	≤39.4	L3-8	≤40.1	≤41.5
S1-5	≤37.8	≤35.6	L4-10R	≤41.4	≤33.0
L4-15	≤39.8	≤41.5	S4-12	≤38.5	≤35.8
L5-15P	≤39.4	≤41.3	S1-5W	≤40.7	≤40.7
E3-14	≤41.7	≤37.0	ML4-6	≤40.2	≤33.3
E4-13	≤42.1	≤34.6	MC1-5	≤40.5	≤40.9
MT4-6	≤41.9	≤35.7	MS1-5	≤41.1	≤40.9



NOTE: Values following the “±” mark indicate the expanded uncertainty with a confidence lever of 95%, $t_{.975}=1.5$.

APPENDIX F Guidance and Manufacture's Declaration

1. Guidance and manufacturer's declaration – electromagnetic emissions

The SonoMax Series is intended for use in the electromagnetic environment specified below. The customer or the user of the SonoMax Series should assure that it is used in such an environment.

Emissions test	Compliance	Electromagnetic environment – guidance
RF emissions CISPR 11	Group 1	The SonoMax Series uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class A	The SonoMax Series is suitable for use in medical establishments including domestic establishments and those directly connected to the public high-voltage power supply network that used for non-domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Complies	



NOTE: The EMISSIONS characteristics of this equipment make it suitable for use in industrial areas and hospitals (CISPR 11 class A). If it is used in a residential environment (for which CISPR 11 class B is normally required) this equipment might not offer adequate protection to radio-frequency communication services. The user might need to take mitigation measures, such as relocating or re-orienting the equipment.

2. Guidance and manufacturer's declaration – electromagnetic immunity

The SonoMax Series is intended for use in the electromagnetic environment specified below. The customer or the user of the SonoMax Series should assure that it is used in such an environment.

Immunity test	Compliance level	Electromagnetic environment – guidance
Electrostatic discharge (ESD) IEC 61000-4-2	± 8 kV contact discharge ±2 kV, ±4 kV, ±8 kV, ±15 kV air discharge	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the

		relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	± 2 kV , 100 kHz, for AC power port	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	± 1 kV line(s) to line(s) ± 2 kV line(s) to earth	Mains power quality should be that of a typical commercial or hospital environment.
interruptions and voltage variations on power supply input lines IEC 61000-4-11	0 % UT; 0,5 cycle at 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315°; 0 % UT, 1 cycle and 70 % UT, 25/30 cycles Single phase: at 0° 0 % UT, 250/300 cycles	Mains power quality should be that of a typical commercial or hospital environment. If the user of the SonoMax series requires continued operation during power mains interruptions, it is recommended that the SonoMax series be powered from an uninterruptible power supply or a battery.
Power frequency magnetic field (50/60 Hz) IEC 61000-4-8	30 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE <i>UT</i> is the a.c. mains voltage prior to application of the test level.		

3 Guidance and manufacturer's declaration – electromagnetic immunity

The SonoMax Series is intended for use in the electromagnetic environment specified below. The customer or the user of the SonoMax Series should assure that it is used in such an environment.

3.1. Immunity Test	Compliance level	Electromagnetic environment – guidance
Radiated RF EM fields	3V/m, 80MHz-2.7GHz 80%AM at 1kHz	Portable and mobile RF communications

IEC 61000-4-3		equipment should be used no closer to any part of the SonoMax series, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Conducted RF IEC 61000-4-6	150 kHz to 80 MHz, 3VrmsISM band, 6Vrms	Recommended separation distance $d = 1,2 \sqrt{P}$ $d = 1,2 \sqrt{P} \quad 80 \text{ MHz to } 800 \text{ MHz}$ $d = 2,3 \sqrt{P} \quad 800 \text{ MHz to } 2,5 \text{ GHz}$ where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m) Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, should be less than the compliance level in each frequency range.^a Interference may occur in the vicinity of equipment marked with the following symbol: 
NOTE 1: At 80 MHz and 800 MHz, the higher frequency range applies. NOTE 2: These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.		
a. Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the SonoMax Series is used exceeds the applicable RF compliance level above, the SonoMax Series should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the SonoMax Series. b. Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.		

**Recommended separation distances between
portable and mobile RF communications equipment and the SonoMax Series**

Nowadays, many RF wireless equipment have being used in various healthcare locations where medical equipment and/or systems are used. When they are used in close proximity to medical equipment and/or systems, the medical equipment and/or systems" basic safety and essential performance may be affected. The Digital Color Doppler Ultrasound System has been tested with the immunity test level in the below table and meet the related requirements of IEC 60601-1-2:2014. The customer and/or user should help keep a minimum distance between RF wireless communications equipment and the Digital Color Doppler Ultrasound System recommended below.

Test frequency (MHz)	Band (MHz)	Service	Modulation	Maximum power (W)	Distance (m)	Immunity test level (V/m)
385	380-390	TETRA 400	Pulse modulation 18Hz	1.8	0.3	27
450	430-470	GMRS 460 FRS 460	FM ± 5 kHz deviation 1 kHz sine	2	0.3	28
710	704-787	LTE Band 13, 17	Pulse modulation 217Hz	0.2	0.3	9
745						
780						
810	800-960	GSM 800/900, TETRA 800, iDEN 820, CDMA 850, LTE Band 5	Pulse modulation 18Hz	2	0.3	28
870						
930						
1720		GSM 1800; CDMA 1900;				
1845						

1970	1700-1990	GSM 1900; DECT; LTE Band 1, 3, 4, 25; UMTS	Pulse modulation 217Hz	2	0.3	28
2450	2400-2570	Bluetooth, WLAN, 802.11 b/g/n, RFID 2450, LTE Band 7	Pulse modulation 217Hz	2	0.3	28
5240	5100-5800	WLAN 802.11 a/n	Pulse modulation 217Hz	0.2	0.3	9
5500						
5785						



WARNING: PORTABLE RF communications equipment (including peripherals such as antenna cables and external antennas) should be used no closer than 30 cm (12 inches) to any part of the [ME EQUIPMENT or ME SYSTEM], including cables specified by the MANUFACTURER. Otherwise, degradation of the performance of this equipment could result.

APPENDIX G Measurement Results Summary

Measurement	Useful Range	Accuracy
Distance	Full Screen	<±5%
Circumference: trace method, ellipse method	Full Screen	<±5%
Area: trace method, ellipse method	Full Screen	<±10%
Volume	Full screen	<±10%
Angle	Full screen	<±5%
Time	Full Screen	<±5%
Heart rate	Full Screen	<±5%
Velocity	Full Screen	<±10%

APPENDIX H Pulsed Wave Doppler Velocity Measurement

Results Summary

Probe type: L4-10, Scan mode: PW, fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.26	103.82	156.34
Measurement Error (%)	2.60%	3.82%	4.23%

Probe type: L4-10, Scan mode: CFM,fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.95	103.64
Measurement Error (%)	2.80%	3.90%	3.64%

Probe type: L4-10R, Scan mode: PW, fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.52	156.27
Measurement Error (%)	2.65%	3.52%	4.18%

Probe type: L4-10R, Scan mode: CFM,fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.27	51.93	103.72
Measurement Error (%)	2.72%	3.85%	3.72%

Probe type: L6-18, Scan mode: PW, fc: 10MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.29	103.65	156.35
Measurement Error (%)	2.85%	3.65%	4.23%

Probe type: L6-18, Scan mode: CFM,fc: 10MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.30	52.03	103.78
Measurement Error (%)	2.95%	4.05%	3.78%

Probe type: L3-8, Scan mode: PW, fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.85	155.93
Measurement Error (%)	2.72%	3.85%	3.95%

Probe type: L3-8, Scan mode: CFM,fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.36	51.78	103.92
Measurement Error (%)	3.58%	3.55%	3.92%

Probe type: L4-15B, Scan mode: PW, fc: 9.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.30	103.75	155.48
Measurement Error (%)	2.95%	3.75%	3.65%

Probe type: L4-15B, Scan mode: CFM,fc: 9.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.37	51.98	103.82
Measurement Error (%)	3.65%	3.95%	3.82%

Probe type: L3-10, Scan mode: PW, fc: 8.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.37	103.86	156.30
Measurement Error (%)	3.72%	3.86%	4.20%

Probe type: L3-10, Scan mode: CFM,fc: 8.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.98	103.89
Measurement Error (%)	2.79%	3.96%	3.89%

Probe type: L5-14, Scan mode: PW, fc: 8.0MHz, Power: 100%			
---	--	--	--

Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.25	103.69	156.83
Measurement Error (%)	2.52%	3.69%	4.55%

Probe type: L5-14, Scan mode: CFM,fc: 8.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.36	51.91	103.35
Measurement Error (%)	3.58%	3.82%	3.35%

Probe type: L4-15, Scan mode: PW, fc: 8.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.37	103.82	156.35
Measurement Error (%)	3.68%	3.82%	4.23%

Probe type: L4-15, Scan mode: CFM,fc: 8.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.30	51.92	103.95
Measurement Error (%)	2.95%	3.84%	3.95%

Probe type: L5-15P, Scan mode: PW, fc: 10.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.78	156.53
Measurement Error (%)	2.73%	3.78%	4.35%

Probe type: L5-15P, Scan mode: CFM,fc: 10.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.78	103.49
Measurement Error (%)	2.77%	3.56%	3.49%

Probe type: L11-20, Scan mode: PW, fc: 18.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.29	51.77	156.17

Measurement Error (%)	2.88%	3.54%	4.11%
-----------------------	-------	-------	-------

Probe type: L11-20, Scan mode: CFM,fc: 18.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.27	51.93	103.82
Measurement Error (%)	2.69%	3.85%	3.82%

Probe type: L6-15i, Scan mode: PW, fc: 9.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.28	103.92	156.80
Measurement Error (%)	2.80%	3.92%	4.53%

Probe type: L6-15i, Scan mode: CFM,fc: 9.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.29	51.75	103.34
Measurement Error (%)	2.85%	3.50%	3.34%

Probe type: C1-5, Scan mode: PW, fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.26	103.63	156.23
Measurement Error (%)	2.58%	3.63%	4.15%

Probe type: C1-5, Scan mode: CFM,fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.27	51.89	103.95
Measurement Error (%)	2.69%	3.77%	3.95%

Probe type: C1-6, Scan mode: PW, fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.49	156.26
Measurement Error (%)	2.58%	3.49%	4.17%

Probe type: C1-6, Scan mode: CFM,fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.27	51.91	103.73
Measurement Error (%)	2.69%	3.82%	3.73%

Probe type: C1-7, Scan mode: PW, fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.28	103.79	156.23
Measurement Error (%)	2.75%	3.79%	4.15%

Probe type: C1-7, Scan mode: CFM,fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.81	103.44
Measurement Error (%)	2.75%	3.62%	3.44%

Probe type: C1-7P, Scan mode: PW, fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.36	103.26	156.27
Measurement Error (%)	3.56%	3.26%	4.18%

Probe type: C1-7P, Scan mode: CFM,fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.26	51.89	103.69
Measurement Error (%)	2.59%	3.78%	3.69%

Probe type: C2-6, Scan mode: PW, fc: 3.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.22	103.52	156.35
Measurement Error (%)	2.15%	3.52%	4.23%

Probe type: C2-6, Scan mode: CFM,fc: 3.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec

Measured Target Velocity (cm/sec)	10.28	51.95	103.64
Measurement Error (%)	2.80%	3.90%	3.64%

Probe type: C4-11, Scan mode: PW, fc: 6.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.72	156.20
Measurement Error (%)	2.74%	3.72%	4.13%

Probe type: C4-11, Scan mode: CFM,fc: 6.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.26	51.91	103.79
Measurement Error (%)	2.55%	3.82%	3.79%

Probe type: E4-10, Scan mode: PW, fc: 6.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.25	103.62	156.80
Measurement Error (%)	2.53%	3.62%	4.53%

Probe type: E4-10, Scan mode: CFM,fc: 6.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.95	103.64
Measurement Error (%)	2.80%	3.90%	3.64%

Probe type: E3-14, Scan mode: PW, fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.88	156.35
Measurement Error (%)	2.65%	3.88%	4.23%

Probe type: E3-14, Scan mode: CFM,fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.26	51.91	103.72
Measurement Error (%)	2.59%	3.81%	3.72%

Probe type: E4-13, Scan mode: PW, fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.26	103.85	156.44
Measurement Error (%)	2.63%	3.85%	4.29%

Probe type: E4-13, Scan mode: CFM,fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.29	51.96	103.64
Measurement Error (%)	2.85%	3.92%	3.64%

Probe type: BL3-12, Scan mode: PW, fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.28	103.85	156.53
Measurement Error (%)	2.75%	3.85%	4.35%
Probe type: BL3-12, Scan mode: CFM,fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.91	103.61
Measurement Error (%)	2.77%	3.82%	3.61%

Probe type: VE4-10, Scan mode: PW, fc: 6.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.25	103.68	156.32
Measurement Error (%)	2.52%	3.68%	4.21%

Probe type: VE4-10, Scan mode: CFM,fc: 6.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.29	51.96	103.61
Measurement Error (%)	2.87%	3.92%	3.61%

Probe type: V2-6, Scan mode: PW, fc: 4.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec

Measured Target Velocity (cm/sec)	10.26	103.28	156.77
Measurement Error (%)	2.55%	3.28%	4.51%

Probe type: V2-6, Scan mode: CFM,fc: 4.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.91	103.75
Measurement Error (%)	2.76%	3.82%	3.75%

Probe type: CW2, Scan mode: PW, fc: 2.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.26	103.56	156.23
Measurement Error (%)	2.61%	3.56%	4.15%

Probe type: CW2, Scan mode: CFM,fc: 2.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.26	51.93	103.55
Measurement Error (%)	2.63%	3.85%	3.55%

Probe type: T4-6, Scan mode: PW, fc: 5.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.73	156.80
Measurement Error (%)	2.66%	3.73%	4.53%

Probe type: T4-6, Scan mode: CFM,fc: 5.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.30	51.91	103.58
Measurement Error (%)	2.95%	3.82%	3.58%

Probe type: MT4-6, Scan mode: PW, fc: 5.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.26	103.62	156.50
Measurement Error (%)	2.59%	3.62%	4.33%

Probe type: MT4-6, Scan mode: CFM,fc: 5.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.29	51.93	103.72
Measurement Error (%)	2.92%	3.85%	3.72%

Probe type: S1-5P, Scan mode: PW, fc: 3.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.75	156.54
Measurement Error (%)	2.65%	3.75%	4.36%

Probe type: S1-5P, Scan mode: CFM,fc: 3.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.91	103.57
Measurement Error (%)	2.79%	3.82%	3.57%

Probe type: S1-5, Scan mode: PW, fc: 2.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.27	103.78	156.50
Measurement Error (%)	2.72%	3.78%	4.33%

Probe type: S1-5, Scan mode: CFM,fc: 2.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.29	51.98	103.60
Measurement Error (%)	2.87%	3.95%	3.60%

Probe type: S2-8, Scan mode: PW, fc: 5.3MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.26	103.75	156.23
Measurement Error (%)	2.62%	3.75%	4.15%

Probe type: S2-8, Scan mode: CFM,fc: 5.3MHz, Power: 100%			
--	--	--	--

Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.98	103.66
Measurement Error (%)	2.83%	3.95%	3.66%

Probe type: S4-12, Scan mode: PW, fc: 5.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.28	103.95	156.78
Measurement Error (%)	2.82%	3.95%	4.52%

Probe type: S4-12, Scan mode: CFM,fc: 5.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.25	51.98	103.46
Measurement Error (%)	2.53%	3.95%	3.46%

Probe type: L5-10W, Scan mode: PW, fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.30	103.72	156.53
Measurement Error (%)	2.95%	3.72%	4.35%

Probe type: L5-10W, Scan mode: CFM,fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.29	51.92	103.75
Measurement Error (%)	2.85%	3.83%	3.75%

Probe type: S1-5W, Scan mode: PW, fc: 2.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.25	103.58	156.93
Measurement Error (%)	2.53%	3.58%	4.62%

Probe type: S1-5W, Scan mode: CFM,fc: 2.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.83	103.76

Measurement Error (%)	2.83%	3.66%	3.76%
-----------------------	-------	-------	-------

Probe type: ML4-6, Scan mode: PW, fc: 8.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.26	103.82	156.35
Measurement Error (%)	2.60%	3.82%	4.23%

Probe type: ML4-6, Scan mode: CFM,fc: 8.0MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.98	103.43
Measurement Error (%)	2.75%	3.95%	3.43%

Probe type: MC1-5, Scan mode: PW, fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.26	103.85	156.80
Measurement Error (%)	2.57%	3.85%	4.53%

Probe type: MC1-5, Scan mode: CFM,fc: 3.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.28	51.93	103.52
Measurement Error (%)	2.80%	3.86%	3.52%

Probe type: MS1-5, Scan mode: PW, fc: 2.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.25	103.62	156.80
Measurement Error (%)	2.53%	3.62%	4.53%

Probe type: MS1-5, Scan mode: CFM,fc: 2.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.26	51.62	103.54
Measurement Error (%)	2.56%	3.23%	3.54%

Probe type: LA1-15, Scan mode: PW, fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 100 cm/sec	V3: 150 cm/sec
Measured Target Velocity (cm/sec)	10.35	103.92	156.50
Measurement Error (%)	3.53%	3.92%	4.33%

Probe type: LA1-15, Scan mode: CFM,fc: 7.5MHz, Power: 100%			
Phantom Target Velocity (cm/sec)	V1: 10 cm/sec	V2: 50 cm/sec	V3: 100 cm/sec
Measured Target Velocity (cm/sec)	10.31	51.97	103.84
Measurement Error (%)	3.06%	3.93%	3.84%